



Manufacturer's Name: Agilent Technologies, Inc.
Manufacturer's Address: 1900 Garden of the Gods Road
Supplier's Address: Colorado Springs, CO 80907
USA

Declares under sole responsibility that the product as originally delivered

Product Name: 100 MHz High voltage differential probe
Model Number: N2790A
Product Options: This declaration covers all options of the above product(s)

complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

EMC Directive (2004/108/EC)

and conforms with the following product standards:

EMC	Standard	Limit
	IEC 61326-1:2005 / EN 61326-1:2006	
	CISPR 11 / EN 55011	Group 1 Class A
	IEC 61000-4-2 / EN 61000-4-2	4 kV CD, 8 kV AD
	IEC 61000-4-3 / EN 61000-4-3	3 V/m, 80-1000 MHz & 1.4-2 GHz, 1V/m 2-2.7 GHz
	IEC 61000-4-4 / EN 61000-4-4	0.5 kV signal lines, 1 kV power lines
	IEC 61000-4-5 / EN 61000-4-5	0.5 kV line-line, 1 kV line-ground
	IEC 61000-4-6 / EN 61000-4-6	3 V, 0.15-80 MHz
	IEC 61000-4-11 / EN 61000-4-11	1 cycle, >95%

Canada: ICES-001:2004

Australia/New Zealand: AS/NZS CISPR 11:2004

The product was tested in a typical configuration with Agilent Technologies test systems.

Safety IEC 61010-1:2001 / EN 61010-1:2001
Canada: CAN/CSA-C22.2 No. 61010-1-04
USA: ANSI/UL 61010-1:2004
UL94

Supplementary Information:

This DoC applies to above-listed products placed on the EU market after:

February 3, 2009

Date

Kirk Smith

Quality Manager

For further information, please contact your local Agilent Technologies sales office, agent or distributor,
or Agilent Technologies Deutschland GmbH, Herrenberger Straße 130, D 71034 Böblingen, Germany.

Product Regulations

EMC

IEC 61326-1:2004 / EN 61326-1 :2006

Product Safety

IEC 61010-1:2001 / EN 61010-1:2001

Canada: CAN/CSA-C22.2 No. 61010-1-04

USA: ANSI/UL 61010-1:2004

UL94

Sound Pressure Level

Less than 60 dBA

Regulatory Information for Canada

Canada: ICES-001:2004

This ISM device complies with Canadian ICES-001.

Cet appareil ISM est conforme à la norme NMB-001 du Canada.

Regulatory Information for Australia/New Zealand

This ISM device complies with Australian/New Zealand AS/NZS CISPR 11:2004



Manual Part Number: N2790-92000, February 2009



Printed in Japan
