

This handbook provides important information for persons planning to take a Florida state certification exam listed below. Policies, procedures and information in this handbook supersede previous editions. Please review this information carefully; you are responsible for understanding the contents of this handbook.

# EXAMINATION HANDBOOK

*for Florida State Certification Exams  
Administered by ARRT in*

# 2020

- Radiography
- Nuclear Medicine Technology
- Radiation Therapy
- Limited Scope of Practice in Radiography

*Basic X-Ray Machine Operator (Core, Chest, Extremities)*

*Basic X-Ray Machine Operator (Core, Podiatric)*

## ***Important Notice: State Licensing is Not ARRT Credentialing***

---

*A passing score on a state licensing examination does not make a candidate eligible for ARRT certification and registration. Candidates seeking ARRT certification and registration must have submitted an application directly to ARRT and must have met all other criteria for ARRT certification and registration. Those seeking only state licensing must meet criteria established by the state. Test scores earned as a state candidate may not be used for ARRT certification and registration.*

# How to Use This Handbook

## ▼ **Licensing vs.**

### ● **Certification and Registration**

*The information contained in this handbook pertains to Florida certification examinations and processing.*

*These Florida certification exams, their eligibility, or application process bear **no relation** in any way to national credentialing in radiologic technology offered by ARRT.*

This Examination Handbook is designed to help Florida state certification candidates understand and prepare for their examination. It is published for each calendar year, but changes to content specifications and policies and procedures may occur during the year so be sure to check [www.staterhc.org](http://www.staterhc.org) for updates. To ensure that your exam experience is as successful as possible, you will want to read the following information very carefully and keep the handbook for future reference.

The information in this handbook supersedes that in any prior publications of ARRT. Earlier versions may contain outdated information. It is your responsibility to obtain a handbook corresponding to the year in which your examination is to be taken, as noted on the front cover of the handbook; and to familiarize yourself with the contents.

ARRT does not discriminate against candidates on the basis of their race, color, creed, religion, national origin, sex, marital status, status with regard to public assistance, familial status, disability, sexual orientation, gender identity, veteran status, age or any other legally protected basis.

## **Watch for These Symbols**



This exclamation point is your pointer to key pieces of information you need to know.



This icon tips you to ways you can streamline your journey through the examination process.

## **NCCA Accreditation**

ARRT's Radiography, Nuclear Medicine Technology, Radiation Therapy, Sonography, Computed Tomography, and Registered Radiologist Assistant certification and registration programs have earned accreditation by the National Commission for Certifying Agencies (NCCA), the accrediting body of the Institute for Credentialing Excellence (ICE).

To receive NCCA accreditation, ARRT demonstrated that this certification and registration program met strict standards in accordance with ICE's mission to promote excellence in competency assurance for practitioners in all occupations and professions.

For more information on ICE/NCCA and their accreditation programs, visit them at [www.credentialingexcellence.org](http://www.credentialingexcellence.org).

ARRT is unable to respond to questions regarding licensing requirements for the state of Florida.

- Direct questions regarding your state license application, the FL DOH Rad Tech Certification Office 180-day eligibility period, or changes to your name, address, social security number, or date of birth to:

**Florida Department of Health  
Radiologic Technology Certification Office**

**4052 Bald Cypress Way Bin C85  
Tallahassee, FL 32399-3285**

**Phone: (850) 488-0595**

**Email: [MQA.rad-tech@flhealth.gov](mailto:MQA.rad-tech@flhealth.gov)**

**Website: [www.floridahealth.gov/licensing-and-regulation/radiologic-technology/](http://www.floridahealth.gov/licensing-and-regulation/radiologic-technology/)**

- After carefully reading this handbook, direct questions regarding examination procedures or ADA testing accommodations to:  
**Attn: FLStateRHC  
ARRT  
1255 Northland Drive  
St. Paul, MN 55120-1155  
Phone: (651) 687-0048, ext. 8525**
- For exam window change requests, complete the Examination Window Extension Request form found at [www.staterhc.org](http://www.staterhc.org) and fax to (651) 681-3294.

*For information about **national credentialing in radiologic technology**, contact:*

**The American Registry of Radiologic Technologists®  
1255 Northland Drive, St. Paul, Minnesota 55120-1155  
Phone: (651) 687-0048  
[www.arrt.org](http://www.arrt.org)**

# Table of Contents

## Florida Certification Exams

Certification and Registration vs  
State-Related Licensing ..... 4  
ARRT Exams for State Licensing ..... 4

## Upholding Exam Security

Why Does Security Matter So Much? ..... 5  
*Candidate Agreement on Exam Disclosure*  
Disclosing Exam Information ..... 5  
*Help Us Protect Exam Security*

## Application Procedures

Application for Examination ..... 5-6  
*Paying by Credit Card*  
*Exam Fees*  
One Exam at a Time ..... 6  
Testing Accommodations ..... 6-7  
*Testing Accommodations: Ask Up Front*

## Before the Examination

Candidate Status Report ..... 7  
*Status Report Info Incorrect?*  
Limited Scope Candidate Status Report ..... 7  
Address or Name Changes ..... 8  
*Name/Address Change? Notify the State*  
FL DOH Rad Tech Certification Office 180-Day  
Eligibility Period ..... 8  
ARRT 90-Day Examination Window ..... 8  
*State Exam Eligibility vs. ARRT Exam Window*  
Extending an Exam Window ..... 8-9  
*Window Extension: Notify ARRT*  
Test Centers ..... 9  
Study Materials ..... 9

## The Examination Appointment

Scheduling Your Appointment ..... 9-10  
*Internet Scheduling*  
*Follow-Up and Confirm Appointment*  
Missing Your Appointment ..... 10  
Canceling or Rescheduling ..... 10-11  
*Calling to Reschedule? Remember to Cancel  
Exception*

## Exam Administration Day

Test Center Environment ..... 11  
*What to Expect on Exam Day*  
Follow Procedures ..... 12  
Arrive Early ..... 12  
*Invalid ID? Cancel Appointment*  
ID, Photo, Signature, PVR ..... 12  
*Acceptable Forms of Identification*  
*Palm Vein Recognition Replaces Fingerprint*  
Assignment to Testing Station ..... 13  
Calculators and Notes ..... 13  
*Don't Bring Yours*  
Requesting Assistance ..... 13  
Exam Timing ..... 13-14  
*Pace Yourself*  
Test Center Misconduct and  
Score Cancellation ..... 14-15  
*Non-Disclosure Agreement*  
What if the Test Center is Closed? ..... 16  
*Severe Weather Looming?*  
Taking the Exam ..... 16  
Candidate Comments ..... 16  
Leaving the Test Center ..... 16  
Appeals of Exam Administration ..... 17

## After the Examination

Cancellation of Scores ..... 17  
Score Reporting ..... 17  
Interpreting Scores ..... 18  
*How/When Will I Get My Exam Results?*  
*State Attempts Count Against Three-Attempt Limit*  
Appeals of Exam Scoring ..... 18  
Re-examination ..... 18

## APPENDICES

A. Content Specifications — Radiography ..... 21-34  
B. Content Specifications — Nuclear Medicine Technology .. 35-45  
C. Content Specifications — Radiation Therapy ..... 46-54  
D. Content Specifications — Limited Scope of Practice  
in Radiography ..... 55-72  
E. Sample Questions for the Examination for the Limited Scope of  
Practice in Radiography ..... 73  
F. Test Centers ..... 74  
G. Sample Candidate Status Report ..... 75-76  
H. Sample Limited Scope Candidate Status Report ..... 77-78  
I. Computer-Based Testing Overview ..... 79-82  
J. Limited Scope Computer-Based Testing Overview ..... 83-86  
K. Potential Exam Disclosure Scenarios ..... 87  
L. ARRT Rules Agreement ..... 88  
M. Checklist ..... 89

# Florida Certification Exams for State Licensing

Florida state certification exams are administered by The American Registry of Radiologic Technologists®, but state licensing is *not* ARRT credentialing.

## **Certification and Registration vs. State-Related Licensing**

More than 75 percent of the states have licensing laws covering the practice of radiologic technology. In those states, you must obtain a state license before you can work as a radiologic technologist. In addition, many states use ARRT exam scores and/or credentials when making licensing decisions.

Terminology used in establishing the authority of a technologist is often confusing.

ARRT uses the term “certification and registration” when an individual satisfies all eligibility requirements — which include ethics, education and examination. If you wish to become certified and registered with ARRT, you need to submit an application directly to ARRT. Submitting an application to an individual state licensing agency would not make you eligible for ARRT certification and registration.

Although you may have earned your ARRT credential, this does not automatically mean that you are eligible to work in your state. Most states have their own licensing policies and procedures that you must meet in order to work in the state. Verify with the state licensing agency in the state where you plan to work to make sure you meet their eligibility requirements.

Exam scores earned as a state candidate may not be used for later application to ARRT for certification and registration; however, if you attempt to pass an exam as a state licensing candidate it will be counted as an attempt for purposes of ARRT’s three-attempt, three-year limit for certification and registration (see page 18 for details).

ARRT is unable to respond to questions regarding licensing requirements for the state of Florida. Direct questions regarding your state license application, the FL DOH Rad Tech Certification Office 180-day eligibility period, or changes to your name, address, social security number, or date of birth to (850) 488-0595. Full contact information can be found on the inside front cover of this handbook.

## **ARRT Exams for State Licensing**

The ARRT is a national, voluntary certification and registration organization that administers examinations both for its own credentialing programs and on behalf of states for their use in issuing state licenses. ARRT-developed exams are produced in collaboration with content experts from various specialties. The examinations consist of questions designed to measure the knowledge and cognitive skills underlying the intelligent performance of the major tasks typically required of a radiologic technologist. In addition to the Limited Scope of Practice in Radiography exam, ARRT makes its Radiography, Nuclear Medicine Technology, Radiation Therapy, Sonography and Magnetic Resonance Imaging exams, as well as post-primary exams available to states for state licensing purposes.

# Upholding Exam Security

ARRT has strict security regulations and takes exam security seriously. ARRT prohibits you from cheating on your exam or taking action that would help another candidate cheat. If you violate the regulations, you can face legal action and/or risk being banned from future testing.

## Why Does Security Matter So Much?

It's a matter of public health.

Security is critical to ensuring that the examination is an accurate and reliable measure of the critical knowledge and cognitive skills underlying the tasks typically required for the practice of medical imaging, interventional procedures, and radiation therapy. In fact, subverting the integrity of ARRT's exams is illegal, based on a Minnesota law that went into effect on August 1, 2010. More information can be found by visiting [www.staterhc.org](http://www.staterhc.org).

Ask yourself: Would you want a loved one to receive care from an individual who passed the ARRT-administered exam because they got a sneak peek at questions and memorized the answers rather than having learned all the critical content that the questions scientifically sample?

### ▼ Candidate Agreement On Exam Disclosure

Disclosing exam information using language that is substantially similar to that used in questions and/or answers on the ARRT examinations is considered an attempt to subvert the integrity of the exam when such information is gained as a direct result of having been a candidate. This includes (but is not limited to) disclosures to: students in educational programs, graduates of educational programs, educators, or anyone else involved in the preparation of candidates to sit for the exam. It is also considered an attempt to subvert the integrity of the exam to receive, from a candidate, exam information that uses language that is substantially similar to that used in questions and/or answers on the ARRT exam from a candidate, whether requested or not, or to relay such information.

### ▼ Help Us Protect Exam Security

If you know of any situations in which the security of ARRT exam materials might be compromised, we invite you to visit [www.staterhc.org](http://www.staterhc.org).

### Disclosing Exam Information: The Bright Line Between What's OK and What's Not

Candidates for state licensing and/or permit examinations see language in the ARRT state licensing examination handbooks, as well as the non-disclosure screens at the test center that clarify what they are agreeing to comply with regarding exam security. This language is reproduced in the box on page 15.

Failure to comply with these agreements can trigger an ARRT investigation which may lead up to the invalidation of the results of the current and any prior examinations. This could also permanently bar the candidate from all future exams as well as the appropriate state licensing agency being notified. Violating these agreements could also lead to legal action. Appendix K contains a list of potential exam disclosure scenarios.

If you have any questions about your responsibilities under ARRT's exam disclosure policy, visit [www.staterhc.org](http://www.staterhc.org). A video depicting the consequences of violating this policy is available at [www.arrt.org/video-library](http://www.arrt.org/video-library).

NOTE: ARRT reserves the right to bar state candidates from examination who are currently sanctioned by the ARRT.

# Application Procedures

## Application for Examination

### Application to State

Before paying for an ARRT-administered examination, you must submit your license application and the appropriate application fee directly to the FL DOH Rad Tech Certification Office for eligibility determination. ARRT does not determine eligibility for Florida licensing candidates.

NOTE: DO NOT include the examination fee, which is different than the FL DOH Rad Tech Certification Office application fee.

After the FL DOH Rad Tech Certification Office has determined your eligibility for examination, the FL DOH Rad Tech Certification Office will mail a confirmation letter

to you with your eligibility information and instructions for paying for your ARRT-administered state certification examination. The FL DOH Rad Tech Certification Office will also forward your eligibility information to the ARRT. A 180-day state eligibility period will start on the day the FL DOH Rad Tech Certification Office confirms your eligibility. Questions regarding the 180-day state eligibility period should be directed to the FL DOH Rad Tech Certification Office (not ARRT) listed on the inside cover of this handbook.

## Paying by

### Credit Card

*If you choose to pay by credit card, you must log on to a secure website. First time users must click on the "Register" button on the upper right side of the screen. Previous users to the site should enter their user ID and password. You are responsible for remembering your user ID and password for future use.*

## Exam Fees

*Fees paid to ARRT are non-refundable under any circumstances, nor can they be transferred to another discipline of examination.*

### **Exam Fee to ARRT**

You have two options for examination payment to ARRT. One way to pay is with a credit card by logging onto a secure website at [www.staterhc.org](http://www.staterhc.org). If you have not visited this site previously, you will be required to create a personal profile by clicking on the "Register" button on the upper right side of the home page before providing your credit card information. Previous users should enter their user ID and password, which was entered during the original log in procedure. **If you require test accommodations, credit card payment is not an option.**

The other option is to send the exam fee directly to ARRT by submitting a copy of your FL DOH Rad Tech Certification Office eligibility letter and a cashier's check or money order for the appropriate fee amount (personal checks and business checks are not accepted and will be returned).

Send the letter and fee to Attn: StateRHC, ARRT, 1255 Northland Drive, St. Paul, MN 55120. You must report any name and address discrepancies on your eligibility letter to the FL DOH Rad Tech Certification Office. Do not note name and address changes on the copy of the eligibility letter you send to ARRT — ARRT is not authorized to make changes to your information on file without official notification from the FL DOH Rad Tech Certification Office. ARRT will return to you any payment received without a copy of the FL DOH Rad Tech Certification Office eligibility letter. Please allow approximately 10 business days for processing your exam fee.

NOTE: Exam fees must be received by the ARRT at least 30 days prior to the last day of your FL DOH Rad Tech Certification Office 180-day eligibility period. Exam fees received with less than 30 days remaining will not be processed and will be returned to you.

After ARRT processes your exam payment, a Candidate Status Report (CSR) and handbook will be mailed to you. Your CSR will provide your permanent ID number. This ID number is required to schedule your exam appointment.

## **One Exam at a Time**

You may apply for one exam at a time. That means if you're planning to take a state exam (administered by ARRT) and an ARRT certification exam, you must choose which one to take first. If you choose to take your state exam first, your application and fee for an ARRT certification exam will be held and not processed until you complete your state exam. Similarly, if you have been assigned an exam window for an ARRT certification exam, your fee for your state exam will be returned to you. Once you complete your ARRT exam, you can re-submit your fee for your state licensing exam.

## **Testing Accommodations**

To comply with the Americans With Disabilities Act (ADA), we'll provide testing accommodations if our partner organization, Paradigm Testing, determines that you meet ADA requirements. Exam accommodations include any changes to standard testing procedures, including requests for additional time, a reader, as well as medical aids such as insulin pumps, Pico magnifiers, lumbar pillows, asthma inhalers, etc.

If you need accommodations, you must submit a *Request for Test Accommodations* form (located at [www.staterhc.org](http://www.staterhc.org)) along with a copy of your FL-DOH Rad Tech Certification Office eligibility letter and cashier's check or money order (personal and

**TIP**

### **Testing Accommodations: Ask Up Front**

*Requests for testing accommodations must be submitted to ARRT with a copy of your FL DOH Rad Tech Certification Office eligibility letter and fee. Candidates requesting testing accommodations may not pay by credit card.*

business checks are not accepted.) If you are requesting testing accommodations, you may not pay for your exam using the credit card option.

Once ARRT receives your information, we'll send you instructions (including a 10-digit alpha/numeric authorization code) explaining how to submit your online request for accommodations to Paradigm Testing. You'll also be required to submit documentation verifying the reason you need accommodations.

ARRT will not assign you an exam window until Paradigm notifies us of the approval or denial of your request for accommodations. If Paradigm denies your request, you have the option to appeal their decision.

If you request accommodations with ARRT, but do not submit your documentation to Paradigm, we will return your fee after 180 days, or sooner, depending on when your FL-DOH Rad Tech Certification Office 180-day eligibility period ends. If you apply for accommodations and then decide you no longer wish to go through the approval process, you will need to notify ARRT by fax (651.681.3293, attn: StateRHC Coordinator) to process your exam without test accommodations.

Remember, you must submit a *Request for Test Accommodations* form to ARRT each time you are eligible for examination. If you don't submit the form each time with a copy of your FL-DOH Rad Tech Certification Office eligibility letter and fee, you'll have to take your exam without accommodations for the assigned window.

If you asked for and received approval for ADA accommodations, ARRT will send you a letter with instructions on scheduling your appointment with Pearson VUE. You will need to bring your approval letter with you to your exam appointment and give it to test center staff when checking in.

## Before the Examination

Familiarize yourself with exam procedures explained in this handbook and on your CSR before scheduling your exam at any of hundreds of test centers across the U.S. and internationally.

### ▼ **Status Report Info**

#### ● **Incorrect?**

*If the information on your Candidate Status Report is incorrect, contact the FL DOH Rad Tech Certification Office (not ARRT) right away — and before scheduling a testing appointment.*

### **Candidate Status Report (CSR)**

ARRT will mail you a Candidate Status Report (CSR; see Appendix G for sample) after your examination fee is processed at the ARRT. The CSR contains your identification information, your permanent ARRT-assigned ID number, and examination window dates.

Be sure your name on the CSR matches your IDs that you will bring for admission to the test center (see "Acceptable Forms of Identification" on page 12). Notify the FL DOH Rad Tech Certification Office immediately and before scheduling your exam appointment if any identification information is incorrect or does not match your IDs. Do not contact ARRT with identification changes.

Do not schedule your exam appointment until you receive a new CSR and verify that the information is correct. You may then proceed to schedule your exam appointment.

If you lose your CSR or do not receive it within the anticipated processing time of four weeks, contact ARRT at (651) 687-0048, ext. 8525.

### **Limited Scope Candidate Status Report And State-Assigned Exam Modules**

Notify the FL DOH Rad Tech Certification Office (not ARRT) immediately — and before scheduling your appointment — if the modules listed on your Limited Scope CSR (see Appendix H for a sample) do not match the modules that you think you should be taking. ARRT cannot make changes to your limited scope modules without official notification from the FL DOH Rad Tech Certification Office and if you have an appointment scheduled. It is your responsibility to verify you have been assigned the correct exam modules before scheduling your appointment.

## Name/Address Change?

### Notify the State

*You should notify the FL DOH Rad Tech Certification Office (not ARRT) immediately of any name and/or address changes and before scheduling an exam appointment.*

## State Exam Eligibility

### vs. ARRT Exam Window

*Your 180-day license eligibility period, determined by the FL DOH, is different than the 90-day examination window assigned by ARRT.*

## Address or Name Changes

You must notify the FL DOH Rad Tech Certification Office (not ARRT) immediately of any changes to your name or address as submitted on your license application form. Changes cannot be processed by ARRT, the Pearson VUE Call Center or at the test center.

At the test center, the name on your IDs must match your name as it appears on your CSR (the only permissible exception is middle initial versus middle name, as long as the first letters match). Name change requests must be directed to the FL DOH Rad Tech Certification Office at least 10 business days before a scheduled appointment to allow enough time for your information to be submitted to Pearson VUE for processing. Requests received less than 10 business days before your exam appointment may not be processed in time, which may result in your being turned away from the test center and forfeiting your fee. If the name on your IDs doesn't match your CSR, cancel your appointment (see page 10) and correct the discrepancy with the FL DOH Rad Tech Certification Office. Don't schedule a new appointment until you receive a new CSR and verify the changes are correct.

## FL DOH Rad Tech Certification Office 180-Day Eligibility Period

Florida statutes state that an application expires six months after initial filing. Candidates for examination must complete their exam within 180 days of the date that appears on the eligibility letter sent to them by the FL DOH Rad Tech Certification Office. No extensions to the 180-day eligibility period can be granted. Please note that the 180-day eligibility period is different than the ARRT-assigned 90-day exam window. If you are unable to complete your exam within the ARRT-assigned 90-day exam window, you may request a window extension. See "Extending an Exam Window" below for details.

## ARRT 90-Day Examination Window

ARRT will assign you a 90-day exam window. You should schedule your exam appointment for a date within the 90-day exam window printed on your CSR. Please be aware that the ARRT-assigned 90-day exam window is different than the 180-day FL DOH Rad Tech Certification Office eligibility period (see "Application for Examination" on pages 5-6). Generally, examination windows begin on the Wednesday after payment is processed (not received) by ARRT, and extend for 90 calendar days. For example, if an exam payment is processed on Thursday, April 16, 2020, the examination window will begin on Wednesday, April 22, 2020, and end on Monday, July 20, 2020.

Your exam window will close automatically after 90 calendar days, or if you miss an appointment, if an appointment is not canceled in time, you fail to comply with the non-disclosure agreement at the test center (see page 15), the name on your IDs do not match the name on your CSR, or if you have an invalid ID. In addition your fee is forfeited. To open a new window, you would have to re-apply through the FL-DOH Rad Tech Certification Office and submit a new exam fee to ARRT.

## Extending an Exam Window

If circumstances make it impossible for you to schedule your examination during your ARRT-assigned 90-day exam window, you may request a window extension.

If you have an existing appointment, you must cancel it before requesting a window extension, scheduling a new exam date, or changing the test center location. (See "Canceling or Rescheduling Your Appointment" on pages 10-11.)

ARRT requires you complete the Window Extension Request Form and fax it to ARRT. ARRT must receive the request on or before the last day of your current ARRT 90-day examination window. If your window expires on a weekend or holiday, your request must be received on or before the last business day prior to the expiration date. (Saturday and Sunday are not considered ARRT business days.) The Window

**TIP**

### Window Extension?

#### Notify ARRT

*If you need to request a window extension, you must make sure you have time remaining in your state-assigned 180-day eligibility period. A completed Examination Window Extension Request form must be faxed to the ARRT before the last day of your current ARRT 90-day exam window. You must cancel any scheduled exam appointment before requesting a window extension. Your window cannot be extended beyond the last day of your 180-day eligibility period.*

Extension Request Form is located at the bottom of the Florida home page at [www.staterhc.org](http://www.staterhc.org). If you provide an email address, a confirmation of receipt of your request will be sent. If you do not provide an email address, you should follow-up with a phone call to (651) 687-0048, ext. 8525, to confirm that your fax has been received. Your new ARRT exam window will begin on the day ARRT processes the extension request. ARRT will not accept requests for specific window dates.

NOTE: ARRT cannot process requests it receives after the last day of your current window. Window extensions will be processed only if sufficient time remains in your FL DOH Rad Tech Certification Office 180-day license eligibility period. ARRT cannot extend your 90-day exam window beyond your FL DOH Rad Tech Certification Office 180-day license eligibility period under any circumstances. It is your responsibility to know when your 180-day eligibility period expires. This information appears on your CSR.

## Test Centers

ARRT examinations are administered by Pearson VUE, the electronic testing business of Pearson Education. Their network of more than 200 high-security test centers is specifically designed and built for professional licensure and certification markets in the U.S. and its territories. Their international test centers are equipped to deliver ARRT exams in selected cities in Canada, Europe, Asia and Australia.

A geographic list of test center locations appears in Appendix F of this handbook, but please keep in mind that location changes may occur after publication. Current test center locations and driving directions may be viewed at [www.pearsonvue.com/arrt](http://www.pearsonvue.com/arrt).

## Study Materials

Use the content specifications in the appendices of this handbook to prepare for examination. You may purchase a text and workbook for the Basic X-Ray Machine Operator exam from the FL DOH Rad Tech Certification Office. A link to this syllabus can be found on [www.staterhc.org](http://www.staterhc.org). Please note that information provided by the FL DOH Rad Tech Certification Office was not provided or is not endorsed by ARRT. ARRT does NOT provide specific lists of study materials or textbooks for any ARRT-developed exams, nor do we endorse any review programs, mock registries or study guides. It is recommended you use a variety of references when preparing for your exam.

## Scheduling Your Appointment

Pearson VUE schedules appointments on a first-come, first-served basis. Once you receive your CSR, you may schedule your appointment one of two ways:

- call the Pearson VUE Call Center at the toll-free phone number shown on your CSR (Monday–Friday, 7 a.m.–7 p.m. CT); or
- online at [www.pearsonvue.com/arrt](http://www.pearsonvue.com/arrt) (see “tip” box on next page for details on scheduling an appointment through the Internet).

Even if you don’t want to take your exam immediately, it’s better to schedule early to obtain your choice of exam date. If you delay too long in scheduling your examination, you may not find an available appointment prior to the expiration date. If your window is allowed to expire, your file is closed, and you must contact the FL DOH Rad Tech Certification Office for instructions on re-application (see “Extending an Exam Window” on pages 8-9 for details).

You will be providing and receiving a great deal of important information when scheduling your appointment with Pearson VUE. It is your responsibility to manage that information each step along the way.

### Have Your Information Available

Have your CSR at hand when going online or calling to schedule. You cannot schedule a testing appointment until you receive your CSR (see page 7). You will be able to select a test center from those listed in Appendix F or on the Pearson VUE website.

# The Examination Appointment

Once you receive your Candidate Status Report from ARRT, you’re ready to schedule the appointment.

**TIP****Internet Scheduling**

After you have been notified of your eligibility to sit for the exam, you may schedule online at [www.pearsonvue.com/arrt](http://www.pearsonvue.com/arrt). When you arrive at the web page, the process will differ depending on if you're a first-time or returning user.

First-time users should click on the "Create an Account" link, where you will be asked for your ID number and personal information from your Candidate Status Report. Make sure the information you enter on the screen matches the information on the front of your CSR. When creating your profile, follow the prompts until you have completed the process and can select the "Finish" link. You will be provided a link to follow the prompts for scheduling your exam.

Returning users should click on the "Sign In" link. If you have forgotten your password, click on the "Forgot my Password" link and follow the prompts.

To schedule online, candidates must provide an email address. Otherwise, phone the Pearson VUE Call Center directly to schedule an appointment.

When calling to schedule your appointment, you will be asked to verify your name as listed on your current CSR and provide your ARRT-assigned ID number appearing on your CSR (ARRT does not provide ID numbers over the phone.) Calls may be recorded for quality assurance purposes.

Pearson VUE Call Center staff will help you schedule a date and time for your exam. Test centers are generally open Monday through Friday between the hours of 8 a.m. and 6 p.m. Some test centers offer extended evening or weekend hours.

NOTE: Call Center staff cannot make changes (except adding email and phone info) to the application information you provided to the FL DOH Rad Tech Certification Office. (See "Address or Name Changes" on page 8.)

**Confirm Your Scheduling Information**

Space is provided on the back of your CSR for you to write the date, time, confirmation number and name of the Call Center representative. Pearson VUE will email a letter confirming your appointment.

The letter will include the address, phone number, and directions to the test center, as well as the name, date, and time of your exam and other important information. Driving directions are also available at [www.pearsonvue.com/arrt](http://www.pearsonvue.com/arrt).

NOTE: Occasionally the email confirmation may be filtered into a SPAM folder based on the security settings of your email account. Be aware that the email confirmation comes from [PearsonVUEconfirmation@pearson.com](mailto:PearsonVUEconfirmation@pearson.com). If you do not receive an email confirmation from VUE immediately after scheduling, check your email address on file, your filter settings and/or contact the VUE Call Center to confirm your appointment date and time, and request that a new confirmation email be sent.

ARRT and the FL DOH Rad Tech Certification Office are not able to confirm exam dates, times, or locations for your examination, nor can they provide driving directions to test centers.

**Missing Your Appointment**

If you fail to keep your appointment or fail to reschedule it as detailed in the next section, your file will close and you will forfeit your examination fee. Neither ARRT nor the FL DOH Rad Tech Certification Office are responsible for appointment time discrepancies between you and the test center.

**Canceling or Rescheduling Your Appointment**

You may cancel or reschedule an appointment up to 24 hours (one business day) prior to the scheduled appointment — either by phoning (800) 632-9055 (leaving a voicemail on an answering machine is not acceptable) or at [www.pearsonvue.com/arrt](http://www.pearsonvue.com/arrt) (be sure to follow the prompts to complete the process). If you make a new appointment, follow up by phoning the Call Center to confirm it. See the "Follow-Up and Confirm Your Exam Appointment" box at left. Pearson VUE will immediately send you an email confirmation each time an appointment is made, changed, or canceled. If you do not receive a confirmation, contact Pearson VUE immediately to confirm the transaction. Pearson VUE charges a \$10 fee for exam appointments that are canceled or rescheduled; fees will be collected by credit card payment (American Express, Mastercard, Visa, or Discover) at the time the appointment is canceled or rescheduled. This includes all changes made online or via the Pearson VUE Call Center.

The table on the next page shows that appointments for a given time on the scheduled exam day must be canceled by that same time on the preceding business day:

**Follow-Up and Confirm****Your Exam Appointment**

You are responsible for confirming the date, time, and location of your exam with Pearson VUE. If you don't receive an email confirmation immediately after scheduling, contact the Pearson VUE Call Center to confirm over the phone and request that a duplicate confirmation letter be sent.

This applies to appointments scheduled via the Call Center as well as those scheduled through the Internet.



**Calling to Reschedule?  
— Remember to Cancel**

*Just because you call to reschedule a testing appointment does not necessarily mean that your initial appointment is automatically canceled. And an uncanceled appointment is your responsibility, potentially resulting in forfeiting the application fee.*

*If you call Pearson VUE intending to reschedule a testing appointment, your initial appointment will remain in effect until you formally approve a new appointment date/time. If you can't find an appropriate alternative appointment and plan to call back later, your initial appointment will still be on the books.*

*Play it safe when changing your appointment. Be sure to specifically request that the initial appointment is canceled. You will receive an email confirmation immediately after your cancellation request is processed.*

| Scheduled Exam Day | Cancel/Change Deadline (same time as appointment) |
|--------------------|---|
| Monday             | Friday of the preceding week                      |
| Tuesday            | Monday of the same week                           |
| Wednesday          | Tuesday of the same week                          |
| Thursday           | Wednesday of the same week                        |
| Friday             | Thursday of the same week                         |
| Saturday           | Friday of the same week                           |



**Exception**

*Due to call center hours, if your appointment is in a time zone ahead of Central (i.e., Eastern or further east), you must cancel any 8 a.m. appointment by 7 p.m. CT two days in advance.*

For example, if your exam is scheduled for 9 a.m. on Monday, you must call by 9 a.m. on Friday to cancel your appointment. VUE will follow-up with a confirmation email detailing your cancellation or appointment change information.

NOTE: National holidays and weekends are not considered business days.

If you fail to appear for your scheduled appointment and do not reschedule through the procedure above, you will forfeit your examination fee. To receive a new eligibility letter, you must contact the FL DOH Rad Tech Certification Office. Neither ARRT nor the FL DOH Rad Tech Certification Office are responsible for appointment errors.

ARRT does not grant exceptions for missed appointments under any circumstance.

# Exam Administration Day

Here's a preview of what you'll encounter when you open the test center's front door on the day of your state licensing exam appointment.



**What to Expect on Exam Day**

*Watch the "What to Expect on Exam Day" video at [www.arrt.org/video-library](http://www.arrt.org/video-library) to familiarize yourself with the process. NOTE: You will not see a preliminary exam score as depicted in the video. (See page 17 for score information.)*

## Test Center Environment

Pearson VUE test centers provide computerized testing for many organizations. Be aware that other exams may be administered in the test center at the same time as ARRT examinations.

Most test centers are located in buildings comprised of several other offices. Waiting areas at the test centers are small. Friends, relatives or children will not be permitted to wait in the test center or to contact you during your examination.

Test center personnel try to maintain a comfortable temperature in the testing rooms. In spite of these efforts, the room may be too cool or too warm for an individual's preference, so dress accordingly. Be aware that outerwear (overcoat, windbreaker, jacket, etc.) is not allowed in the testing room; however, clothing typically worn indoors (sweater, sweatshirt without a hood, blazer, etc.) is allowed.

Keep in mind that there will be other people at the test center taking exams, so typing, coughing and/or people entering and leaving the testing room may be heard. It is impossible to provide a completely noise-free exam environment. If you feel these distractions may be disruptive to your testing, be sure to request earplugs before beginning your exam. Noise reduction headphones can also be provided.

## Acceptable Forms of Identification

**PRIMARY:** Must be government-issued, have **pre-printed name, photo, and signature**, and not be expired.

- Government-issued driver's license
- State ID card
- Passport
- Military ID\*

\*Barcode for signature acceptable with Military IDs only.

**SECONDARY:** Must have **pre-printed name and signature** and not be expired.

- Government-issued IDs (e.g., U.S. social security card)
- Employee ID or work badge
- Bank automated teller machine (ATM) card
- School ID
- Credit card
- Any form of ID on the primary list

**Very Important!** Please note that Permanent Residence Cards ("Green Cards") or any other IDs that do not have your signature **will not be accepted** at the test center as valid primary or secondary identification.

TIP

### Invalid ID? Cancel Appointment

If you arrive at the test center with invalid forms of ID, you will not be admitted to take the test, and you will forfeit your fee.

Your ID is invalid if:

- it is not current (expired), or
- the name on ID varies from that on your ARRT file, or
- there is no signature, or
- it is not an accepted form of ID (see box above).

Cancel your appointment in time and re-schedule when you have acquired two forms of valid ID.

### Follow Procedures

Test center personnel adhere to designated procedures to ensure that their operations meet ARRT criteria for standardized testing. Review the following information before the examination to become familiar with the procedures.

### Arrive Early

Having already confirmed the location of the test center, plan your schedule and route to ensure that you arrive at least 30 minutes before your scheduled appointment, to allow time for check-in procedures. Be sure to allow ample time for your commute, especially if inclement weather is a factor.

If you arrive at the test center 15 minutes after your scheduled appointment, you may be required to forfeit the appointment. If an appointment is forfeited, the test center will report to ARRT your failure to take the examination and your file will close. ARRT does not refund exam fees on forfeited appointments.

If you wish to be assigned a new exam window, you must contact the FL DOH Rad Tech Certification Office for new exam eligibility information.

### Palm Vein Recognition

#### Replaces Fingerprint

As of January 1, 2011, a new biometric procedure was added to the admissions process, replacing the fingerprint process. Called palm vein technology, it scans the veins inside the hand to create a digital template that represents your vein pattern. The pattern reader uses a safe, near-infrared light source, similar to a television remote.

Repeat candidates that have had their digital fingerprint captured at a previous exam will be required to provide a fingerprint match and then enroll in the palm vein recognition system. Exams taken after this initial procedure will use the palm vein system only.

All other steps of the admissions process will remain the same. Each candidate will still be required to bring a valid photo ID and secondary ID, be photographed, and sign a digital signature pad.

Learn more about the process at [www.rrt.org](http://www.rrt.org).

### ID, Photo, Signature, Palm Vein Recognition (PVR)

When you arrive at the test center, you will be required to show two forms of identification, both of which show your signature and your pre-printed name as it appears on your CSR. One of the IDs must be a current official government-issued photo ID. See above for examples of the two types of IDs required.

Your name on your government-issued ID must be the same as that on record with ARRT, as reflected on your most recent CSR. Your ID may contain your full middle name as long as the middle initial on your CSR matches the first letter of your middle name. If your name has a cultural variation, ensure that the same variation appears on the CSR and both IDs.

If you arrive without proper ID or with discrepancies in your name listed on the IDs, you will not be admitted to the test center. You will not be allowed to re-schedule your exam appointment and will forfeit your examination fee. If you are admitted with questionable ID, you may have your score canceled following investigation by ARRT.

Upon checking in, you will be asked to provide a digital signature, which constitutes a) your consent for ARRT and/or Pearson VUE to retain and transmit personal data and exam responses; and b) your agreement to abide by the ARRT Rules Agreement, which will be presented to you prior to your exam.

You will also have your palm vein scanned and be photographed. If you leave the testing area for any reason, your palm will be scanned upon leaving and again before re-entering.

The palm-vein information and photo are for authentication purposes only. The information is kept confidential and not shared with any organization.

## Assignment to Testing Station

Test center personnel will give you a short orientation, provide you with a copy of the ARRT Rules Agreement (see Appendix L) to read, and then escort you to an assigned workstation. You must remain in your assigned seat during the examination, except when authorized to leave by a test center staff member.

You will be required to keep all personal items in a secure locker. Don't wear jewelry that may be noisy or disruptive in the testing room. You will be asked to remove jewelry that is wider than  $\frac{1}{4}$ " as such items can pose a threat to exam security. If you bring a phone or other electronic device, turn off the device and store it in your locker. You may not access any electronic device until you have completed your exam and are ready to leave the test center. You cannot access items placed in a secure locker or anywhere else in the test center building for the duration of your exam unless you receive written pre-approval from ARRT. This includes breaks. Test centers assume no responsibility for candidates' personal belongings.

If you need to leave the testing room for personal reasons, you must first raise your hand to get test center staff's permission. No additional time is allowed to make up for lost time due to this reason. Test center staff is required to file an incident report with ARRT on any candidate that leaves the testing room for more than 10 minutes.

Test center personnel are not trained to answer specific questions related to ARRT examination content.

## Calculators and Notes

Personal calculators are not permitted. Both scientific and basic four-function calculators are provided on the computer, or you may request a basic four-function calculator from test center personnel. Appendix I presents facsimiles of the computer calculator. Examples of calculators are also presented in the tutorial at the beginning of the exam.

Test center personnel will provide an erasable note board and pen, which may be replaced as needed during testing but may not be removed from the testing room at any time. Do not use the note board until after responding to the non-disclosure agreement and you may not hold your note board up to the screen when responding to questions. Scratch paper, pens, or pencils are not allowed in the testing room.

## Requesting Assistance

Raise your hand to notify test center personnel if:

- you need assistance adjusting the computer screen's brightness or contrast;
- you would like a hand-held calculator;
- you need earplugs;
- an image appears too large to be fully viewed;
- you suspect a problem with the computer;
- you need another erasable note board;
- you need a break;
- you have completed your exam; or
- you need a staff member for any other reason.

## Exam Timing

Time allowed for completing an examination is based on the number of questions on the exam. The table on the next page indicates how much time has been allocated for each of the different examinations. The column labeled "exam time" indicates how much time has been allocated to answer the questions on the examination. The column labeled "total time" adds 30 minutes to testing time to allow the candidate the 18 minutes designated for the tutorial, followed by two minutes to respond to the non-disclosure agreement and 10 minutes designated for the survey after the examination has been completed. This extra 30 minutes is for completion of the tutorial and survey and cannot be used to answer examination questions. Voluntary breaks are subtracted from the allowed testing time; that is, the clock is not stopped during voluntary breaks.

TIP

### **Don't Bring Yours... Calculators Provided**

*Personal calculators aren't permitted, so don't even bring one to the test center. You can use theirs.*

TIP

### **Pace Yourself...**

*It's important to use your time economically. Time remaining is displayed in the upper right corner of the computer screen. See "Taking the Exam" on page 16 and Appendix I for more information.*

ARRT recommends that you complete the tutorial to familiarize yourself with the testing program and the online calculators. You must also click "A" for the non-disclosure agreement (see box on page 15), which appears after the tutorial and before starting your exam.

| <b>Discipline</b>           | <b>Exam Time</b> | <b>Total Time</b> |
|-----------------------------|------------------|-------------------|
| Radiography                 | 3.5 hours        | 4.0 hours         |
| Nuclear Medicine Technology | 3.5 hours        | 4.0 hours         |
| Radiation Therapy           | 3.5 hours        | 4.0 hours         |

#### Limited Scope of Practice in Radiography

Basic X-Ray Machine (Core, Chest, Extremities)

Basic X-Ray Machine/Podiatry (Core, Podiatric)

**Time allowed for Limited Scope of Practice in Radiography.** Each module is separately timed. The amount of time is determined by the number of questions in a module, at a rate of one minute per question. For example, the Core module has 115 total questions, so you have up to 115 minutes to complete it. The Chest module has 25 total questions, and 25 minutes are allowed for completion. It is important to pace yourself so that you complete each module within the allotted time. NOTE: Breaks are not scheduled between modules. That is, the clock will continue ticking after completing one module and moving to the next module.

**Which Modules.** The computer will present only those modules that were assigned to you by your state licensing agency. Those same modules are listed on your CSR. If you feel you have not been assigned the correct modules, contact the FL DOH Rad Tech Certification Office — not ARRT — immediately and before scheduling your appointment.

**Review Session.** The computer requires that you answer every question. If you are unsure of an answer to a question, you can “mark” the question and come back to it later. After you have answered all questions in a module, a review screen allows you to go back to any question you marked. You can change answers during the review. When done reviewing questions, you can end the module. Extra time is not given for the review session; it must be completed during the time allowed for that specific module. A sample review screen is printed in Appendix J.

**End Module/End Exam.** Once you end the review session, the module ends. You will not be able to go back and review questions in that module. At this point, one of two things happen: 1) if you have additional modules to complete, the next module will appear; 2) if you do not have additional modules to complete, the exam ends.

**Pilot Questions.** Pilot questions are unscored questions embedded in the test. ARRT uses data from these pilot questions to evaluate new questions. This is a cost-effective way to develop test materials for future candidates, just as past candidates assisted in piloting questions for today. These questions are not identified as pilot questions, and they appear just like any other question on the test. Up to 20 percent of your test may be unscored pilot questions, and ARRT has allotted extra time for you to complete them. Your answers to these questions will not affect test scores.

**Copyrighted Exam Material.** Law prohibits any attempt to reproduce all or part of the examinations. Anyone caught removing exam materials from the test center, whether by physical removal or by reproducing materials from memory, will be prosecuted to the full extent of the law and will be permanently barred from future examinations.

## Test Center Misconduct and Score Cancellation

Numerous security measures are enforced during the exam administration to ensure the integrity of ARRT exams. Be aware that you will be observed at all times while completing the exam. This includes direct observation by test center staff, as well as video and audio recording of the testing session.

### Zero Tolerance Policy

ARRT has a zero tolerance policy regarding possession of cell phones and other electronic devices in the test center, as well as candidates leaving the test center building prior to completing the examination and attempting to re-enter the test center. Automatic score cancellation will result for any candidate violating this policy.

1. Under no circumstances are candidates permitted to access cell phones or any other type of electronic device after check-in at the test center. Test center personnel are instructed to dismiss any candidate found in possession of an electronic device after the candidate has completed the check-in procedures. This includes candidates on breaks.

Such electronic devices include, but are not limited to:

- cellular phones;
- media players;
- compact disc players or any other electronic communication/recording/listening device;
- removable storage devices;
- personal digital assistants (PDAs);
- calculator or computing watches;
- scan pens;
- laptop computers, tablets or any computer device; and
- photographic devices.

If a candidate is found possessing, or otherwise having access to, a cell phone or any other type of electronic device during the administration of their exam, the candidate will not be allowed to continue testing and the test center administrator will file an incident report. Possession of a cell phone or any other type of electronic listening device after check-in will result in automatic score cancellation.

2. If test center staff observes a candidate leaving the test center building and re-entering the test center prior to completing the exam, the candidate will not be allowed to continue testing and the test center administrator will file an incident report. Leaving the test center building and attempting to re-enter the test center will result in automatic score cancellation.
3. Candidates should not bring papers, pamphlets, books, notebooks or study guides into the test center. If you bring these items, they must remain in your locker for the duration of your exam. If you are found in possession of, or otherwise having access to, any prohibited item during the administration of your exam, you will not be allowed to continue testing and the test center administrator will file an incident report. This will also result in automatic score cancellation.
4. For any candidate demonstrating misconduct or irregular behavior during or in connection with the examination — as evidenced by observation, statistical analysis of exam responses or otherwise — the ARRT will withhold examination scores and may revoke or suspend a certificate, deny or reject an application for renewal of certification and registration, censure or take any other appropriate action. This includes permanently barring the candidate from all future examinations, terminating candidate participation in the exam and invalidating the results of that exam and any prior exam.

Examples of misconduct or irregular behavior include, but are not limited to:

- Removing items from a secured locker without prior authorization
- Giving or receiving unauthorized help
- Attempting to take the examination for someone else
- Failing to follow test center staff instructions
- Tampering with the operation of the computer or attempting to use it for any function other than completing the examination
- Attempting to remove exam content (in any format) from the test center
- Creating a disturbance of any kind
- Accessing notes, books, study guides or unauthorized electronic devices

If found to be in violation of this policy, you may find yourself part of an ARRT ethics investigation, or even a federal court lawsuit for copyright infringement and/or breach of contract.

## **Non-Disclosure Agreement**

*After the tutorial, a non-disclosure agreement will appear on the computer screen. You must accept the terms of the agreement in order to proceed with the exam. By accepting these terms, you agree not to disclose exam questions in any form or remove them from the test center. You have two minutes to indicate your acceptance of the agreement. If you do not respond within two minutes, the exam will end and you will have to submit a re-application form and fee to obtain a new exam window.*

*The agreement states: “This exam is confidential and is protected by copyright law. You are expressly prohibited from disclosing, publishing, reproducing, or transmitting this exam, in whole or in part, in any form or by any means, oral or written, electronic or mechanical, for any purpose.”*

*The screen will instruct you to click the “A” (for Accept) button to symbolize your signature and to accept the terms. Selecting “A” will allow you to continue with the exam. If you do not accept these terms, click “N” (for Not Accept) to let test center staff know that you are through with the exam. If you click “N” but later decide to examine at a future date, you will need to submit a re-application form and fee.*

*Learn more about the non-disclosure agreement in the “What to Expect on Exam Day” video at [www.arry.org/video-library](http://www.arry.org/video-library).*

TIP

### **Severe Weather Looming?**

*If you anticipate severe weather and your appointment is more than 24 hours out, consider rescheduling to avoid transportation hassles. If you miss your appointment due to weather and the test center was open, you will forfeit your exam fee and you will need to contact the FL DOH Rad Tech Certification Office for a new eligibility period.*

## **What if the Test Center is Closed?**

If you are unsure whether a test center is closed because of inclement weather or some other factor, phone Pearson VUE's Call Center at (800) 632-9055. If the test center is open, it is your responsibility to keep your appointment. If it is closed, you will be given the opportunity to reschedule your appointment. In the event of a test center closing, Pearson VUE will contact you via the email address you provided during scheduling to reschedule your exam appointment. You may also call Pearson VUE to reschedule your exam.

## **Taking the Exam**

### **Order of Questions**

ARRT examinations present questions in random order, which is consistent with the purposes of education and evaluation. When an individual learns an important concept, the intent is that he or she will take that knowledge beyond a specific context or environment and generalize that knowledge to the practice setting.

### **Question Format**

Most exam items are standard multiple-choice with one best answer. ARRT is also introducing new formats on a limited basis. Some items may require that you select multiple answers from a list or use the mouse to sort a list of options into a particular order. A few items may require that you identify anatomic structures on an image by placing the mouse arrow (cursor) over the correct location on the screen and clicking. Others may require you to answer a multiple-choice question after viewing a short video clip. Appendix I provides additional information on exam item formats.

### **Selecting Answers**

An answer must be recorded for a question before the computer allows display of the next question. You may flag questions for later review if you are unsure of the answer. For further information, refer to Appendix I.

### **Pacing**

It's important to use your time economically. Time remaining is displayed in the upper right corner of the computer screen. If a question is difficult, guess at the answer, flag the question for review, and go on to the next question. When you have finished the examination and there is still time left, go back to the questions that you flagged and review them by clicking on the "Review Flagged" button (see details in Appendix I).

### **Guessing**

Exam scores are based upon the total number of correct answers. Therefore, it is to your advantage to answer every question, even if that means selecting an answer of which you are not sure. You must indicate some response to each question before the computer will proceed to the next question.

## **Candidate Comments**

You may comment on a specific question at the time you answer the question by clicking on the "Comment" button at the top of that page. No additional testing time is allowed during the exam for making comments on questions.

You may comment on your test center experience in the evaluation survey at the end of your exam.

## **Leaving the Test Center**

When you are finished with the examination and evaluation survey, raise your hand and test center staff will collect the erasable note board before dismissing you. Do not leave your seat until you have been dismissed. You may not remove note boards from the testing room. Your palm will be scanned again before leaving the test center.

## Appeals of Exam Administration

ARRT makes every effort to assure that examinations are fairly administered in a comfortable and safe environment.

On rare occasions, candidates may encounter technical difficulties at the test center. If you experience a technical difficulty, notify the test center administrator immediately. Test center personnel will make every effort to correct any difficulties as quickly as possible.

Should the test center experience a loss of power, back-up systems are in place, so every reasonable effort will be made to retrieve testing data. Once power is restored, you will be able to continue your testing session from the point where you were interrupted. If you are unable to continue the testing session due to severe technical difficulties, reasonable accommodations will be made, including re-scheduling of an exam appointment. ARRT will evaluate individual requests for re-scheduling at no cost.

If you believe that your examination was administered in a manner that substantially deviated from normal testing procedures, you may request a review of the procedures. If you experience a problem, verify with the test center administrator before you leave the test center that they will file a report regarding your issue.

If you wish to request a review, submit a completed Appeal Request Form (at [www.staterhc.org](http://www.staterhc.org)) detailing the specific nature of the alleged deviation from normal testing procedures. Be sure to include your email address, along with the examination discipline, administration date and test center location.

Because ARRT will investigate complaints only if they are received before your results have been released, you have only two days to submit the request. You may fax the appeal form to (651) 681-3295.

If ARRT finds that any such deviation unfairly interfered with your ability to complete the exam to the best of your ability in the allotted time, your original score will be canceled and you will be allowed to retake the examination at no cost. Under no circumstances will your score be adjusted based upon the findings of the review.

## After the Examination

After the examination, all exam data is returned to ARRT, where scoring and analysis is completed. ARRT follows strict procedures to ensure accuracy of scoring.

### Cancellation of Scores

ARRT may withhold or cancel scores if there is evidence that the security of the examination has been compromised. Such action may be necessary even in the absence of evidence indicating that a candidate was knowingly involved in the compromising activities. ARRT expects candidates to cooperate in any investigation. Once scores are cancelled, they are not available for reporting at a later date.

Some scores may be rendered invalid because of circumstances beyond a candidate's control, such as technical difficulties. ARRT investigates each of these situations. When this results in a cancellation of scores, ARRT arranges for a makeup administration of the exam at no additional cost.

### Score Reporting

You will not see a preliminary score at the end of your exam at the test center. ARRT does not release examination scores to state candidates. Your score information is forwarded to the FL DOH Rad Tech Certification Office which, in turn, determines the pass/fail status and posts the scores on their website at [www.floridahealth.gov/licensing-and-regulation/radiologic-technology/exam-grade-report/index.html](http://www.floridahealth.gov/licensing-and-regulation/radiologic-technology/exam-grade-report/index.html). The FL DOH Rad Tech Certification Office no longer mails paper statements. Contact the FL DOH Rad Tech Certification Office (not ARRT) if your results have not been posted to their website within four weeks.

**TIP****How/When Will I Get Exam Results?**

*Examination results are sent to the FL DOH Rad Tech Certification Office for final pass/fail determination and will appear on the FL-DOH website. Please allow up to four weeks for processing, then contact the FL DOH Rad Tech Certification Office — not ARRT — for information on your exam results. (See inside cover of handbook for contact information.)*

**State Attempts****Count Against Three-Attempt Limit**

*As of January 1, 2010, if you attempt the exam as a state candidate instead of taking the exam for ARRT certification and registration, you will have the state attempt count as an attempt toward future ARRT certification and registration attempts.*

**Interpreting Scores**

ARRT uses “scaled scores” to report examination results for the Radiography, Nuclear Medicine Technology, and Radiation Therapy exams. Scaled scores are more meaningful than raw scores (i.e., number or percentage correct) because they take into account the difficulty of a particular exam compared to other forms of the same exam. Therefore, a scaled score of 75 represents the same level of exam performance, regardless of what examination form was administered.

Total scores are reported on a scale that ranges from 1 to 99. The total scaled score does not equal the number or percentage of questions answered correctly. A total scaled score of 75 is required to pass the exam. The number of correct answers required to achieve a score of 75 was determined through a standard-setting (or passing score) study. ARRT and panels of consultants periodically review the passing score to assure its validity.

Performance on each section of the exam is also reported using scaled scores. These section scores provide information to candidates regarding their strengths and weaknesses in particular content categories. Pass/fail decisions are not based on individual sections of the exam. Section scores can range from 1 to 10 and are reported in one-tenth point intervals (e.g., 8.1, 8.6). Section scores are intentionally placed on a narrower scale because they are based on fewer exam questions. Therefore, section scores are not as reliable as the total scaled score and should be interpreted with some discretion.

Limited Scope of Practice in Radiography exam results are reported to the state as number correct for each module assigned. The FL DOH Rad Tech Certification Office determines pass/fail scores based on the number correct.

**Appeals of Exam Scoring**

ARRT employs several quality control procedures to ensure that all examinations are scored with complete accuracy. However, you may request a review of the accuracy of the scoring process if you feel an error has occurred.

If you wish a review of scoring, you must complete the Appeal Request form located at StateRHC.org within 30 days of your exam date — detailing the specific reason a scoring error is suspected. Requests must be accompanied by a \$25 fee, payable to ARRT.

ARRT will review your responses to each question, compare those responses to the answer key, and recalculate raw scores. Final passing scores are determined by the FL DOH Rad Tech Certification Office.

ARRT will report its findings to you within 30 days of receiving the written request. If ARRT finds evidence of any scoring error, it will cancel your original score and notify the FL DOH Rad Tech Certification Office of the corrected score.

**Re-examination**

If you fail the examination, do not appear as scheduled, answer "no" or do not respond to the non-disclosure agreement, allow your 90-day exam window to expire, or you were turned away due to invalid IDs, you should contact the FL DOH Rad Tech Certification Office for information on your examination eligibility. Once it has been determined you are eligible for re-examination, the FL DOH Rad Tech Certification Office will send you a confirmation letter with instructions on how to pay ARRT your new exam fee. Once ARRT processes your new exam fee, a new handbook and CSR indicating your new 90-day exam window will be mailed to you.

# Appendices

|    |  |       |
|----|--|-------|
| A. | Content Specifications — Radiography .....   | 21-34 |
| B. | Content Specifications — Nuclear Medicine Technology..   | 35-45 |
| C. | Content Specifications — Radiation Therapy .....   | 46-54 |
| D. | Content Specifications — Limited Scope of Practice<br>in Radiography .....                     | 55-72 |
| E. | Sample Questions for the Examination for the Limited Scope of<br>Practice in Radiography ..... | 73    |
| F. | Test Centers .....   | 74    |
| G. | Sample Candidate Status Report .....   | 75-76 |
| H. | Sample Limited Scope Candidate Status Report.....  | 77-78 |
| I. | Computer-Based Testing Overview.....   | 79-82 |
| J. | Limited Scope Computer-Based Testing Overview .....  | 83-86 |
| K. | Potential Exam Disclosure Scenarios .....  | 87    |
| L. | ARRT Rules Agreement .....   | 88    |
| M. | Checklist .....  | 89    |

This page intentionally left blank.



# Radiography Examination

The purpose of *The American Registry of Radiologic Technologists® (ARRT®) Radiography Examination* is to assess the knowledge and cognitive skills underlying the intelligent performance of the tasks typically required of radiographers. Using a nationwide survey, the ARRT periodically conducts a practice analysis to develop a task inventory which delineates or lists the job responsibilities typically required of radiographers.<sup>1</sup> An advisory committee then determines the knowledge and cognitive skills needed to perform the tasks on the task inventory and these are organized into the content categories within this document. The document is used to develop the examination. The results of the most recent practice analysis have been applied to this document. Every content category can be linked to one or more activities on the task inventory. The complete task inventory is available at [arrt.org](http://arrt.org).

The following table presents the four major content categories covered on the examination, and indicates the number of test questions in each category. The remaining pages list the specific topics addressed within each category, with the approximate number of test questions allocated to each topic appearing in parentheses.

This document is not intended to serve as a curriculum guide. Although ARRT programs for certification and registration and educational programs may have related purposes, their functions are clearly different. Educational programs are generally broader in scope and address the subject matter that is included in these content specifications, but do not limit themselves to only this content.

| Content Category                                | Number of Scored Questions <sup>2</sup> |
|---|---|
| Patient Care                                    | 33                                      |
| Patient Interactions and Management             |   |
| Safety  | 53                                      |
| Radiation Physics and Radiobiology <sup>3</sup> |   |
| Radiation Protection                            |   |
| Image Production                                | 50                                      |
| Image Acquisition and Technical Evaluation      |   |
| Equipment Operation and Quality Assurance       |   |
| Procedures                                      | 64                                      |
| Head, Spine and Pelvis Procedures               |   |
| Thorax and Abdomen Procedures                   |   |
| Extremity Procedures                            |   |
| <b>Total</b>                                    | <b>200</b>                              |

<sup>1</sup> A special debt of gratitude is due to the hundreds of professionals participating in this project as committee members, survey respondents and reviewers.

<sup>2</sup> Each exam includes an additional 20 unscored (pilot) questions.

<sup>3</sup> SI units will become the primary (principle) units of radiation measurement used on the radiography examination in 2017.



## Patient Care (33)

### 1. Patient Interactions and Management (33)

#### A. Ethical and Legal Aspects

1. patient's rights
  - a. informed consent (\*e.g., written, oral, implied)
  - b. confidentiality (HIPAA)
  - c. American Hospital Association (AHA) Patient Care Partnership (Patient's Bill of Rights)
    1. privacy
    2. extent of care (e.g., DNR)
    3. access to information
    4. living will, health care proxy, advanced directives
    5. research participation
  - d. legal issues
    - a. verification (e.g., patient identification, compare order to clinical indication)
    - b. common terminology (e.g., battery, negligence, malpractice, beneficence)
    - c. legal doctrines (e.g., respondeat superior, res ipsa loquitur)
    - d. restraints versus immobilization
    - e. manipulation of electronic data (e.g., exposure indicator, processing algorithm, brightness and contrast, cropping or masking off anatomy)
2. ARRT Standards of Ethics

#### B. Interpersonal Communication

1. modes of communication
  - a. verbal/written
  - b. nonverbal (e.g., eye contact, touching)
2. challenges in communication
  - a. interactions with others
    1. language barriers
    2. cultural and social factors
    3. physical or sensory impairments
    4. age
    5. emotional status, acceptance of condition
  - b. explanation of medical terms
  - c. strategies to improve understanding
3. patient education

- a. explanation of current procedure (e.g., purpose, exam length)
- b. verify informed consent when necessary
- c. pre- and post-examination instructions (e.g., preparation, diet, medications and discharge instructions)
- d. respond to inquiries about other imaging modalities (e.g., CT, MRI, mammography, sonography, nuclear medicine, bone densitometry regarding dose differences, types of radiation, patient preps)

#### C. Physical Assistance and Monitoring

1. patient transfer and movement
  - a. body mechanics (e.g., balance, alignment, movement)
  - b. patient transfer techniques
2. assisting patients with medical equipment
  - a. infusion catheters and pumps
  - b. oxygen delivery systems
  - c. other (e.g., nasogastric tubes, urinary catheters, tracheostomy tubes)
3. routine monitoring
  - a. vital signs
  - b. physical signs and symptoms (e.g., motor control, severity of injury)
  - c. fall prevention
  - d. documentation

#### D. Medical Emergencies

1. allergic reactions (e.g., contrast media, latex)
2. cardiac or respiratory arrest (e.g., CPR)
3. physical injury or trauma
4. other medical disorders (e.g., seizures, diabetic reactions)

\*The abbreviation "e.g.," is used to indicate that examples are listed in parentheses, but that it is not a complete list of all possibilities.

(Patient Care continues on the following page.)



## Patient Care (continued)

### E. Infection Control

1. cycle of infection
  - a. pathogen
  - b. reservoir
  - c. portal of exit
  - d. mode of transmission
    1. direct
      - a. droplet
      - b. direct contact
    2. indirect
      - a. airborne
      - b. vehicle borne–fomite
      - c. vector borne–mechanical or biological
  - e. portal of entry
  - f. susceptible host
2. asepsis
  - a. equipment disinfection
  - b. equipment sterilization
  - c. medical aseptic technique
  - d. sterile technique
3. CDC Standard Precautions
  - a. hand hygiene
  - b. use of personal protective equipment (e.g., gloves, gowns, masks)
  - c. safe injection practices
  - d. safe handling of contaminated equipment/surfaces
  - e. disposal of contaminated materials
    1. linens
    2. needles
    3. patient supplies
    4. blood and body fluids
4. transmission-based precautions
  - a. contact
  - b. droplet
  - c. airborne
5. additional precautions
  - a. neutropenic precautions (reverse isolation)
  - b. healthcare associated (nosocomial) infections

### F. Handling and Disposal of Toxic or Hazardous Material

1. types of materials
  - a. chemicals
  - b. chemotherapy
2. safety data sheet (e.g., material safety data sheets)

### G. Pharmacology

1. patient history
  - a. medication reconciliation (current medications)
  - b. premedications
  - c. contraindications
  - d. scheduling and sequencing examinations
2. administration
  - a. routes (e.g., IV, oral)
  - b. supplies (e.g., enema kits, needles)
3. venipuncture
  - a. venous anatomy
  - b. supplies
  - c. procedural technique
4. contrast media types and properties (e.g., iodinated, water soluble, barium, ionic versus non-ionic)
5. appropriateness of contrast media to exam
  - a. patient condition (e.g., perforated bowel)
  - b. patient age and weight
  - c. laboratory values (e.g., BUN, creatinine, GFR)
6. complications/reactions
  - a. local effects (e.g., extravasation/infiltration, phlebitis)
  - b. systemic effects
    1. mild
    2. moderate
    3. severe
  - c. emergency medications
  - d. radiographer's response and documentation



## Safety (53)

### 1. Radiation Physics and Radiobiology (22)

- A. Principles of Radiation Physics
  - 1. x-ray production
    - a. source of free electrons (e.g., thermionic emission)
    - b. acceleration of electrons
    - c. focusing of electrons
    - d. deceleration of electrons
  - 2. target interactions
    - a. bremsstrahlung
    - b. characteristic
  - 3. x-ray beam
    - a. frequency and wavelength
    - b. beam characteristics
      - 1. quality
      - 2. quantity
      - 3. primary versus remnant (exit)
    - c. inverse square law
    - d. fundamental properties (e.g., travel in straight lines, ionize matter)
  - 4. photon interactions with matter
    - a. Compton effect
    - b. photoelectric absorption
    - c. coherent (classical) scatter
    - d. attenuation by various tissues
      - 1. thickness of body part
      - 2. type of tissue (atomic number)
- B. Biological Aspects of Radiation
  - 1. SI units of measurement (NCRP #160)
    - a. absorbed dose (Gy)
    - b. dose equivalent (Sv)
    - c. exposure (C/kg)
    - d. effective dose (Sv)
    - e. air kerma (Gy)
  - 2. radiosensitivity
    - a. dose-response relationships
    - b. relative tissue radiosensitivities (e.g., LET, RBE)
    - c. cell survival and recovery (LD<sub>50</sub>)
    - d. oxygen effect
  - 3. somatic effects
    - a. short-term versus long-term effects
    - b. acute versus chronic effects
    - c. carcinogenesis
    - d. organ and tissue response (e.g., eye, thyroid, breast, bone marrow, skin, gonadal)
  - 4. acute radiation syndromes
    - a. hemopoietic
    - b. gastrointestinal (GI)
    - c. central nervous system (CNS)
  - 5. embryonic and fetal risks
  - 6. genetic impact
    - a. genetically significant dose
    - b. goals of gonadal shielding

(Safety continues on the following page.)

**Safety (continued)****2. Radiation Protection (31)**

- |  |  |
|--|--|
| <p><b>A. Minimizing Patient Exposure</b></p> <ol style="list-style-type: none"> <li>1. exposure factors             <ol style="list-style-type: none"> <li>a. kVp</li> <li>b. mAs</li> <li>c. automatic exposure control (AEC)</li> </ol> </li> <li>2. shielding             <ol style="list-style-type: none"> <li>a. rationale for use</li> <li>b. types</li> <li>c. placement</li> </ol> </li> <li>3. beam restriction             <ol style="list-style-type: none"> <li>a. purpose of primary beam restriction</li> <li>b. types (e.g., collimators)</li> </ol> </li> <li>4. filtration             <ol style="list-style-type: none"> <li>a. effect on skin and organ exposure</li> <li>b. effect on average beam energy</li> <li>c. NCRP recommendations (NCRP #102, minimum filtration in useful beam)</li> </ol> </li> <li>5. patient considerations             <ol style="list-style-type: none"> <li>a. positioning</li> <li>b. communication</li> <li>c. pediatric</li> <li>d. morbid obesity</li> </ol> </li> <li>6. radiographic dose documentation</li> <li>7. image receptors</li> <li>8. grids</li> <li>9. fluoroscopy             <ol style="list-style-type: none"> <li>a. pulsed</li> <li>b. exposure factors</li> <li>c. grids</li> <li>d. positioning</li> <li>e. fluoroscopy time</li> <li>f. automatic brightness control (ABC) or automatic exposure rate control (AERC)</li> <li>g. receptor positioning</li> <li>h. magnification mode</li> <li>i. air kerma display</li> <li>j. last image hold</li> <li>k. dose or time documentation</li> <li>l. minimum source-to-skin distance (21 CFR)</li> </ol> </li> <li>10. dose area product (DAP) meter</li> </ol> | <p><b>B. Personnel Protection (ALARA)*</b></p> <ol style="list-style-type: none"> <li>1. sources of radiation exposure             <ol style="list-style-type: none"> <li>a. primary x-ray beam</li> <li>b. secondary radiation                 <ol style="list-style-type: none"> <li>1. scatter</li> <li>2. leakage</li> </ol> </li> <li>c. patient as source</li> </ol> </li> <li>2. basic methods of protection             <ol style="list-style-type: none"> <li>a. time</li> <li>b. distance</li> <li>c. shielding</li> </ol> </li> <li>3. protective devices             <ol style="list-style-type: none"> <li>a. types</li> <li>b. attenuation properties</li> <li>c. minimum lead equivalent (NCRP #102)</li> </ol> </li> <li>4. special considerations             <ol style="list-style-type: none"> <li>a. mobile units</li> <li>b. fluoroscopy                 <ol style="list-style-type: none"> <li>1. protective drapes</li> <li>2. protective Bucky slot cover</li> <li>3. cumulative timer</li> <li>4. remote-controlled fluoroscopy</li> </ol> </li> <li>c. guidelines for fluoroscopy and mobile units (NCRP #102, 21 CFR)                 <ol style="list-style-type: none"> <li>1. fluoroscopy exposure rates (normal and high-level control)</li> <li>2. exposure switch guidelines</li> </ol> </li> </ol> </li> <li>5. radiation exposure and monitoring             <ol style="list-style-type: none"> <li>a. dosimeters                 <ol style="list-style-type: none"> <li>1. types</li> <li>2. proper use</li> </ol> </li> <li>b. NCRP recommendations for personnel monitoring (NCRP #116)                 <ol style="list-style-type: none"> <li>1. occupational exposure</li> <li>2. public exposure</li> <li>3. embryo/fetus exposure</li> <li>4. dose equivalent limits</li> <li>5. evaluation and maintenance of personnel dosimetry records</li> </ol> </li> </ol> </li> <li>6. handling and disposal of radioactive material</li> </ol> |
|--|--|

\* (August 24, 2016) Note: Although it is the radiographer's responsibility to apply radiation protection principles to minimize bioeffects for both patients and personnel, the ALARA concept is specific to personnel protection and is listed only for that section.



## Image Production (50)

### 1. Image Acquisition and Technical Evaluation (21)

#### A. Selection of Technical Factors Affecting Radiographic Quality

Refer to *Attachment C* to clarify terms that may occur on the exam.

(X indicates topics covered on the examination.)

|                                      | 1.<br>Receptor<br>Exposure | 2.<br>Contrast | 3.<br>Spatial<br>Resolution | 4.<br>Distortion |
|--------------------------------------|----------------------------|----------------|-----------------------------|------------------|
| a. mAs                               | X                          |                |                             |                  |
| b. kVp                               | X                          | X              |                             |                  |
| c. OID                               |                            | X (air gap)    | X                           | X                |
| d. SID                               | X                          |                | X                           | X                |
| e. focal spot size                   |                            |                | X                           |                  |
| f. grids*                            | X                          | X              |                             |                  |
| g. tube filtration                   | X                          | X              |                             |                  |
| h. beam restriction                  | X                          | X              |                             |                  |
| i. motion                            |                            |                | X                           |                  |
| j. anode heel effect                 | X                          |                |                             |                  |
| k. patient factors (size, pathology) | X                          | X              | X                           | X                |
| l. angle (tube, part, or receptor)   |                            |                | X                           | X                |

\* Includes conversion factors for grids

#### B. Technique Charts

1. anatomically programmed technique
2. caliper measurement
3. fixed versus variable kVp
4. special considerations
  - a. casts
  - b. pathologic factors
  - c. age (e.g., pediatric, geriatric)
  - d. body mass index (BMI)
  - e. contrast media

#### C. Automatic Exposure Control (AEC)

1. effects of changing exposure factors on radiographic quality
2. detector selection
3. anatomic alignment
4. exposure adjustment (e.g., density, +1 or -1)

#### D. Digital Imaging Characteristics

1. spatial resolution (equipment related)
  - a. pixel characteristics (e.g., size, pitch)
  - b. detector element (DEL) (e.g., size, pitch, fill factor)

#### c. matrix size

#### d. sampling frequency

#### 2. contrast resolution (equipment related)

#### a. bit depth

#### b. modulation transfer function (MTF)

#### c. detective quantum efficiency (DQE)

#### 3. image signal (exposure related)

#### a. dynamic range

#### b. quantum noise (quantum mottle)

#### c. signal to noise ratio (SNR)

#### d. contrast to noise ratio (CNR)

#### E. Image Identification

#### 1. methods (e.g., radiographic, electronic)

#### 2. legal considerations

(e.g., patient data, examination data)

(Image Production continues on the following page.)



## Image Production (continued)

### 2. Equipment Operation and Quality Assurance (29)

- A. Imaging Equipment
1. components of radiographic unit (fixed or mobile)
    - a. operating console
    - b. x-ray tube construction
      1. electron source
      2. target materials
      3. induction motor
    - c. automatic exposure control (AEC)
      1. radiation detectors
      2. back-up timer
      3. exposure adjustment (e.g., density, +1 or –1)
      4. minimum response time
    - d. manual exposure controls
    - e. beam restriction
  2. x-ray generator, transformers and rectification system
    - a. basic principles
    - b. phase, pulse and frequency
    - c. tube loading
  3. components of fluoroscopic unit (fixed or mobile)
    - a. image receptors
      1. image intensifier
      2. flat panel
    - b. viewing systems
    - c. recording systems
    - d. automatic brightness control (ABC) or automatic exposure rate control (AERC)
    - e. magnification mode
    - f. table
  4. components of digital imaging
    - a. CR components
      1. plate (e.g., photo-stimulable phosphor (PSP))
      2. plate reader
    - b. DR image receptors
      1. flat panel
      2. charge coupled device (CCD)
      3. complementary metal oxide semiconductor (CMOS)
  5. accessories
    - a. stationary grids
    - b. Bucky assembly
    - c. compensating filters
- B. Image Processing and Display
1. raw data (pre-processing)
    - a. analog-to-digital converter (ADC)
    - b. quantization
    - c. corrections (e.g., rescaling, flat fielding, dead pixel correction)
    - d. histogram
  2. corrected data for processing
    - a. grayscale
    - b. edge enhancement
    - c. equalization
    - d. smoothing
  3. data for display
    - a. values of interest (VOI)
    - b. look-up table (LUT)
  4. post-processing
    - a. brightness
    - b. contrast
    - c. region of interest (ROI)
    - d. electronic cropping or masking
    - e. stitching
  5. display monitors
    - a. viewing conditions (e.g., viewing angle, ambient lighting)
    - b. spatial resolution (e.g., pixel size, pixel pitch)
    - c. brightness and contrast
  6. imaging informatics
    - a. DICOM
    - b. PACS
    - c. RIS (modality work list)
    - d. HIS
    - e. EMR or EHR

(Image Production continues on the following page.)



## Image Production (continued)

- C. Criteria for Image Evaluation of Technical Factors
  - 1. exposure indicator
  - 2. quantum noise (quantum mottle)
  - 3. gross exposure error (e.g., loss of contrast, saturation)
  - 4. contrast
  - 5. spatial resolution
  - 6. distortion (e.g., size, shape)
  - 7. identification markers (e.g., anatomical side, patient, date)
  - 8. image artifacts
  - 9. radiation fog
- D. Quality Control of Imaging Equipment and Accessories
  - 1. beam restriction
    - a. light field to radiation field alignment
    - b. central ray alignment
  - 2. recognition and reporting of malfunctions
  - 3. digital imaging receptor systems
    - a. maintenance (e.g., detector calibration, plate reader calibration)
    - b. QC tests (e.g., erasure thoroughness, plate uniformity, spatial resolution)
    - c. display monitor quality assurance (e.g., grayscale standard display function, luminance)
  - 4. shielding accessories (e.g., lead apron, glove testing)



## Procedures (64)

This section addresses imaging procedures for the anatomic regions listed below. Questions will cover the following topics:

1. Positioning (e.g., topographic landmarks, body positions, path of central ray, immobilization devices, respiration).
2. Anatomy (e.g., including physiology, basic pathology, and related medical terminology).
3. Procedure adaptation (e.g., body habitus, body mass index, trauma, pathology, age, limited mobility).
4. Evaluation of displayed anatomical structures (e.g., patient positioning, tube-part-image receptor alignment).

The specific radiographic positions and projections within each anatomic region that may be covered on the examination are listed in *Attachment A*. A guide to positioning terminology appears in *Attachment B*.

### 1. Head, Spine and Pelvis Procedures (18)

- A. Head
1. skull
  2. facial bones
  3. mandible
  4. zygomatic arch
  5. temporomandibular joints
  6. nasal bones
  7. orbits
  8. paranasal sinuses
- B. Spine and Pelvis
1. cervical spine
  2. thoracic spine
  3. scoliosis series
  4. lumbar spine
  5. sacrum and coccyx
  6. myelography
  7. sacroiliac joints
  8. pelvis and hip
  9. hysterosalpingography

### 2. Thorax and Abdomen Procedures (21)

- A. Thorax
1. chest
  2. ribs
  3. sternum
  4. soft tissue neck
- B. Abdomen and GI Studies
1. abdomen
  2. esophagus
  3. swallowing dysfunction study
  4. upper GI series, single or double contrast
  5. small bowel series
  6. contrast enema, single or double contrast
  7. surgical cholangiography
  8. ERCP

### C. Urological Studies

1. cystography
2. cystourethrography
3. intravenous urography
4. retrograde urography

### 3. Extremity Procedures (25)

#### A. Upper Extremities

1. fingers
2. hand
3. wrist
4. forearm
5. elbow
6. humerus
7. shoulder
8. scapula
9. clavicle
10. acromioclavicular joints

#### B. Lower Extremities

1. toes
2. foot
3. calcaneus
4. ankle
5. tibia/fibula
6. knee/patella
7. femur
8. long bone measurement

#### C. Other

1. bone age
2. bone survey (e.g. metastatic, child abuse)
3. arthrography



**Attachment A**

**Radiographic Positions and Projections**

**1. Head, Spine and Pelvis**

**A. Head**

1. Skull
  - a. AP axial (Towne)
  - b. lateral
  - c. PA axial (Caldwell)
  - d. PA
  - e. submentovertex (full basal)
  - e. PA axial (Haas)
  - f. trauma cross-table (horizontal beam) lateral
  - g. trauma AP axial (reverse Caldwell)
  - h. trauma AP
  - i. trauma AP axial (Towne)
2. Facial Bones
  - a. lateral
  - b. parietoacanthial (Waters)
  - c. PA axial (Caldwell)
  - d. modified parietoacanthial (modified Waters)
  - e. trauma acanthioparietal (reverse Waters)
3. Mandible
  - a. axiolateral oblique
  - b. PA
  - c. AP axial (Towne)
  - d. PA axial
  - e. PA (modified Waters)
  - f. submentovertex (full basal)
4. Zygomatic Arch
  - a. submentovertex (full basal)
  - b. parietoacanthial (Waters)
  - c. AP axial (modified Towne)
  - d. oblique inferosuperior (tangential)
5. Temporomandibular Joints
  - a. axiolateral oblique (modified Law)
  - b. axiolateral (modified Schuller)
  - c. AP axial (modified Towne)
6. Nasal Bones
  - a. parietoacanthial (Waters)
  - b. lateral
  - c. PA axial (Caldwell)
7. Orbits
  - a. parietoacanthial (Waters)
  - b. lateral
  - c. PA axial (Caldwell)
  - d. modified parietoacanthial (modified Waters)
8. Paranasal Sinuses
  - a. lateral, horizontal beam
  - b. PA axial (Caldwell), horizontal beam
  - c. parietoacanthial (Waters), horizontal beam
  - d. submentovertex (full basal), horizontal beam

**B. Spine and Pelvis**

1. Cervical Spine
  - a. AP axial
  - b. AP open mouth
  - c. lateral
  - d. cross-table (horizontal beam) lateral
  - e. PA axial obliques
  - f. AP axial obliques
  - g. lateral swimmers
  - h. lateral flexion and extension
  - i. AP dens (Fuchs)
2. Thoracic Spine
  - a. AP
  - b. lateral, breathing
  - c. lateral, expiration
3. Scoliosis Series
  - a. AP or PA
  - b. lateral
4. Lumbar Spine
  - a. AP
  - b. PA
  - c. lateral
  - d. L5-S1 lateral spot
  - e. posterior oblique
  - f. anterior oblique
  - e. AP axial L5-S1
  - g. AP right and left bending
  - h. lateral flexion and extension
5. Sacrum and Coccyx
  - a. AP axial sacrum
  - b. AP axial coccyx
  - c. lateral sacrum and coccyx, combined
  - d. lateral sacrum or coccyx, separate
6. Myelography
7. Sacroiliac Joints
  - a. AP
  - b. posterior oblique
  - c. anterior oblique
8. Pelvis and Hip
  - a. AP hip only
  - b. cross-table (horizontal beam) lateral hip
  - c. unilateral frog-leg, non-trauma
  - d. axiolateral inferosuperior, trauma (Clements-Nakayama)
  - e. AP pelvis
  - f. AP pelvis, bilateral frog-leg
  - g. AP pelvis, axial anterior pelvic bones (inlet, outlet)
  - h. anterior oblique pelvis, acetabulum (Judet)
9. Hysterosalpingography

**2. Thorax and Abdomen**

**A. Thorax**

1. Chest
  - a. PA or AP upright
  - b. lateral upright
  - c. AP lordotic
  - d. AP supine
  - e. lateral decubitus
  - f. anterior and posterior obliques
2. Ribs
  - a. AP and PA, above and below diaphragm
  - b. anterior and posterior obliques
3. Sternum
  - a. lateral
  - b. RAO
4. Soft Tissue Neck
  - a. AP upper airway
  - b. lateral upper airway

**B. Abdomen and GI Studies**

1. Abdomen
  - a. AP supine
  - b. AP upright
  - c. lateral decubitus
  - d. dorsal decubitus
2. Esophagus
  - a. RAO
  - b. left lateral
  - c. AP
  - d. PA
  - e. LAO
3. Swallowing Dysfunction Study
4. Upper GI series\*
  - a. AP scout
  - b. RAO
  - c. PA
  - d. right lateral
  - e. LPO
  - f. AP
5. Small Bowel Series
  - a. PA scout
  - b. PA (follow through)
  - c. ileocecal spots
6. Contrast Enema\*
  - a. left lateral rectum
  - b. left lateral decubitus
  - c. right lateral decubitus
  - d. LPO and RPO
  - e. PA
  - f. RAO and LAO
  - g. AP axial (sigmoid)
  - h. PA axial (sigmoid)
  - i. PA post-evacuation
7. Surgical Cholangiography
8. ERCP

\*single or double contrast



- C. Urological Studies**
1. Cystography
    - a. AP
    - b. LPO and RPO
    - c. lateral
    - d. AP axial
  2. Cystourethrography
    - a. AP voiding
      - cystourethrogram female
    - b. RPO voiding
      - cystourethrogram male
  3. Intravenous Urography
    - a. AP, scout, and series
    - b. RPO and LPO
    - c. post-void
  4. Retrograde Urography
    - a. AP scout
    - b. AP pyelogram
    - c. AP ureterogram
- 3. Extremities**
- A. Upper Extremities**
1. Fingers
    - a. PA entire hand
    - b. PA finger only
    - c. lateral
    - d. medial and/or lateral oblique
    - e. AP thumb
    - f. medial oblique thumb
    - g. lateral thumb
  2. Hand
    - a. PA
    - b. lateral
    - c. lateral oblique
  3. Wrist
    - a. PA
    - b. lateral oblique
    - c. lateral
    - d. PA–ulnar deviation
    - e. PA axial (Stecher)
    - f. tangential carpal canal (Gaynor-Hart)
  4. Forearm
    - a. AP
    - b. lateral
  5. Elbow
    - a. AP
    - b. lateral
    - c. lateral oblique
    - d. medial oblique
    - e. AP partial flexion
    - f. trauma axial laterals (Coyle)
  6. Humerus
    - a. AP
    - b. lateral
    - c. neutral
    - d. transthoracic lateral
  7. Shoulder
    - a. AP internal and external rotation
    - b. inferosuperior axial (Lawrence)
    - c. posterior oblique (Grashey)
    - d. AP neutral
    - e. scapular Y
  8. Scapula
    - a. AP
    - b. lateral
  9. Clavicle
    - a. AP
    - b. AP axial
    - c. PA axial
  10. Acromioclavicular Joints – AP
    - Bilateral With and Without Weights
- B. Lower Extremities**
1. Toes
    - a. AP, entire forefoot
    - b. AP or AP axial toe
    - c. oblique toe
    - d. lateral toe
    - e. sesamoids, tangential
  2. Foot
    - a. AP axial
    - b. medial oblique
    - c. lateral oblique
    - d. lateral
    - e. AP axial weight bearing
    - f. lateral weight bearing
  3. Calcaneus
    - a. lateral
    - b. plantodorsal, axial
    - c. dorsoplantar, axial
  4. Ankle
    - a. AP
    - b. mortise
    - c. lateral
    - d. medial oblique
    - e. AP stress views
    - f. AP weight bearing
    - g. lateral weight bearing
  5. Tibia/Fibula
    - a. AP
    - b. lateral
  6. Knee/patella
    - a. AP
    - b. Lateral
    - c. AP weight bearing
    - d. lateral oblique
    - e. medial oblique
    - f. PA axial–intercondylar fossa (Holmblad)
    - g. PA axial–intercondylar fossa (Camp Coventry)
    - h. AP axial–intercondylar fossa (Béclère)
    - i. PA patella
    - j. tangential (Merchant)
    - k. tangential (Settegast)
    - l. tangential (Hughston)
  7. Femur
    - a. AP
    - b. lateral
  8. Long Bone Measurement
- C. Other**
1. Bone Age
  2. Bone Survey
  3. Arthrography



**Attachment B**  
**Standard Terminology**  
**for Positioning and Projection**

**Radiographic View:** Describes the body part as seen by the image receptor or other recording medium, such as a fluoroscopic screen. Restricted to the discussion of a *radiograph* or *image*.

**Radiographic Position:** Refers to a specific body position, such as supine, prone, recumbent, erect or Trendelenburg. Restricted to the discussion of the *patient's physical position*.

**Radiographic Projection:** Restricted to the discussion of the *path of the central ray*.

POSITIONING TERMINOLOGY

A. Lying Down

1. *supine* – lying on the back
2. *prone* – lying face downward
3. *decubitus* – lying down with a horizontal x-ray beam
4. *recumbent* – lying down in any position

B. Erect or Upright

1. *anterior position* – facing the image receptor
2. *posterior position* – facing the radiographic tube

C. Either Upright or Recumbent


1. oblique torso positions
  - a. anterior oblique (facing the image receptor)
    - i. *left anterior oblique (LAO)* body rotated with the left anterior portion closest to the image receptor
    - ii. *right anterior oblique (RAO)* body rotated with the right anterior portion closest to the image receptor
  - b. posterior oblique (facing the radiographic tube)
    - i. *left posterior oblique (LPO)* body rotated with the left posterior portion closest to the image receptor
    - ii. *right posterior oblique (RPO)* body rotated with the right posterior portion closest to the image receptor
2. oblique extremity positions
  - a. lateral (external) rotation from either prone or supine, outward rotation of the extremity
  - b. medial (internal) rotation from either prone or supine, inward rotation of the extremity




RADIOGRAPHY EXAMINATION  
CONTENT SPECIFICATIONS

ARRT® BOARD APPROVED: **JANUARY 2016**  
IMPLEMENTATION DATE: **JANUARY 2017**


Anteroposterior Projection




Posteroanterior Projection




Right Lateral Position




Left Lateral Position




Left Posterior Oblique Position




Right Posterior Oblique Position



Left Anterior Oblique Position



Right Anterior Oblique Position





**Attachment C**  
**ARRT Standard Definitions**

|                     |  |
|---------------------|--|
| Digital Radiography | Digital Radiography includes both computed radiography and direct radiography<br><i>Computed Radiography (CR)</i> systems use storage phosphors to temporarily store energy representing the image signal. The phosphor then undergoes a process to extract the latent image.<br><i>Direct Radiography (DR)</i> systems have detectors that directly capture and readout an electronic image signal.   |
| Spatial Resolution  | The sharpness of the structural edges recorded in the image.   |
| Receptor Exposure   | The amount of radiation striking the image receptor.   |
| Brightness          | Brightness is the measurement of the luminance of an area in a radiographic image displayed on a monitor. It is calibrated in units of candela (cd) per square meter   |
| Contrast            | Contrast is the visible difference between any two selected areas of brightness levels within the displayed radiographic image. It is determined primarily by the processing algorithm (mathematical codes used by the software to provide the desired image appearance). The default algorithm determines the initial processing codes applied to the image data.<br><i>Grayscale</i> refers to the number of brightness levels (or gray shades) visible on an image and is linked to the bit depth of the system.<br><i>Long Scale</i> is the term used when slight differences between gray shades are present (low contrast) but the total number of gray shades is great.<br><i>Short Scale</i> is the term used when considerable or major differences between gray shades are present (high contrast) but the total number of gray shades is small. |
| Dynamic Range       | The range of exposures that may be captured by a detector.   |
| Receptor Contrast   | The fixed characteristic of the receptor. Most digital receptors have an essentially linear response to exposure. This is impacted by <b>contrast resolution</b> (the smallest exposure change or signal difference that can be detected). Ultimately, contrast resolution is limited by the <b>quantization</b> (number of bits per pixel) of the analog-to-digital convertor.  |
| Exposure Latitude   | The range of exposures which produces quality images at appropriate patient dose.  |
| Subject Contrast    | The magnitude of the signal difference in the remnant beam as a result of the different absorption characteristics of the tissues and structures making up that part.  |



# Nuclear Medicine Technology Examination

The purpose of *The American Registry of Radiologic Technologists® (ARRT®) Nuclear Medicine Technology Examination* is to assess the knowledge and cognitive skills underlying the intelligent performance of the tasks typically required of the nuclear medicine technologist. Using a nationwide survey, the ARRT periodically conducts a practice analysis to develop a task inventory which delineates or lists the job responsibilities typically required of nuclear medicine technologists.<sup>1</sup> An advisory committee then determines the knowledge and cognitive skills needed to perform the tasks on the task inventory and these are organized into the content categories within this document. The document is used to develop the examination. The results of the most recent practice analysis have been applied to this document. Every content category can be linked to one or more activities on the task inventory. The complete task inventory is available at [arrt.org](http://arrt.org).

The following table presents the four major content categories covered on the examination, and indicates the number of test questions in each category. The remaining pages list the specific topics addressed within each category, with the approximate number of test questions allocated to each topic appearing in parentheses.

This document is not intended to serve as a curriculum guide. Although ARRT programs for certification and registration and educational programs may have related purposes, their functions are clearly different. Educational programs are generally broader in scope and address the subject matter that is included in these content specifications, but do not limit themselves to only this content.

| Content Category                              | Number of Scored Questions <sup>2</sup> |
|---|---|
| Patient Care                                  | 20                                      |
| Safety <sup>3</sup>                           | 22                                      |
| Image Production                              | 38                                      |
| Procedures                                    | 120                                     |
| Radionuclides and Radiopharmaceuticals        |   |
| Cardiac Procedures                            |   |
| Endocrine and Oncology Procedures             |   |
| Gastrointestinal and Genitourinary Procedures |   |
| Other Imaging Procedures                      |   |
| <b>Total</b>                                  | <b>200</b>                              |

<sup>1</sup> A special debt of gratitude is due to the hundreds of professionals participating in this project as committee members, survey respondents, and reviewers.

<sup>2</sup> Each exam includes an additional 20 unscored (pilot) questions. On the pages that follow, the approximate number of test questions allocated to each content category appears in parentheses.

<sup>3</sup> SI and conventional units of radiation measurement will continue to be used on the nuclear medicine technology examination.



## Patient Care (20)

### 1. Patient Interactions and Management (20)

#### A. Ethical and Legal Aspects

1. patient's rights
  - a. informed consent (\*e.g., written, oral, implied)
  - b. confidentiality (HIPAA)
  - c. American Hospital Association (AHA) Patient Care Partnership (Patient's Bill of Rights)
    1. privacy
    2. extent of care (e.g., DNR)
    3. access to information
    4. living will, health care proxy, advanced directives
    5. research participation
2. legal issues
  - a. verification (e.g., patient identification, compare order to clinical indication, exam coding)
  - b. common terminology (e.g., battery, negligence, malpractice, beneficence)
  - c. legal doctrines (e.g., respondeat superior, res ipsa loquitur)
  - d. restraints versus immobilization
3. ARRT Standards of Ethics

#### B. Interpersonal Communication

1. modes of communication
  - a. verbal/written
  - b. nonverbal (e.g., eye contact, touching)
2. challenges in communication
  - a. interaction with others
    1. language barriers
    2. cultural and social factors
    3. physical and sensory impairments
    4. age
    5. emotional status, acceptance of condition
  - b. explanation of medical terms
  - c. strategies to improve understanding

#### 3. patient education

- a. explanation of current procedure (e.g., risks, benefits)
- b. verify informed consent when necessary
- c. pre- and post-examination instructions (e.g., preparations, diet, medications, discharge instructions)
- d. respond to inquiries about other imaging modalities (e.g., CT, MRI, mammography, sonography, radiography, bone densitometry)

#### C. Physical Assistance and Monitoring

1. patient transfer and movement
  - a. body mechanics (e.g., balance, alignment, movement)
  - b. patient transfer techniques
2. assisting patients with medical equipment
  - a. infusion catheters and pumps
  - b. oxygen delivery systems
  - c. other (e.g., nasogastric tubes, urinary catheters, tracheostomy tubes)
3. Routine Monitoring
  - a. vital signs
  - b. physical signs and symptoms (e.g., motor control, severity of injury)
  - c. fall prevention
  - d. documentation
  - e. immobilization
  - f. sedation

#### D. Medical Emergencies

1. allergic reactions (e.g., pharmaceuticals, latex)
2. cardiac or respiratory arrest (e.g., CPR)
3. physical injury or trauma
4. other medical disorders (e.g., seizures, diabetic reactions)

\*e.g., This is used here and in the remainder of this document to indicate examples of the topics covered, but not a complete list.

(Patient Care continues on the following page.)



## Patient Care (continued)

- E. Infection Control
  - 1. cycle of infection
    - a. pathogen
    - b. reservoir
    - c. portal of exit
    - d. mode of transmission
      - 1. direct
        - a. direct contact
        - b. droplet
      - 2. indirect
        - a. airborne
        - b. vehicle-borne – fomite
        - c. vector-borne – mechanical or biological
    - e. portal of entry
    - f. susceptible host
  - 2. asepsis
    - a. equipment disinfection
    - b. equipment sterilization
    - c. medical aseptic technique
    - d. sterile technique
  - 3. CDC Standard Precautions
    - a. hand hygiene
    - b. use of personal protective equipment (e.g., gloves, gowns, masks)
    - c. safe injection practices
    - d. safe handling of contaminated equipment/surfaces
    - e. disposal of contaminated materials
      - 1. linens
      - 2. needles
      - 3. patient supplies
      - 4. blood and body fluids
  - 4. transmission-based precautions
    - a. contact
    - b. droplet
    - c. airborne
  - 5. additional precautions
    - a. neutropenic precautions (reverse isolation)
    - b. healthcare associated (nosocomial) infections



## Safety (22)

### 1. Radiation Physics, Radiobiology, and Regulations (22)

#### A. Physical Properties of Radioactive Materials

1. decay of radioactivity
  - a. atomic structure
  - b. decay modes (e.g., alpha, beta, gamma)
  - c. decay rate
  - d. half-life
  - e. parent-daughter relationship
2. interaction of radiation with matter
  - a. photoelectric effect
  - b. Compton scattering
  - c. pair production and annihilation
  - d. internal conversion
  - e. Auger electron
  - f. bremsstrahlung

#### B. Biological Effects of Radiation

1. cellular biology
2. effects of radiation on cells
  - a. direct and indirect action
  - b. radiolysis of water
  - c. LET and RBE
3. stochastic and deterministic effects
4. acute effects of total body radiation
  - a. radiation sickness
  - b. hemopoietic syndrome
  - c. gastrointestinal syndrome
  - d. central nervous system syndrome
5. long term effects of radiation
  - a. somatic
  - b. genetic
6. relative tissue and organ sensitivity (e.g., law of Bergonié and Tribondeau)
7. effects of radiation on embryo/fetus

#### C. Basic Concepts of Radiation Protection

1. units of radiation exposure
2. principles of time, distance, and shielding
3. personnel protection equipment (e.g., gloves, lab coats)
4. personnel monitoring devices
  - a. types
  - b. use, care, and placement
5. ALARA
6. release of patients

#### D. NRC Regulations for Radiation Exposure

1. occupational
2. public
3. pregnancy or nursing
4. internal dosimetry and bioassays
5. personnel exposure records

#### E. Medical Events

1. definition
2. NRC regulations for reporting and notification

#### F. Area/Facilities Monitoring

1. Basic Concepts
  - a. units of measurement
  - b. exposure rates
  - c. definition of contaminated area
2. Survey Equipment and Techniques
  - a. well counters
  - b. survey meters
  - c. wipe test technique
3. NRC Regulations
  - a. frequency of surveys and wipes
  - b. classification of areas
    1. work
    2. treatment
    3. storage
  - c. posting of signs (e.g., types, locations)
  - d. documentation of survey and wipes results
    1. interpretation
    2. reporting (corrective action)
    3. record retention
4. Radioactive Spills
  - a. major spills
  - b. minor spills
  - c. processes for decontamination
  - d. reporting procedures

(Safety continues on the following page.)



## **Safety (continued)**

- G. Radioactive Materials
  - 1. inspection of incoming and outgoing materials (e.g., DOT and NRC regulations)
    - a. shipping labels
    - b. measurement of exposure rate
    - c. measurement of surface contamination
    - d. removable contamination limits/trigger levels
    - e. documentation
  - 2. storage
    - a. radiopharmaceuticals
    - b. sealed sources
    - c. consequences of improper storage
  - 3. disposal of radioactive waste
    - a. release to environment
    - b. decay in storage
    - c. transfer to authorized recipient
- H. Disposal of Pharmaceuticals
  - 1. expired pharmaceuticals
  - 2. partially used pharmaceuticals



## Image Production (38)

### 1. Instrumentation (38)

- A. Survey Meter
  - 1. operating principles
    - a. Geiger Müller
    - b. ionization chambers (cutie pies)
  - 2. quality control
    - a. frequency and types of checks
    - b. interpretation and record keeping
- B. Dose Calibrator
  - 1. operating principles
  - 2. quality control
    - a. frequency and types of checks
      - 1. accuracy
      - 2. constancy
      - 3. linearity
      - 4. geometry
    - b. interpretation and record keeping
- C. Scintillation Detector System
  - 1. operating principles
    - a. well counter
    - b. uptake probe (e.g., thyroid, surgical)
  - 2. quality control
    - a. radionuclide source
      - 1. energies
      - 2. type of source
    - b. parameters
      - 1. energy resolution
      - 2. efficiency
      - 3. high voltage calibration
      - 4. resolving time
      - 5. sensitivity
      - 6. energy linearity
      - 7. chi-square
    - c. interpretation and record keeping
- D. Gas and Aerosol Delivery Systems
  - 1. operating principles
  - 2. exhaust system (e.g., negative pressure, gas traps)
  - 3. interpretation and record keeping
- E. Gamma Camera
  - 1. operating principles
  - 2. quality control
    - a. frequency and types of checks
    - b. performance characteristics
      - 1. flood field uniformity
      - 2. high count uniformity correction
      - 3. spatial linearity
      - 4. spatial resolution
      - 5. energy resolution (e.g., FWHM)
      - 6. detector sensitivity
      - 7. extrinsic versus intrinsic methods
      - 8. center of rotation
      - 9. SPECT phantom measurements
    - c. interpretation and record keeping
  - 3. image acquisition
    - a. detector system
      - 1. count or time mode
      - 2. detector orientation
      - 3. photopeak energy setting and window width
      - 4. multi-energy acquisition
    - b. collimator selection
      - 1. types (e.g., parallel hole, pinhole)
      - 2. parameters (e.g., energy, resolution, sensitivity)
    - c. dynamic/static acquisition
      - 1. matrix selection
      - 2. framing (e.g., number and length)
      - 3. gating
      - 4. list mode
    - d. SPECT acquisition
      - 1. angular sampling/number of views (e.g., 180° versus 360°)
      - 2. matrix selection
      - 3. attenuation correction
      - 4. duration of acquisition

(Image Production continues on the following page.)



## Image Production (continued)

### F. PET/CT Scanner

1. PET operating principles
2. PET quality control
  - a. frequency and types of checks
  - b. characterization and correction calibration
    1. energy window calibration
    2. gain setting
    3. reference (blank) scan
    4. normalization calibration
    5. absolute activity (well counter) calibration
  - c. interpretation and record keeping
3. PET image acquisition
  - a. 2D versus 3D
  - b. list mode
  - c. respiratory gating
  - d. time-of-flight
4. CT operating principles\*
5. CT quality control\*
  - a. tube warm-up
  - b. CT number (water phantom)
6. CT image acquisition\*
  - a. kVp
  - b. mA
  - c. pitch
  - d. slice thickness
  - e. noise and uniformity
  - f. artifacts

### G. Data Processing

1. quantitative analysis (e.g., region of interest selection, ejection fraction, time activity curves, SUV)
2. reconstruction
  - a. registration (image fusion)
  - b. orientation
  - c. filter parameters
  - d. attenuation correction
  - e. gated images
  - f. motion correction
3. image management
  - a. archiving
  - b. PACS
  - c. HIS/RIS

\*Diagnostic CT is not assessed on the Nuclear Medicine Technology Examination. CT content is assessed for attenuation correction/anatomic localization.



## Procedures (120)

### 1. Radionuclides and Radiopharmaceuticals (24)

#### A. Production of Radionuclides

1. methods
  - a. reactor
  - b. accelerator
  - c. cyclotron
  - d. generator
2. purity
  - a. radionuclide
  - b. chemical
3. physical form (e.g., gas, solution, capsule)

#### B. Radiopharmaceutical Characteristics

1. method of localization
  - a. capillary blockade
  - b. active transport
  - c. phagocytosis
  - d. diffusion
  - e. compartmentalization
  - f. chemisorption
  - g. receptor binding
  - h. antigen antibody
  - i. filtration
  - j. metabolism
  - k. sequestration
2. half-life
  - a. physical
  - b. biological
  - c. effective
3. biodistribution
  - a. pharmacokinetics
  - b. critical organs
  - c. target organs

#### C. Preparation and Administration

1. kit preparation
  - a. labeling process
    1. principles
      - a. oxidation/reduction
      - b. pH
      - c. time for reaction
      - d. temperature
    2. compounding techniques
      - a. venting
      - b. heating
      - c. mixing
      - d. USP 797 regulations
    3. factors that affect labeling quality
  - b. shelf life and storage

- c. quality control
  1. radiochemical purity
  2. particle size
  3. specific activity (e.g., millicuries per mass)
  4. color and clarity
2. calculation of radiopharmaceutical and pharmaceutical dosage
  - a. units
    1. conversions
    2. calculations
  - b. volume determination
    1. formula
    2. decay tables
    3. concentration
    4. activity
3. pharmaceutical and radiopharmaceutical administration
  - a. preparation
    1. syringe
    2. needle selection
    3. shielding
  - b. radiopharmaceutical label
    1. name of radiopharmaceutical
    2. assay date and time
    3. lot number and expiration date
    4. concentration
    5. volume
    6. activity
  - c. administration techniques
    1. routes
    2. aseptic
    3. uniform distribution (e.g., mixing, agitation)
    4. complications and reactions
    5. documentation

(Procedures continue on the following page.)



**Procedures (continued)**

**TYPE OF STUDY**

**2. Cardiac Procedures (24)**

- A. Gated Blood Pool
- B. Myocardial Perfusion
- C. Viability

**3. Endocrine and Oncology Procedures (28)**

- A. Endocrine
  - 1. thyroid uptake/imaging
  - 2. parathyroid
  - 3. neuroendocrine
  - 4. adrenal imaging
- B. Tumor
  - 1. whole body
  - 2. SPECT or SPECT/CT
  - 3. PET/CT
  - 4. lymphoscintigraphy
- C. Therapy
  - 1. procedures
    - a. palliative bone
    - b. thyroid ablation
    - c. hyperthyroidism
    - d. non-Hodgkin lymphoma
    - e. selective internal radiation therapy with hepatic artery perfusion study (HAPS)
  - 2. regulations

**FOCUS OF QUESTIONS**

Questions about a specific study or procedure may address any of the following factors:

- A. Instrumentation
  - detector system
  - data acquisition
  - data analysis
  - ancillary equipment
- B. Radiopharmaceuticals and Pharmaceuticals
  - selection
  - dosage
  - administration
  - biodistribution
- C. Patient Preparation, Monitoring, and Education
  - indications and contraindications
  - pregnancy and nursing
  - dietary restrictions
  - adverse reactions
  - medications
  - age specific considerations
  - lab values
- D. Imaging Techniques
  - anatomical landmarks
  - views
  - patient-detector orientation
  - fusion imaging
- E. Anatomy and Pathophysiology
  - general anatomy
  - cross-sectional anatomy

(Procedures continue on the following page.)



## Procedures (continued)

### TYPE OF STUDY

#### 4. Gastrointestinal and Genitourinary Procedures (20)

- A. Gastric Emptying
- B. Gastroesophageal Reflux
- C. Meckel Diverticulum
- D. GI Bleed
- E. Hepatobiliary
- F. RBC Hemangioma
- G. Damaged RBC Spleen
- H. Liver/Spleen
- I. Renal Function
- J. Renal Cortical
- K. Radionuclide Cystogram

#### 5. Other Imaging Procedures (24)

- A. Abscess/Infection
- B. Bone
  - 1. planar
  - 2. 3-phase
  - 3. whole body
  - 4. SPECT or SPECT/CT
  - 5. PET/CT
- C. Central Nervous System
  - 1. brain death
  - 2. SPECT or SPECT/CT
  - 3. PET/CT
  - 4. cisternography/CSF leak
  - 5. shunt patency
- D. Lung
  - 1. perfusion
  - 2. ventilation – gas and aerosol
  - 3. quantitative

### FOCUS OF QUESTIONS

Questions about a specific study or procedure may address any of the following factors:

- A. Instrumentation
  - detector system
  - data acquisition
  - data analysis
  - ancillary equipment
- B. Radiopharmaceuticals and Pharmaceuticals
  - selection
  - dosage
  - administration
  - biodistribution
- C. Patient Preparation, Monitoring, and Education
  - indications and contraindications
  - pregnancy and nursing
  - dietary restrictions
  - adverse reactions
  - medications
  - age specific considerations
  - lab values
- D. Imaging Techniques
  - anatomical landmarks
  - views
  - patient-detector orientation
  - fusion imaging
- E. Anatomy and Pathophysiology
  - general anatomy
  - cross-sectional anatomy



## Attachment A: Nuclear Medicine Pharmaceuticals\*

### Radiopharmaceuticals

1. Tc-99m sodium pertechnetate
2. Tc-99m HDP
3. Tc-99m MDP
4. Tc-99m sestamibi
5. Tc-99m tetrofosmin
6. Tc-99m labeled RBCs
7. Tc-99m DTPA
8. Tc-99m DMSA
9. Tc-99m MAG3
10. Tc-99m HMPAO (Ceretek™)
11. Tc-99m ECD (Neurolite®)
12. Tc-99m HMPAO (Ceretek™) tagged WBCs
13. Tc-99m MAA
14. Tc-99m sulfur colloid
15. Tc-99m disofenin
16. Tc-99m mebrofenin (Choletec®)
17. In-111 DTPA
18. In-111 oxine labeled WBCs
19. In-111 pentetretotide (OctreoScan™)
20. Tl-201 thallous chloride
21. Xe-133 gas
22. I-123 sodium iodide
23. I-131 sodium iodide
24. I-123 ioflupane (DaTscan™)
25. I-123 MIBG
26. Ga-67 gallium citrate
27. F-18 fluorodeoxyglucose (FDG)
28. F-18 sodium fluoride (F-18 NaF)

### Therapeutic Radiopharmaceuticals

29. Y-90 ibritumomab tiuxetan (Zevalin®)
30. Ra-223 dichloride (Xofigo®)
31. I-131 MIBG
32. I-131 sodium iodide
33. Y-90 microspheres (Therasphere®, Sir-Spheres®)

### Interventional Pharmaceuticals

34. Adenosine
35. Aminophylline
36. Dipyridamole (Persantine®)
37. Dobutamine
38. Captopril
39. Furosemide (Lasix®)
40. Sincalide (Kinevac®)
41. Morphine
42. Regadenoson (Lexiscan®)
43. Lugol solution
44. Heparin
45. Recombinant TSH (Thyrogen®)
46. Oral CT contrast media

\*This is a list of commonly used pharmaceuticals that may appear on the exam. However, other pharmaceuticals may appear as practice changes



# Radiation Therapy Examination

The purpose of *The American Registry of Radiologic Technologists® (ARRT®) Radiation Therapy Examination* is to assess the knowledge and cognitive skills underlying the intelligent performance of the tasks typically required of radiation therapists at entry into the profession. Using a nationwide survey, the ARRT periodically conducts a practice analysis to develop a task inventory which delineates or lists the job responsibilities typically required of radiation therapists.<sup>1</sup> An advisory committee then determines the knowledge and cognitive skills needed to perform the tasks on the task inventory and these are organized into the content categories within this document. The document is used to develop the examination. The results of the most recent practice analysis have been applied to this document. Every content category can be linked to one or more activities on the task inventory. The complete task inventory is available at [arrt.org](http://arrt.org).

The following table presents the three major content categories covered on the examination, and indicates the number of test questions in each category. The remaining pages list the specific topics addressed within each category, with the approximate number of test questions allocated to each topic appearing in parentheses.

This document is not intended to serve as a curriculum guide. Although ARRT programs for certification and registration and educational programs may have related purposes, their functions are clearly different. Educational programs are generally broader in scope and address the subject matter that is included in these content specifications, but do not limit themselves to only this content.

| <b>Content Category</b>                             | <b>Number of Scored Questions<sup>2</sup></b> |
|---|---|
| Patient Care  | 47  |
| Patient Interactions                                |   |
| Patient and Medical Record Management               |   |
| Safety  | 49  |
| Radiation Physics, Equipment, and Quality Assurance |   |
| Radiation Protection <sup>3</sup>                   |   |
| Procedures  | 104   |
| Treatment Sites and Tumors                          |   |
| Treatment Volume Localization                       |   |
| Prescription and Dose Calculation                   |   |
| Treatments  |   |
| <b>Total</b>  | <b>200</b>                                    |

<sup>1</sup> A special debt of gratitude is due to the hundreds of professionals participating in this project as committee members, survey respondents and reviewers.

<sup>2</sup> Each exam includes an additional 20 unscored (pilot) questions.

<sup>3</sup> SI units are the primary (principal) units of radiation measurement used on the radiation therapy examination.



## Patient Care (47)

### 1. Patient Interactions (25)

- A. Ethical and Legal Aspects
  - 1. patient's rights
    - a. informed consent  
(\*e.g., written, oral, implied)
    - b. confidentiality (HIPAA)
    - c. American Hospital Association (AHA) Patient Care Partnership (Patient's Bill of Rights)
      - 1. privacy
      - 2. goal of care  
(e.g., definitive, palliative)
      - 3. access to information
      - 4. living will, advanced directive  
(e.g., DNR), health care proxy
      - 5. research participation
  - 2. legal issues
    - a. verification (e.g., patient identification, treatment site, prescription)
    - b. common terminology (e.g., battery, negligence, malpractice, beneficence)
    - c. legal doctrines (e.g., respondeat superior, res ipsa loquitur)
    - d. restraints versus immobilization
  - 3. ARRT Standards of Ethics
- B. Interpersonal Communications
  - 1. modes of professional communication
    - a. verbal/written
    - b. nonverbal (e.g., eye contact, touching)
  - 2. challenges in communication
    - a. interaction with others
      - 1. language barriers
      - 2. cultural and social factors
      - 3. physical or sensory impairments
      - 4. age
      - 5. emotional status, acceptance of condition (e.g., stage of grief)
    - b. explanation of medical terms
  - 3. patient education
    - a. explanation of treatment
    - b. strategies to improve understanding
    - c. treatment compliance  
(e.g., positioning, skin marks)
  - 4. support services
    - a. hospice
    - b. other professionals  
(e.g., dietitian, clergy, social services)

### C. Physical Assistance

- 1. patient transfer and movement
  - a. body mechanics (e.g., balance, alignment, movement)
  - b. patient transfer techniques
  - c. fall prevention
- 2. assisting patients with medical equipment
  - a. infusion catheters and pumps
  - b. oxygen delivery systems
  - c. other (e.g., nasogastric tubes, urinary catheters, tracheostomy tubes)

### D. Medical Emergencies

- 1. allergic reactions
  - a. contrast media
    - 1. contraindications
    - 2. adverse reactions
  - b. other (e.g., latex)
- 2. cardiac or respiratory arrest  
(e.g., CPR)
- 3. physical injury or trauma
- 4. other medical disorders  
(e.g., seizures, diabetic reactions)

\*e.g., This is used here and in the remainder of this document to indicate examples of the topics covered, but not a complete list.

(Patient Care continues on the following page.)



## Patient Care (continued)

### E. Infection Control

1. cycle of infection
  - a. pathogen
  - b. reservoir
  - c. portal of exit
  - d. mode of transmission
    1. direct
      - a. direct contact
      - b. droplet
    2. indirect
      - a. airborne
      - b. vehicle-borne – fomite
      - c. vector-borne – mechanical or biological
  - e. portal of entry
  - f. susceptible host
2. asepsis
  - a. equipment disinfection
  - b. equipment sterilization
  - c. medical aseptic technique
  - d. sterile technique
3. CDC Standard Precautions
  - a. hand hygiene
  - b. use of personal protective equipment (e.g., gloves, gowns, masks)
  - c. safe needle practices
  - d. safe handling of contaminated materials
  - e. disposal of contaminated materials
    1. linens
    2. needles
    3. patient supplies
    4. blood and body fluids
4. transmission-based precautions
  - a. contact
  - b. droplet
  - c. airborne
5. additional precautions
  - a. neutropenic precautions (reverse isolation)
  - b. healthcare associated (nosocomial) infections

### F. Handling and Disposal of Toxic or Hazardous Material

1. types of materials
  - a. metals (e.g., block alloy)
  - b. chemicals
  - c. chemotherapy
2. material safety data sheet (MSDS)

(Patient Care continues on the following page.)



## Patient Care (continued)

### 2. Patient and Medical Record Management (22)

- A. Evaluation
  - 1. epidemiology and etiology
    - a. cancer risk factors
    - b. prevalence and incidence
  - 2. cancer screening
  - 3. signs and symptoms
  - 4. history and physical examination
  - 5. imaging studies (e.g., CT, MRI, PET/CT)
  - 6. other diagnostic studies
    - a. lab results
    - b. surgical reports
    - c. pathology reports
- B. Assessment
  - 1. treatment side effects
    - a. signs and symptoms
    - b. causes
    - c. management
  - 2. blood studies
    - a. types of studies (e.g., CBC, BUN, creatinine)
    - b. factors affecting blood values
  - 3. dietary counseling
    - a. common problems
    - b. causes
    - c. dietary management
  - 4. routine monitoring
    - a. weight
    - b. vital signs
    - c. signs and symptoms
    - d. documentation

### C. Documentation

- 1. information included in treatment record
  - a. prescription
  - b. monitor units
  - c. target dose (daily and accumulated)
  - d. energy and type of radiation
  - e. date
  - f. time of day for b.i.d. treatment
  - g. fraction
  - h. elapsed days
  - i. field number and description
  - j. doses to other regions of interest
  - k. set-up instructions
- 2. elements of record keeping
  - a. patient identification
  - b. accountability (e.g., signatures)
  - c. accuracy and legibility
  - d. variance from prescription (errors, prescription changes)
  - e. medical events (definition and required documentation)
- 3. charge capture terminology
  - a. professional and technical components
  - b. CPT® principles



## Safety (49)

### 1. Radiation Physics, Equipment, and Quality Assurance (20)

- A. Sources of Radiation
  - 1. radioactive material
  - 2. machine-produced radiation
- B. Basic Properties of Radiation
  - 1. wave characteristics
  - 2. attenuation
  - 3. inverse-square law
  - 4. x-ray beam quality
- C. Interactions with Matter
  - 1. photon interactions (e.g., Compton, photoelectric effect)
  - 2. electron interactions
  - 3. particle interactions (e.g., proton, neutron)
- D. Components and Operation
  - 1. linear accelerator
  - 2. CT simulator
- E. Quality Control Procedures
  - 1. warm-up and inspection of linear accelerators and CT simulators
    - a. interlock systems
    - b. safety lights
    - c. emergency switches
    - d. critical machine parameters (e.g., pressure, temperature)
    - e. electrical and mechanical hazards
    - f. imaging systems
  - 2. radiation output verification
    - a. methods
    - b. frequency
    - c. effect of environment (e.g., humidity) on measurements
  - 3. light and treatment field checks
    - a. light and radiation field agreement
    - b. collimator indicator agreement
    - c. multileaf collimator performance
    - d. sidelight/laser accuracy check (isocenter)
  - 4. rotation check
    - a. safety procedures
    - b. operation of gantry/console
  - 5. evaluation of quality assurance results
    - a. interpretation
    - b. course of action
    - c. documentation

### 2. Radiation Protection (29)

- A. Biological Effects of Radiation
  - 1. radiosensitivity
  - 2. dose-response relationships
  - 3. somatic effects
    - a. cellular
    - b. tissue (e.g., hemopoietic, skin, reproductive organs)
    - c. embryonic and fetal risks
    - d. carcinogenesis
    - e. early versus late effects
    - f. acute versus chronic effects
- B. Radiation Tissue Tolerance
  - 1. tolerance levels (TD<sub>5/5</sub>)
  - 2. adverse effects
  - 3. dose to critical structures
  - 4. radiobiological factors (e.g., dose, fractionation, volume)
  - 5. biological factors (e.g., age, anatomic variation, medical conditions)
  - 6. medical factors (e.g., prior surgery, pacemakers)
  - 7. other factors (e.g., radiosensitizers, radioprotectors)
  - 8. contribution from other sources
    - a. chemotherapy
    - b. brachytherapy<sup>1</sup>
    - c. other fields (e.g., prior or abutting)
    - d. radiation effect modifiers

<sup>1</sup> Only basic concepts related to common uses of brachytherapy are covered, including dose to surrounding tissue and radiation protection issues. Specific procedures and isotope characteristics are not covered.

(Safety continues on the following page.)



## Safety (continued)

- C. Measurement of Radiation
  - 1. units of measurement
    - a. absorbed dose
    - b. dose equivalent
    - c. exposure
  - 2. instrumentation
    - a. ionization chamber
    - b. Geiger-Müller detector
    - c. TLD/OSL (optically stimulated luminescence)
    - d. diodes
    - e. neutron detectors
- D. Fundamental Principles
  - 1. ALARA
  - 2. basic methods of protection (time, distance, shielding)
- E. Personnel Monitoring
  - 1. NCRP recommendations for personnel monitoring (report #116)
    - a. occupational exposure
    - b. public exposure
    - c. embryo/fetus exposure
  - 2. maintenance and evaluation of personnel dosimetry records
- F. Facilities and Area Monitoring
  - 1. NRC regulations (10 CFR, parts 20 and 35)
    - a. classification of areas (restricted, controlled, unrestricted)
    - b. required postings (signs)
    - c. area monitoring devices
  - 2. barrier requirements
    - a. primary
    - b. secondary
- G. Handling and Disposal of Radioactive Materials



## Procedures (104)

### 1. Treatment Sites and Tumors (26)

- A. Anatomy, Pathophysiology, Lymphatic Drainage, and Metastatic Patterns
  1. brain and spinal cord
  2. head and neck (includes thyroid and salivary glands)
  3. breast
  4. lung
  5. abdomen, pelvis, GI, and GU
    - a. esophagus, stomach, small bowel, large bowel, rectum, and anus
    - b. pancreas, adrenals, liver, and gallbladder
    - c. ureters, kidneys, bladder, and urethra
  6. reproductive
    - a. prostate, testes
    - b. endometrium, cervix, ovaries, uterus, vagina, and vulva
  7. skeletal
  8. miscellaneous
    - a. lymphoma (Hodgkin and non-Hodgkin)
    - b. sarcomas (bone and soft tissue)
    - c. multiple myeloma
    - d. skin
    - e. leukemia
    - f. mycosis fungoides
    - g. bone marrow transplant
    - h. benign (e.g., heterotopic bone, keloid, AVM)
    - i. oncologic emergencies (e.g., whole brain, SVC, cord compression)
- B. Tumor Classification
  1. histopathologic types (e.g., benign, sarcomas, carcinomas)
  2. histopathologic grade
    - a. purpose (differentiation and growth rate)
    - b. grading system (e.g., GX, G1-G4)
  3. staging (basic concepts; not specific sites)
    - a. purpose
    - b. components (e.g., TNM, I-IV)

### 2. Treatment Volume Localization (18)

- A. Treatment Techniques and Anatomic Relationships
  1. radiation therapy techniques
  2. sectional and topographic anatomy
  3. critical organs
  4. patient positioning and immobilization
  5. types and uses of contrast media
- B. CT Simulation
  1. CT image acquisition (e.g., mA, slice thickness, and spacing)
  2. CT image processing and display (e.g., reconstruction, window level, field of view, CT number)
  3. contour volume and isocenter determination
  4. image transmission, storage, and retrieval
  5. programmable lasers
- C. Documentation of Simulation Procedure
  1. anatomic position
  2. equipment orientation
  3. accessory equipment
  4. field parameters
  5. set-up diagrams or photographs
  6. temporary and/or permanent reference marks

(Procedures continues on the following page.)



## Procedures (continued)

### 3. Prescription and Dose Calculation (24)

- A. Treatment Prescription
  - 1. total target dose
  - 2. fractionation schedules
  - 3. beam energy
  - 4. types of radiation
  - 5. treatment volume (e.g., GTV, CTV, PTV)
  - 6. number of fields
  - 7. fixed/rotational fields
  - 8. field weighting
  - 9. field orientation
  - 10. treatment unit capabilities and limitations
  - 11. plan modifications
  - 12. beam modifiers
- B. Geometric Parameters and Patient Measurements
  - 1. field size and shape
  - 2. target depth
  - 3. patient thickness
  - 4. SSD, SAD
  - 5. collimator setting
  - 6. abutting fields (e.g., gap calculations)
  - 7. fusion with outside diagnostic studies
- C. Dose Calculation and Verification
  - 1. selection of energy
  - 2. equivalent square (open and blocked field)
  - 3. scatter factors (e.g., collimator, phantom)
  - 4.  $D_{max}$
  - 5. percentage depth dose
  - 6. TAR, TMR
  - 7. SSD, SAD
  - 8. inverse square
  - 9. extended distance factors
  - 10. wedges (e.g., wedge angle or factor)
  - 11. off-axis calculation
  - 12. isodose curve characteristics (e.g., penumbra, DVH)
  - 13. factors for beam modifiers (e.g., tray factor, bolus, compensator)
  - 14. inhomogeneity correction factors
  - 15. rotational factors
  - 16. machine output data
  - 17. verification and documentation

### 4. Treatments (36)

- A. Treatment Options (indications, benefits, risks)
  - 1. chemotherapy
  - 2. surgery
  - 3. radiation therapy
    - a. external beam
    - b. brachytherapy<sup>2</sup>
  - 4. multimodality treatment
- B. Verification and Application of the Treatment Plan
  - 1. patient position
  - 2. isocenter
  - 3. treatment parameters (e.g., beam orientation, energy)
  - 4. prescription
  - 5. modality
    - a. 2D
    - b. 3D
    - c. 4D (e.g., respiratory gating)
    - d. IMRT
    - e. arc therapy
    - f. stereotactic
  - 6. imaging procedures
    - a. kV imaging
    - b. cone beam CT (CBCT)
    - c. MV imaging

<sup>2</sup> Only basic concepts related to common uses of brachytherapy are covered, including dose to surrounding tissue and radiation protection issues. Specific procedures and isotope characteristics are not covered.

(Procedures continues on the following page.)



## Procedures (continued)

- C. Treatment Machine Set-Up
  - 1. auxiliary set-up devices
    - a. couch indexing
    - b. positioning aids
    - c. alignment lasers
  - 2. machine operations
    - a. SSD, SAD
    - b. collimator or cone settings
    - c. optical or mechanical distance indicator
    - d. gantry angle
    - e. collimator angle
    - f. field light
    - g. treatment couch
    - h. console controls
    - i. pendant controls
- D. Treatment Accessories
  - 1. beam modifiers
    - a. compensating filters
    - b. shielding
    - c. blocks (e.g., thickness, half value layer (HVL), half-value thickness (HVT))
    - d. multileaf collimation
    - e. bolus
    - f. wedges (enhanced dynamic wedge, physical wedge)
  - 2. immobilization devices
    - a. custom
    - b. standard
  - 3. parameters
    - a. SSD, SAD, depth
    - b. gantry, collimator, and field size settings
    - c. beam energy and type
- E. Treatment Administration
  - 1. patient monitoring
    - a. visual (mirror, TV monitor)
    - b. two-way voice communication system
    - c. back-up systems
    - d. monitoring regulations
    - e. emergency situations
  - 2. record and verify systems
  - 3. image acquisition and registration
  - 4. site verification
  - 5. dose verification (e.g., diodes, film)
  - 6. equipment malfunctions
    - a. types (e.g., radiation, electrical, mechanical, software)
    - b. troubleshooting and correction
    - c. documentation and reporting



# Limited Scope of Practice in Radiography

The purpose of the *Limited Scope of Practice in Radiography Examination*, which is developed and administered by *The American Registry of Radiologic Technologists (ARRT)* on behalf of state licensing agencies, is to assess the knowledge and cognitive skills underlying the intelligent performance of the tasks typically required of operators of radiographic equipment used to radiograph selected anatomic regions (chest, extremities, etc.). ARRT administers the examination to state approved candidates under contractual arrangement with the state and provides the results directly to the state. This examination is not associated with any type of certification and registration by the ARRT.

The knowledge and skills covered by the examination were determined by administering a comprehensive practice analysis survey to a nationwide sample of radiographers and adopting a subset of the tasks developed for the radiography task inventory as the limited scope task inventory. The task inventory appears in *Attachment D* of this document. The content specifications for the limited scope examination identify the knowledge areas underlying performance of the tasks on the limited scope task inventory. Every content category can be linked to one or more activities on the task inventory.

It is the philosophy of the ARRT that an individual licensed in limited scope radiography possess the same knowledge and cognitive skill, in his or her specific area of radiography, as radiographers. The modules covered by the examination are outlined below. Subsequent pages describe in detail the topics covered within each module. All candidates take the CORE module of the examination and one or more PROCEDURE modules, depending on the type of license for which they have applied.

| Core Module  | Number of Scored Questions <sup>1</sup> | Testing Time |
|--|---|--------------|
| Patient Care   | 18                                      |              |
| <i>Patient Interactions and Management (18)</i>        |   |              |
| Safety <sup>2</sup>                                    | 40                                      |              |
| <i>Radiation Physics and Radiobiology (12)</i>         |   |              |
| <i>Radiation Protection (28)</i>                       |   |              |
| Image Production                                       | 42                                      |              |
| <i>Image Acquisition and Technical Evaluation (20)</i> |   |              |
| <i>Equipment Operation and Quality Assurance (22)</i>  |   |              |
| Total for Core Module                                  | 100                                     | 1 hr, 55 min |
| <b>Procedure Modules</b>                               |   |              |
| 1. Chest   | 20                                      | 25 min       |
| 2. Extremities   | 25                                      | 30 min       |
| 3. Skull/Sinuses                                       | 20                                      | 25 min       |
| 4. Spine   | 25                                      | 30 min       |
| 5. Podiatric   | 20                                      | 25 min       |

<sup>1</sup> The core module includes an additional 15 unscored (pilot) questions. Each of the procedure modules has five additional unscored questions.

<sup>2</sup> SI units will become the primary (principle) units of radiation measurement used on the limited scope of practice in radiography examination in 2018.

1



## Patient Care

### 1. Patient Interactions and Management

#### A. Ethical and Legal Aspects

1. patient's rights
  - a. informed consent (\*e.g., written, oral, implied)
  - b. confidentiality (HIPAA)
  - c. American Hospital Association (AHA) Patient Care Partnership (Patient's Bill of Rights)
    1. privacy
    2. extent of care (e.g., DNR)
    3. access to information
    4. living will, health care proxy, advanced directives
    5. research participation
2. legal issues
  - a. verification (e.g., patient identification, compare order to clinical indication)
  - b. common terminology (e.g., battery, negligence, malpractice, beneficence)
  - c. legal doctrines (e.g., respondeat superior, res ipsa loquitur)
  - d. restraints versus immobilization
  - e. manipulation of electronic data (e.g., exposure indicator, processing algorithm, brightness and contrast, cropping or masking off anatomy)
3. Professional Ethics

#### B. Interpersonal Communication

1. modes of communication
  - a. verbal/written
  - b. nonverbal (e.g., eye contact, touching)
2. challenges in communication
  - a. interactions with others
    1. language barriers
    2. cultural and social factors
    3. physical or sensory impairments
    4. age
    5. emotional status, acceptance of condition
  - b. explanation of medical terms
  - c. strategies to improve understanding
3. patient education (e.g., explanation of current procedure purpose, exam length)

#### C. Physical Assistance and Monitoring

1. patient transfer and movement
  - a. body mechanics (e.g., balance, alignment, movement)
  - b. patient transfer techniques
2. assisting patients with medical equipment (e.g., oxygen delivery systems, urinary catheters)
3. routine monitoring
  - a. vital signs
  - b. physical signs and symptoms (e.g., motor control, severity of injury)
  - c. fall prevention
  - d. documentation

#### D. Medical Emergencies

1. allergic reactions (e.g., contrast media, latex)
2. cardiac or respiratory arrest (e.g., CPR)
3. physical injury or trauma
4. other medical disorders (e.g., seizures, diabetic reactions)

\* The abbreviation "e.g.," is used to indicate that examples are listed in parentheses, but that it is not a complete list of all possibilities.

(Patient Care continues on the following page.)



LIMITED SCOPE OF PRACTICE  
EXAMINATION CONTENT SPECIFICATIONS

ARRT BOARD APPROVED: **JANUARY 2017**  
IMPLEMENTATION DATE: **JANUARY 1, 2018**

**Patient Care (continued)**

- E. Infection Control
    - 1. cycle of infection
      - a. pathogen
      - b. reservoir
      - c. portal of exit
      - d. mode of transmission
        - 1. direct
          - a. droplet
          - b. direct contact
        - 2. indirect
          - a. airborne
          - b. vehicle borne–fomite
          - c. vector borne–mechanical or biological
      - e. portal of entry
      - f. susceptible host
    - 2. asepsis
      - a. equipment disinfection
      - b. equipment sterilization
      - c. medical aseptic technique
      - d. sterile technique
  - 3. CDC Standard Precautions
    - a. hand hygiene
    - b. use of personal protective equipment (e.g., gloves, gowns, masks)
    - c. safe injection practices
    - d. safe handling of contaminated equipment/surfaces
    - e. disposal of contaminated materials
      - 1. linens
      - 2. needles
      - 3. patient supplies
      - 4. blood and body fluids
  - 4. transmission-based precautions
    - a. contact
    - b. droplet
    - c. airborne
  - 5. additional precautions
    - a. neutropenic precautions (reverse isolation)
    - b. healthcare associated (nosocomial) infections
- F. Handling and Disposal of Toxic or Hazardous Material
  - 1. chemicals
  - 2. safety data sheet (e.g., material safety data sheets)



## Safety

### 1. Radiation Physics and Radiobiology

#### A. Principles of Radiation Physics

1. x-ray production
  - a. source of free electrons  
(e.g., thermionic emission)
  - b. acceleration of electrons
  - c. focusing of electrons
  - d. deceleration of electrons
2. target interactions
  - a. bremsstrahlung
  - b. characteristic
3. x-ray beam
  - a. frequency and wavelength
  - b. beam characteristics
    1. quality
    2. quantity
    3. primary versus remnant (exit)
  - c. inverse square law
  - d. fundamental properties  
(e.g., travel in straight lines, ionize matter)
4. photon interactions with matter
  - a. Compton effect
  - b. photoelectric absorption
  - c. coherent (classical) scatter
  - d. attenuation by various tissues
    1. thickness of body part
    2. type of tissue (atomic number)

#### B. Biological Aspects of Radiation

1. SI units of measurement (NCRP Report #160)
  - a. absorbed dose (Gy)
  - b. dose equivalent (Sv)
  - c. exposure (C/kg)
  - d. effective dose (Sv)
2. radiosensitivity
  - a. dose-response relationships
  - b. relative tissue radiosensitivities  
(e.g., LET, RBE)
  - c. cell survival and recovery (LD50)
  - d. oxygen effect
3. somatic effects
  - a. short-term versus long-term effects
  - b. acute versus chronic effects
  - c. carcinogenesis
  - d. organ and tissue response  
(e.g., eye, thyroid, breast, bone marrow, skin, gonadal)
4. acute radiation syndromes
  - a. hemopoietic
  - b. gastrointestinal (GI)
  - c. central nervous system (CNS)
5. embryonic and fetal risks
6. genetic impact
  - a. genetically significant dose
  - b. goals of gonadal shielding

(Safety continues on the following page.)



## Safety (continued)

### 2. Radiation Protection

- A. Minimizing Patient Exposure
  - 1. exposure factors
    - a. kVp
    - b. mAs
  - 2. shielding
    - a. rationale for use
    - b. types
    - c. placement
  - 3. beam restriction
    - a. purpose of primary beam restriction
    - b. types (e.g., collimators)
  - 4. filtration
    - a. effect on skin and organ exposure
    - b. effect on average beam energy
    - c. NCRP recommendations (NCRP #102, minimum filtration in useful beam)
  - 5. patient considerations
    - a. positioning
    - b. communication
    - c. pediatric
    - d. morbid obesity
  - 6. radiographic dose documentation
  - 7. image receptors
  - 8. dose area product (DAP) meter
- B. Personnel Protection (ALARA)\*
  - 1. sources of radiation exposure
    - a. primary x-ray beam
    - b. secondary radiation
      - 1. scatter
      - 2. leakage
    - c. patient as source
  - 2. basic methods of protection
    - a. time
    - b. distance
    - c. shielding
  - 3. protective devices
    - a. types
    - b. attenuation properties
    - c. minimum lead equivalent (NCRP #102)
  - 4. radiation exposure and monitoring
    - a. dosimeters
      - 1. types
      - 2. proper use
    - b. NCRP recommendations for personnel monitoring (NCRP #116)
      - 1. occupational exposure
      - 2. public exposure
      - 3. embryo/fetus exposure
      - 4. dose equivalent limits
      - 5. evaluation and maintenance of personnel dosimetry records

\* Note: Although it is the responsibility of the individual licensed in limited scope radiography to apply radiation protection principles to minimize bioeffects for both patients and personnel, the ALARA concept is specific to personnel protection and is listed only for that section.



LIMITED SCOPE OF PRACTICE  
EXAMINATION CONTENT SPECIFICATIONS

ARRT BOARD APPROVED: **JANUARY 2017**  
IMPLEMENTATION DATE: **JANUARY 1, 2018**

## Image Production

### 1. Image Acquisition and Technical Evaluation

#### A. Selection of Technical Factors Affecting Radiographic Quality

Refer to *Attachment C* to clarify terms that may occur on the exam. (X indicates topics covered on the examination.)

|                                      | 1.<br>Receptor<br>Exposure | 2.<br>Contrast | 3.<br>Spatial<br>Resolution | 4.<br>Distortion |
|--------------------------------------|----------------------------|----------------|-----------------------------|------------------|
| a. mAs                               | X                          |                |                             |                  |
| b. kVp                               | X                          | X              |                             |                  |
| c. OID                               |                            | X (air gap)    | X                           | X                |
| d. SID                               | X                          |                | X                           | X                |
| e. focal spot size                   |                            |                | X                           |                  |
| f. tube filtration                   | X                          | X              |                             |                  |
| g. beam restriction                  | X                          | X              |                             |                  |
| h. motion                            |                            |                | X                           |                  |
| i. anode heel effect                 | X                          |                |                             |                  |
| j. patient factors (size, pathology) | X                          | X              | X                           | X                |
| k. angle (tube, part, or receptor)   |                            |                | X                           | X                |

#### B. Technique Charts

1. anatomically programmed technique
2. caliper measurement
3. fixed versus variable kVp
4. special considerations
  - a. pathologic factors
  - b. age (e.g., pediatric, geriatric)
  - c. body mass index (BMI)

#### C. Digital Imaging Characteristics

1. spatial resolution (equipment related)
  - a. pixel characteristics (e.g., size, pitch)
  - b. detector element (DEL) (e.g., size, pitch, fill factor)
  - c. matrix size
  - d. sampling frequency

#### 2. contrast resolution (equipment related)

- a. bit depth
- b. modulation transfer function (MTF)
- c. detective quantum efficiency (DQE)
3. image signal (exposure related)
  - a. dynamic range
  - b. quantum noise (quantum mottle)
  - c. signal to noise ratio (SNR)
  - d. contrast to noise ratio (CNR)

#### D. Image Identification

1. methods (e.g., radiographic, electronic)
2. legal considerations (e.g., patient data, examination data)

(Image Production continues on the following page.)



LIMITED SCOPE OF PRACTICE  
EXAMINATION CONTENT SPECIFICATIONS

ARRT BOARD APPROVED: **JANUARY 2017**  
IMPLEMENTATION DATE: **JANUARY 1, 2018**

## Image Production (continued)

### 2. Equipment Operation and Quality Assurance

- A. Imaging Equipment
  - 1. components of radiographic unit (fixed or mobile)
    - a. operating console
    - b. x-ray tube construction
      - 1. electron source
      - 2. target materials
      - 3. induction motor
    - c. manual exposure controls
    - d. beam restriction
  - 2. x-ray generator, transformers and rectification system
    - a. basic principles
    - b. tube loading
  - 3. components of digital imaging
    - a. CR components
      - 1. plate (e.g., photo-stimulable phosphor [PSP])
      - 2. plate reader
    - b. DR image receptors
      - 1. flat panel
      - 2. charge coupled device (CCD)
      - 3. complementary metal oxide semiconductor (CMOS)
- B. Image Processing and Display
  - 1. raw data (pre-processing)
    - a. analog-to-digital converter (ADC)
    - b. quantization
    - c. corrections (e.g., rescaling, flat fielding, dead pixel correction)
    - d. histogram
  - 2. corrected data for processing
    - a. grayscale
    - b. edge enhancement
    - c. equalization
    - d. smoothing
  - 3. data for display
    - a. values of interest (VOI)
    - b. look-up table (LUT)
  - 4. post-processing
    - a. brightness
    - b. contrast
    - c. region of interest (ROI)
    - d. electronic cropping or masking
    - e. stitching
  - 5. display monitors
    - a. viewing conditions (e.g., viewing angle, ambient lighting)
    - b. spatial resolution (e.g., pixel size, pixel pitch)
    - c. brightness and contrast
  - 6. imaging informatics
    - a. DICOM
    - b. PACS
    - c. RIS (modality work list)
    - d. HIS
    - e. EMR or EHR

(Image Production continues on the following page.)



LIMITED SCOPE OF PRACTICE  
EXAMINATION CONTENT SPECIFICATIONS

ARRT BOARD APPROVED: **JANUARY 2017**  
IMPLEMENTATION DATE: **JANUARY 1, 2018**

### **Image Production (continued)**

#### **C. Criteria for Image Evaluation of Technical Factors**

1. exposure indicator
2. quantum noise (quantum mottle)
3. gross exposure error (e.g., loss of contrast, saturation)
4. contrast
5. spatial resolution
6. distortion (e.g., size, shape)
7. identification markers (e.g., anatomical side, patient, date)
8. image artifacts
9. radiation fog

#### **D. Quality Control of Imaging Equipment and Accessories**

1. beam restriction
  - a. light field to radiation field alignment
  - b. central ray alignment
2. recognition and reporting of malfunctions
3. digital imaging receptor systems
  - a. maintenance (e.g., detector calibration, plate reader calibration)
  - b. QC tests (e.g., erasure thoroughness, plate uniformity, spatial resolution)
  - c. display monitor quality assurance (e.g., grayscale standard display function, luminance)
4. shielding accessories (e.g., lead apron, glove testing)



LIMITED SCOPE OF PRACTICE  
EXAMINATION CONTENT SPECIFICATIONS

ARRT BOARD APPROVED: **JANUARY 2017**  
IMPLEMENTATION DATE: **JANUARY 1, 2018**

## Procedures

The specific positions and projections within each anatomic region that may be covered on the examination are listed in *Attachment A*. A guide to positioning terminology appears in *Attachment B*.

| <b>PROCEDURE<br/>MODULE</b> <sup>1</sup>   | <b># QUESTIONS<br/>PER MODULE</b> <sup>2</sup> | <b>FOCUS OF QUESTIONS</b> <sup>3</sup>   |
|--|--|--|
| <b>1. Chest</b>  |  |  |
| A. Routine   | 16   | <b>1. Positioning</b> (e.g., topographic landmarks, body positions, path of central ray, immobilization devices, respiration)<br><br>emphasis: high<br><br><b>2. Anatomy</b> (including physiology, basic pathology, and related medical terminology)<br><br>emphasis: medium<br><br><b>3. Evaluation of displayed anatomical structures</b> (e.g., patient positioning, tube-part-image receptor alignment)<br><br>emphasis: medium<br><br><b>4. Procedure adaptation</b> (e.g., body habitus, body mass index, trauma, pathology, age, limited mobility, casts, splints, soft tissue for foreign body, etc.)<br><br>emphasis: low<br><br><b>5. Equipment and Accessories</b> (grids or Bucky, compensating filter, automatic exposure control [AEC], automatic collimation)<br><br>emphasis: low |
| B. Other   | 4  |  |
| <b>TOTAL</b>   | <b>20</b>                                      |  |
| <b>2. Extremities</b>  |  |  |
| A. Lower (toes, foot, calcaneus, ankle, tibia/fibula, knee/ patella, and distal femur) | 11   |  |
| B. Upper (fingers, hand, wrist, forearm, elbow, and humerus)                           | 11   |  |
| C. Pectoral Girdle (shoulder, scapula, clavicle, and acromioclavicular joints)         | 3  |  |
| <b>TOTAL</b>   | <b>25</b>                                      |  |
| <b>3. Skull/Sinuses</b>  |  |  |
| A. Skull   | 8  |  |
| B. Paranasal Sinuses   | 8  |  |
| C. Facial Bones (orbits, nasal bones)  | 4  |  |
| <b>TOTAL</b>   | <b>20</b>                                      |  |
| <b>4. Spine</b>  |  |  |
| A. Cervical Spine  | 8  |  |
| B. Thoracic Spine  | 6  |  |
| C. Lumbar Spine  | 8  |  |
| D. Sacrum, Coccyx, and Sacroiliac Joints   | 2  |  |
| E. Scoliosis Series  | 1  |  |
| <b>TOTAL</b>   | <b>25</b>                                      |  |
| <b>5. Podiatric</b>  |  |  |
| A. Foot and Toes   | 14   |  |
| B. Ankle   | 5  |  |
| C. Calcaneus (os calcis)   | 1  |  |
| <b>TOTAL</b>   | <b>20</b>                                      |  |

### Notes:

- Examinees take one or more procedure modules, depending on the type of license they have applied for. Each procedure module has 20 or 25 scored test questions, depending on the module (see chart above). The number of questions within a module should be regarded as approximate values.
- Each of the procedure modules has five additional unscored questions.
- The procedure modules may include questions about the five areas listed under *FOCUS OF QUESTIONS* on the right side of the chart. The podiatric module does not include questions from the equipment and accessories section.



**Attachment A**  
**Radiographic Positions and Projections**

- I. Chest**
  - A. Chest
    1. PA or AP upright
    2. lateral upright
    3. AP Lordotic
    4. AP supine
    5. lateral decubitus
    6. anterior and posterior obliques
- II. Extremities**
  - A. Toes
    1. AP, entire forefoot
    2. AP or AP axial toe
    3. oblique toe
    4. lateral toe
    5. sesamoids, tangential
  - B. Foot
    1. AP axial
    2. medial oblique
    3. lateral oblique
    4. lateral
    5. AP axial weight bearing
    6. lateral weight bearing
  - C. Calcaneus
    1. lateral
    2. plantodorsal, axial
    3. dorsoplantar, axial
  - D. Ankle
    1. AP
    2. mortise
    3. lateral
    4. medial oblique
    5. AP stress views
    6. AP weight bearing
    7. lateral weight bearing
  - E. Tibia/Fibula
    1. AP
    2. lateral
  - F. Knee/patella
    1. AP
    2. Lateral
    3. AP weight bearing
    4. lateral oblique
    5. medial oblique
    6. PA axial–intercondylar fossa (Holmblad)
    7. PA axial–intercondylar fossa (Camp Coventry)
    8. AP axial–intercondylar fossa (Béclère)
    9. PA patella
    10. Tangential (Merchant)
    11. tangential (Settegast)
    12. tangential (Hughston)
  - G. Femur (Distal)
    1. AP
    2. lateral
  - H. Fingers
    1. PA entire hand
    2. PA finger only
    3. lateral
    4. medial and/or lateral oblique
    5. AP thumb
    6. medial oblique thumb
    7. lateral thumb
  - I. Hand
    1. PA
    2. lateral
    3. lateral oblique
  - J. Wrist
    1. PA
    2. lateral oblique
    3. lateral
    4. PA–ulnar deviation
    5. PA axial (Stecher)
    6. tangential carpal canal (Gaynor-Hart)
  - K. Forearm
    1. AP
    2. lateral
  - L. Elbow
    1. AP
    2. lateral
    3. lateral oblique
    4. medial oblique
    5. AP partial flexion
    6. trauma axial laterals (Coyle)
  - M. Humerus
    1. AP
    2. lateral
    3. neutral
    4. transthoracic lateral
  - N. Shoulder
    1. AP internal and external rotation
    2. inferosuperior axial (Lawrence)
    3. posterior oblique (Grashey)
    4. AP neutral
    5. scapular Y
  - O. Scapula
    1. AP
    2. lateral
  - P. Clavicle
    1. AP
    2. AP axial
    3. PA axial
  - Q. Acromioclavicular Joints – AP Bilateral With and Without Weights
- III. Skull/Sinuses**
  - A. Skull
    1. AP axial (Towne)
    2. lateral
    3. PA axial (Caldwell)
    4. PA
    5. submentovertex (full basal)
  - B. Facial Bones
    1. lateral
    2. parietoacanthial (Waters)
    3. PA axial (Caldwell)
    4. modified parietoacanthial (modified Waters)
  - C. Nasal Bones
    1. parietoacanthial (Waters)
    2. lateral
    3. PA axial (Caldwell)
  - D. Orbits
    1. parietoacanthial (Waters)
    2. lateral
    3. PA axial (Caldwell)
    4. modified parietoacanthial (modified Waters)
  - E. Paranasal Sinuses
    1. lateral, horizontal beam
    2. PA axial (Caldwell), horizontal beam
    3. parietoacanthial (Waters), horizontal beam
    4. submentovertex (full basal), horizontal beam
- IV. Spine**
  - A. Cervical Spine
    1. AP axial
    2. AP open mouth
    3. lateral
    4. PA axial obliques
    5. AP axial obliques
    6. lateral swimmers
    7. lateral flexion and extension
  - B. Thoracic Spine
    1. AP
    2. lateral, breathing
    3. lateral, expiration
  - C. Lumbar Spine
    1. AP
    2. PA
    3. lateral
    4. L5-S1 lateral spot
    5. posterior oblique
    6. anterior oblique
    7. AP axial L5-S1
    8. AP right and left bending
    9. lateral flexion and extension
  - D. Sacrum and Coccyx
    1. AP axial sacrum
    2. AP axial coccyx
    3. lateral sacrum and coccyx, combined
    4. lateral sacrum or coccyx, separate
  - E. Sacroiliac Joints
    1. AP
    2. posterior oblique
    3. anterior oblique
  - F. Scoliosis Series
    1. AP or PA
    2. lateral
- V. Podiatric**
  - A. Foot and Toes
    1. dorsal plantar (DP)\*
    2. medial oblique
    3. lateral oblique
    4. lateral\*
    5. sesamoidal axial\*
  - B. Ankle\*
    1. AP\*
    2. mortise\*
    3. AP medial oblique\*
    4. AP lateral oblique\*
    5. lateral\*
  - C. Calcaneus
    1. axial calcaneal\*
    2. Harris and Beath (ski-jump)\*



LIMITED SCOPE OF PRACTICE  
EXAMINATION CONTENT SPECIFICATIONS

ARRT BOARD APPROVED: **JANUARY 2017**  
IMPLEMENTATION DATE: **JANUARY 1, 2018**

**Attachment B**  
**Standard Terminology**  
**for Positioning and Projection**

**Radiographic View:** Describes the body part as seen by the image receptor or other recording medium, such as a fluoroscopic screen. Restricted to the discussion of a *radiograph* or *image*.

**Radiographic Position:** Refers to a specific body position, such as supine, prone, recumbent, erect or Trendelenburg. Restricted to the discussion of the *patient's physical position*.

**Radiographic Projection:** Restricted to the discussion of the *path of the central ray*.

POSITIONING TERMINOLOGY


- A. Lying Down
  - 1. *supine* – lying on the back
  - 2. *prone* – lying face downward
  - 3. *decubitus* – lying down with a horizontal x-ray beam
  - 4. *recumbent* – lying down in any position
- B. Erect or Upright
  - 1. *anterior position* – facing the image receptor
  - 2. *posterior position* – facing the radiographic tube
- C. Either Upright or Recumbent
  - 1. oblique torso positions
    - a. anterior oblique (facing the image receptor)
      - i. *left anterior oblique (LAO)* body rotated with the left anterior portion closest to the image receptor
      - ii. *right anterior oblique (RAO)* body rotated with the right anterior portion closest to the image receptor
    - b. posterior oblique (facing the radiographic tube)
      - i. *left posterior oblique (LPO)* body rotated with the left posterior portion closest to the image receptor
      - ii. *right posterior oblique (RPO)* body rotated with the right posterior portion closest to the image receptor
  - 2. oblique extremity positions
    - a. lateral (external) rotation from either prone or supine, outward rotation of the extremity
    - b. medial (internal) rotation from either prone or supine, inward rotation of the extremity




LIMITED SCOPE OF PRACTICE  
EXAMINATION CONTENT SPECIFICATIONS

ARRT BOARD APPROVED: **JANUARY 2017**  
IMPLEMENTATION DATE: **JANUARY 1, 2018**


Anteroposterior Projection




Posteroanterior Projection




Right Lateral Position




Left Lateral Position




Left Posterior Oblique Position




Right Posterior Oblique Position



Left Anterior Oblique Position



Right Anterior Oblique Position





LIMITED SCOPE OF PRACTICE  
EXAMINATION CONTENT SPECIFICATIONS

ARRT BOARD APPROVED: **JANUARY 2017**  
IMPLEMENTATION DATE: **JANUARY 1, 2018**

**Attachment C**  
**ARRT Standard Definitions**

|                     |  |
|---------------------|--|
| Digital Radiography | Digital Radiography includes both computed radiography and direct radiography.<br><u>Computed Radiography (CR)</u> systems use storage phosphors to temporarily store energy representing the image signal. The phosphor then undergoes a process to extract the latent image.<br><u>Direct Radiography (DR)</u> systems have detectors that directly capture and readout an electronic image signal.  |
| Spatial Resolution  | The sharpness of the structural edges recorded in the image.   |
| Receptor Exposure   | The amount of radiation striking the image receptor.   |
| Brightness          | Brightness is the measurement of the luminance of an area in a radiographic image displayed on a monitor. It is calibrated in units of candela (cd) per square meter.  |
| Contrast            | Contrast is the visible difference between any two selected areas of brightness levels within the displayed radiographic image. It is determined primarily by the processing algorithm (mathematical codes used by the software to provide the desired image appearance). The default algorithm determines the initial processing codes applied to the image data.<br><u>Grayscale</u> refers to the number of brightness levels (or gray shades) visible on an image and is linked to the bit depth of the system.<br><u>Long Scale</u> is the term used when slight differences between gray shades are present (low contrast) but the total number of gray shades is great.<br><u>Short Scale</u> is the term used when considerable or major differences between gray shades are present (high contrast) but the total number of gray shades is small. |
| Dynamic Range       | The range of exposures that may be captured by a detector.   |
| Receptor Contrast   | The fixed characteristic of the receptor. Most digital receptors have an essentially linear response to exposure. This is impacted by <b>contrast resolution</b> (the smallest exposure change or signal difference that can be detected). Ultimately, contrast resolution is limited by the <b>quantization</b> (number of bits per pixel) of the analog-to-digital convertor.  |
| Exposure Latitude   | The range of exposures which produces quality images at appropriate patient dose.  |
| Subject Contrast    | The magnitude of the signal difference in the remnant beam as a result of the different absorption characteristics of the tissues and structures making up that part.  |



LIMITED SCOPE OF PRACTICE  
EXAMINATION CONTENT SPECIFICATIONS

ARRT BOARD APPROVED: **JANUARY 2017**  
IMPLEMENTATION DATE: **JANUARY 1, 2018**

**Attachment D**  
**Task Inventory for Limited Scope of Practice in Radiography Examination**

| <b>Activity</b>   | <b>Content Categories</b><br>Legend: PC = Patient Care,<br>S = Safety, IP = Image Production,<br>P = Procedures |
|---|---|
| 1. Confirm patient's identity.  | PC.1.A.2.A., PC.1.B., IP.1.D.   |
| 2. Evaluate patient's ability to understand and comply with requirements for the requested examination.   | PC.1.B.   |
| 3. Obtain pertinent medical history.  | PC.1.A.2.A., PC.1.B.  |
| 4. Manage complex interpersonal interactions within the workplace in an effective manner.   | PC.1.B.2.   |
| 5. Review imaging examination request to verify accuracy and completeness of information (e.g., patient history, clinical diagnosis, physician's orders). | PC.1.A.2.A.   |
| 6. Respond as appropriate to imaging study inquiries from patients.   | PC.1.B.   |
| 7. Assume responsibility for medical equipment attached to patients (e.g., IVs, oxygen) during the imaging procedures.                                    | PC.1.C.2.   |
| 8. Follow environmental protection standards for handling and disposing of bio-hazardous materials (e.g., sharps, blood, and body fluids).                | PC.1.E.3.E.   |
| 9. Provide for patient safety, comfort, and modesty.  | PC.1.A., PC.1.C.  |
| 10. Notify appropriate personnel of adverse events or incidents (e.g., patient fall, wrong patient imaged).   | PC.1.A.2.A., PC.1.C.3., IP.1.D.   |
| 11. Communicate scheduling delays to waiting patients.  | PC.1.B.   |
| 12. Demonstrate and promote professional and ethical behavior.  | PC.1.A., PC.1.B.  |
| 13. Verify informed consent as necessary.   | PC.1.A.1.A., PC.1.B.  |
| 14. Communicate relevant information to others (e.g., M.D.s, RNs, other radiology personnel).   | PC.1.A., PC.1.B., PC.1.C.3.D.   |
| 15. Explain procedure instructions to patient or patient's family.  | PC.1.B.3.   |
| 16. Practice Standard Precautions.  | PC.1.E.3.   |
| 17. Follow appropriate procedures when caring for patients with communicable diseases.  | PC.1.E.3., PC.1.E.4., PC.1.E.5.   |
| 18. Use immobilization devices, as needed, to prevent patient movement and/or ensure patient safety.  | PC.1.A.2.D., P.   |
| 19. Use proper body mechanics when assisting a patient.   | PC.1.C.1.A.   |
| 20. Use patient transfer devices when needed.   | PC.1.C.1.B.   |
| 21. Use sterile or aseptic technique when indicated.  | PC.1.E.2.   |
| 22. Follow environmental protection standards for handling hazardous materials.   | PC.1.F.   |
| 23. Obtain vital signs.   | PC.1.C.3.A.   |



LIMITED SCOPE OF PRACTICE  
EXAMINATION CONTENT SPECIFICATIONS

ARRT BOARD APPROVED: **JANUARY 2017**  
IMPLEMENTATION DATE: **JANUARY 1, 2018**

| <b>Activity</b>  | <b>Content Categories</b>  |
|--|--|
|  | Legend: PC = Patient Care,<br>S = Safety, IP = Image Production,<br>P = Procedures |
| 24. Recognize and communicate the need for prompt medical attention.   | PC.1.C.3., PC.1.D.   |
| 25. Administer emergency care.   | PC.1.C.2., PC.1.C.3., PC.1.D.  |
| 26. Explain post-procedural instructions to patient or patient's family.   | PC.1.B.3.  |
| 27. Maintain confidentiality of patient information.   | PC.1.A.1.B., PC.1.A.3.   |
| 28. Clean, disinfect, or sterilize facilities and equipment, and dispose of contaminated items in preparation for next examination.  | PC.1.E.2., PC.1.E.3.   |
| 29. Document required information on patient's medical record (e.g., imaging procedure documentation, images).<br>a. On paper<br>b. Electronically   | PC.1.B.1.A., PC.1.C.3.D.,<br>IP.1.D., IP.2.B.6.                                    |
| 30. Evaluate the need for and use of protective shielding.   | S.2.A.2., S.2.B.3.   |
| 31. Take appropriate precautions to minimize radiation exposure to the patient.  | S.2.A.   |
| 32. Question female patient of child-bearing age about date of last menstrual period or possible pregnancy and take appropriate action (e.g., document response, contact physician).   | PC.1.B., S.1.B.5., S.1.B.6.  |
| 33. Restrict beam to the anatomical area of interest.  | S.2.A.3., IP.1.A.1.G.,<br>IP.1.A.2.G.  |
| 34. Set technical factors to produce diagnostic images and adhere to ALARA.  | S.2.A., IP.1.A., IP.1.B.   |
| 35. Document radiographic procedure dose.  | S.2.A.6., IP.2.B.6.E.  |
| 36. Prevent all unnecessary persons from remaining in area during x-ray exposure.  | S.2.B.4.B.   |
| 37. Take appropriate precautions to minimize occupational radiation exposure.  | S.2.B.   |
| 38. Advocate radiation safety and protection.  | S.1.B., S.2.A., S.2.B.4.B.   |
| 39. Describe the potential risk of radiation exposure when asked.  | PC.1.B.3., S.1.B.  |
| 40. Wear a personnel monitoring device while on duty.  | S.2.B.4.A.   |
| 41. Evaluate individual occupational exposure reports to determine if values for the reporting period are within established limits.   | S.2.B.4.B.   |
| 42. Determine appropriate exposure factors using the following:<br>a. Fixed kVp technique chart<br>b. Variable kVp technique chart<br>c. Calipers (to determine patient thickness for exposure)<br>d. Anatomically programmed technique* | IP.1.A., IP.1.B.   |

\* Applies to specific modules



LIMITED SCOPE OF PRACTICE  
EXAMINATION CONTENT SPECIFICATIONS

ARRT BOARD APPROVED: **JANUARY 2017**  
IMPLEMENTATION DATE: **JANUARY 1, 2018**

| Activity  | Content Categories   |
|---|--|
|   | Legend: PC = Patient Care,<br>S = Safety, IP = Image Production,<br>P = Procedures |
| 43. Select radiographic exposure factors.<br>a. Automatic Exposure Control (AEC)*<br>b. kVp and mAs (manual)  | IP.1.A., IP.1.B., IP.1.C.  |
| 44. Operate radiographic unit and accessories including:<br>a. Fixed unit<br>b. Mobile unit (portable)  | IP.2.A.1., IP.2.A.2., IP.2.A.3.  |
| 45. Operate electronic imaging and record keeping devices including:<br>a. Computed radiography (CR) with photostimulable storage phosphor (PSP) plates<br>b. Direct radiography (DR)<br>c. Picture archiving and communication system (PACS)<br>d. Hospital information system (HIS)<br>e. Radiology information system (RIS)<br>f. Electronic medical record (EMR) system | IP.2.A.3., IP.2.B.   |
| 46. Modify technical factors to correct for noise in a digital image.   | IP.1.D.3.B., IP.2.C.   |
| 47. Remove all radiopaque materials from patient or table that could interfere with the image (e.g., clothing removal, jewelry removal).  | PC.1.B.3.A., IP.2.C.8.   |
| 48. Perform post-processing on digital images in preparation for interpretation.  | IP.2.B.4.  |
| 49. Use radiopaque anatomical side markers at the time of image acquisition.  | IP.1.E., IP.2.C.7.   |
| 50. Add electronic annotations on digital images to indicate position or other relevant information (e.g., time, upright, decubitus, post-void).  | PC.1.A.2.E., IP.1.E., IP.2.C.7.  |
| 51. Select equipment and accessories (e.g., grid*, compensating filter*, shielding) for the examination requested.  | S.2.A.2., P.   |
| 52. Explain breathing instructions prior to making the exposure.*   | PC.1.B.3.A., IP.1.A.3.H., P.   |
| 53. Position patient to demonstrate the desired anatomy using anatomical landmarks.   | P.   |
| 54. Modify exposure factors for circumstances such as involuntary motion, casts and splints*, pathological conditions, or patient's inability to cooperate.   | IP.1.A.3.H., IP.1.A.3.J.,<br>IP.1.B., P.   |
| 55. Verify accuracy of patient identification on image.   | IP.1.E., IP.2.C.7.   |
| 56. Evaluate images for diagnostic quality.   | IP.2.C., IP.2.D., P.   |
| 57. Respond appropriately to digital exposure indicator values.   | IP.2.C.1.  |
| 58. Determine corrective measures if image is not of diagnostic quality and take appropriate action.  | IP.2.C., P.  |
| 59. Identify image artifacts and make appropriate corrections as needed.  | IP.2.C.8.  |
| 60. Store and handle image receptor in a manner which will reduce the possibility of artifact production.   | IP.2.C.8., IP.2.C.9., IP.2.D.3.  |

\* Applies to specific module



LIMITED SCOPE OF PRACTICE  
EXAMINATION CONTENT SPECIFICATIONS

ARRT BOARD APPROVED: **JANUARY 2017**  
IMPLEMENTATION DATE: **JANUARY 1, 2018**

| <b>Activity</b>  | <b>Content Categories</b>  |
|--|--|
|  | Legend: PC = Patient Care,<br>S = Safety, IP = Image Production,<br>P = Procedures |
| 61. Visually inspect, recognize, and report malfunctions in the imaging unit and accessories.  | IP.2.C.8., IP.2.D.2.   |
| 62. Recognize the need for periodic maintenance and evaluation of radiographic equipment affecting image quality and radiation safety (e.g., shielding, image display monitor, light field, central ray detector calibration).   | IP.2.D.  |
| 63. Perform routine maintenance on digital equipment including: <ul style="list-style-type: none"> <li>a. Detector calibration</li> <li>b. CR plate erasure</li> <li>c. Equipment cleanliness</li> <li>d. Test images</li> </ul> | IP.2.D.3.  |
| 64. Adapt radiographic procedures for patient condition (e.g., age, size, trauma, pathology) and location (e.g., mobile, surgical, isolation).   | PC.1.C., PC.1.E., S.2.A.5., IP.1., P.  |
| 65. Select appropriate geometric factors (e.g., SID, OID, focal spot size, tube angle).  | IP.1.A.  |
| <b>Position patient, x-ray tube, and image receptor to perform the following diagnostic examinations:</b>  |  |
| 66. Chest  | P.1.A.   |
| 67. Cervical spine   | P.4.A.   |
| 68. Thoracic spine   | P.4.B.   |
| 69. Scoliosis series   | P.4.E.   |
| 70. Lumbar spine   | P.4.C.   |
| 71. Sacrum/coccyx  | P.4.D.   |
| 72. Sacroiliac joints  | P.4.D.   |
| 73. Skull  | P.3.A.   |
| 74. Facial bones   | P.3.C.   |
| 75. Nasal bones  | P.3.C.   |
| 76. Orbits   | P.3.C.   |
| 77. Paranasal sinuses  | P.3.B.   |
| 78. Toes   | P.2.A., P.5.A.   |
| 79. Foot   | P.2.A., P.5.A.   |
| 80. Calcaneus  | P.2.A., P.5.C.   |
| 81. Ankle  | P.2.A., P.5.B.   |
| 82. Tibia/fibula   | P.2.A.   |

\* Applies to specific modules



LIMITED SCOPE OF PRACTICE  
EXAMINATION CONTENT SPECIFICATIONS

ARRT BOARD APPROVED: **JANUARY 2017**  
IMPLEMENTATION DATE: **JANUARY 1, 2018**

**Content Categories**

Legend: PC = Patient Care,  
S = Safety, IP = Image Production,  
P = Procedures

**Activity**

---

|                              |        |
|------------------------------|--------|
| 83. Knee/patella             | P.2.A. |
| 84. Distal femur             | P.2.A. |
| 85. Fingers                  | P.2.B. |
| 86. Hand                     | P.2.B. |
| 87. Wrist                    | P.2.B. |
| 88. Forearm                  | P.2.B. |
| 89. Elbow                    | P.2.B. |
| 90. Humerus                  | P.2.B. |
| 91. Shoulder                 | P.2.C. |
| 92. Scapula                  | P.2.C. |
| 93. Clavicle                 | P.2.C. |
| 94. Acromioclavicular joints | P.2.C. |

# Sample Questions for the Examination for the Limited Scope of Practice in Radiography

The following questions are similar in format to those in the examinations, but they do not necessarily represent the full range of content or levels of difficulty. An answer key is provided at the bottom of the page.

**DIRECTIONS:** Each of the questions or incomplete statements below is followed by four suggested answers or completions. Select the one which is best in each case.

001. Infection control guidelines require:

- a. hand washing after contact with each patient
- b. destroying all contaminated linen
- c. sterilizing cassettes prior to use
- d. wearing gloves for all patient contact

002. Which of the following sets of technical factors will produce the radiograph with the greatest density?

|    | mA  | msec | SID | kVp |
|----|-----|------|-----|-----|
| a. | 100 | 500  | 48" | 80  |
| b. | 200 | 250  | 36" | 76  |
| c. | 400 | 125  | 40" | 76  |
| d. | 500 | 100  | 40" | 76  |

003. Grid cut-off on a radiograph may be caused by all of the following EXCEPT:

- a. a tube that is off-center
- b. a tube that is angled perpendicular to the lead strips
- c. grid motion being started before the exposure is made
- d. improper distance being used with a focused grid

004. The purpose of the x-ray tube housing is to:


- 1. provide adequate filtration
- 2. provide electrical shielding
- 3. control leakage radiation

- a. 1 & 2 only
- b. 1 & 3 only
- c. 2 & 3 only
- d. 1, 2 & 3

Item 005 refers to the anatomical sketch on the right.

005. The medial condyle is the structure numbered:

- a. 1
- b. 3
- c. 5
- d. 8



|      |      |      |      |      |      |
|------|------|------|------|------|------|
| KEY: | 1. A | 2. B | 3. C | 4. C | 5. C |
|------|------|------|------|------|------|

# Pearson VUE Test Centers

This list may change after publication in this handbook. For an up-to-date list at any time, check the [www.pearsonvue.com/arrt](http://www.pearsonvue.com/arrt) website.

## Alabama

Birmingham  
Decatur  
Dothan  
Mobile  
Montgomery

## Alaska

Anchorage

## Arizona

Chandler  
Phoenix  
Tempe  
Tucson

## Arkansas

Fort Smith  
Little Rock  
Rogers  
Texarkana

## California

Alhambra  
Anaheim  
Culver City  
Daly City  
Fairfield  
Fresno  
Gardena  
Lake Forest  
Milpitas  
Oakland  
Ontario  
Pasadena  
Redding  
Redlands  
Roseville  
Sacramento  
San Diego  
San Dimas  
San Marcos  
San Mateo  
San Francisco  
Santa Maria  
Visalia  
Westlake Village

## Colorado

Colorado Springs  
Greenwood  
Village  
Westminster

## Connecticut

Norwalk  
Wallingford  
Wethersfield

## Dist. of Columbia

Washington

## Delaware

Dover  
Newark

## Florida

Altamonte Springs  
Deerfield Beach  
Doral  
Gainesville  
Jacksonville  
Lakeland  
Orlando  
Pembroke Pines  
Plantation  
Port Charlotte  
St. Petersburg  
Tallahassee  
Tampa

## Georgia

Albany  
Atlanta  
Augusta  
Macon  
Savannah  
Stockbridge

## Hawaii

Honolulu

## Idaho

Boise

## Illinois

Buffalo Grove  
Chicago  
Marion  
Peoria  
Rosemont  
Schaumburg  
Springfield

## Indiana

Crown Point  
Evansville  
Fort Wayne  
Indianapolis  
Terre Haute

## Iowa

Coralville  
Davenport  
Sioux City  
W. Des Moines

## Kansas

Hays  
Overland Park  
Topeka  
Wichita

## Kentucky

Lexington  
Louisville

## Louisiana

Baton Rouge  
Metairie  
New Orleans  
Shreveport

## Maine

Bangor  
Westbrook

## Maryland

Baltimore  
Bethesda  
Columbia  
Salisbury

## Massachusetts

Boston  
Cambridge  
Lexington  
Springfield  
Wellesley  
Worcester

## Michigan

Ann Arbor  
Dearborn  
Grand Rapids  
Lansing  
Marquette  
Southfield  
Troy

## Minnesota

Bloomington  
Brooklyn Park  
Eagan  
Hermantown  
Rochester  
St. Paul

## Mississippi

Jackson  
Tupelo

## Missouri

Columbia  
Kansas City  
Springfield  
St. Louis

## Montana

Billings  
Helena

## Nebraska

Lincoln  
North Platte  
Omaha

## Nevada

Las Vegas  
Reno

## New Hampshire

Concord

## New Jersey

Atlantic City  
Jersey City  
Lyndhurst  
Piscataway  
Princeton

## New Mexico

Albuquerque

## New York

Albany  
Brooklyn  
East Syracuse  
Forest Hills  
Getzville  
Islandia  
Lake Success  
New York  
Rochester  
Staten Island  
Utica  
White Plains  
Williamsville  
Yonkers

## North Carolina

Asheville  
Charlotte  
Durham  
Greenville  
Raleigh  
Wilmington  
Winston-Salem

## North Dakota

Bismarck  
Fargo

## Ohio

Beachwood  
Columbus  
Copley Twp.  
Gahanna  
Independence  
Mason  
Maumee  
Moraine  
Norwood  
Westlake

## Oklahoma

Norman  
Oklahoma City  
Tulsa

## Oregon

Beaverton  
Medford  
Portland  
Salem

## Pennsylvania

Allentown  
Blue Bell  
Erie  
Harrisburg  
King of Prussia  
Lancaster  
Philadelphia  
Pittsburgh  
Scranton  
Warrington  
Washington

## Rhode Island

Warwick

## South Carolina

Columbia  
Greenville  
North Charleston

## South Dakota

Sioux Falls

## Tennessee

Brentwood  
Chattanooga  
Johnson City  
Knoxville  
Memphis  
Nashville

## Texas

Abilene  
Amarillo  
Arlington  
Austin  
Bellaire  
Bryan  
Carrollton  
Corpus Christi  
Dallas  
El Paso  
Harlingen  
Houston  
Hurst  
Lubbock  
McAllen  
Midland  
San Antonio  
Shavano Park  
Sugar Land  
Tyler  
Waco

## Utah

Bountiful  
Draper  
Ogden  
Salt Lake City

## Vermont

South Burlington

## Virginia

Alexandria  
Chesapeake  
Glen Allen  
Lynchburg  
Newport News  
Reston  
Richmond  
Roanoke

## Washington

Renton  
Seattle  
Spokane Valley  
Yakima

## West Virginia

Charleston  
Morgantown

## Wisconsin

Ashwaubenon  
Brookfield  
Eau Claire  
Kenosha  
Madison  
Milwaukee

## Wyoming

Casper

## U.S. Territories

Amer. Samoa  
Pago Pago  
Guam/Tamuning  
N. Mariana  
Islands/Saipan  
Puerto Rico/  
San Juan  
Virgin Islands/  
St. Thomas

## Canada

Calgary, AB  
Edmonton, AB  
Burnaby, BC  
Surrey, BC  
Winnipeg, MB  
Halifax, NS  
Hamilton, ON  
London, ON  
Ottawa, ON  
Regina, SK  
St John's, NL  
Toronto, ON  
Montreal, QU  
Saskatoon, SK  
Vancouver, BC  
Victoria, BC

## International

**Asia/Pacific:**  
Sydney,  
Australia  
Hong Kong,  
Hong Kong  
Mumbai, India  
Osaka, Japan  
Tokyo, Japan  
Seoul, Korea  
Manila,  
Philippines

## Europe:

London,  
England  
Frankfurt,  
Germany  
Istanbul,  
Turkey

## EXAMINATIONS IN RADIOLOGIC TECHNOLOGY

### STATE CANDIDATE STATUS REPORT

Please review the following information very carefully and contact your state licensing agency with any corrections.

*Please read your handbook for complete examination details.*

**YOU MUST USE THE ID NUMBER BELOW WHEN SCHEDULING YOUR APPOINTMENT WITH PEARSON VUE**

**JOHN Q PUBLIC  
APARTMENT 1  
MAIN STREET  
ANYTOWN, USA 00000**

**DATE: 04/15/2020**

**ID#: 999999**

**SOCIAL SECURITY NUMBER: 123-45-6789  
BIRTHDATE: 05/17/1979**

**FOR THE STATE OF: YOUR STATE  
DIRECT QUESTIONS TO: (555) 999-9999**

**EXAMINATION CATEGORY: RADIOGRAPHY  
WINDOW START DATE: 04/22/2020  
WINDOW END DATE: 07/20/2020**

**NUMBER OF PRIOR STATE ATTEMPTS COUNTED  
IN ARRT THREE-ATTEMPT LIMIT: 1**

You have been assigned to take the examination indicated above based upon information you supplied to your state licensing agency. *Please review the above information carefully and contact your state licensing agency at the number listed above if there are any corrections or changes before scheduling your exam.*

At the test center, you will be required to show two forms of identification. One must be a government-issued ID which contains a permanently affixed photo along with a signature and must not be expired. The second ID must contain your pre-printed name and signature and must not be expired. The names appearing on both IDs must match the name appearing at the top of this status report. If your name has a cultural variation, make sure the same variation appears above and on both IDs. Please see the list of acceptable IDs and name requirements in your *Examination Handbook*. Test center administrators have been instructed not to admit anyone to the test center not having the required suitable IDs. Fees will not be refunded if you are denied admission to the test center for failure to provide suitable identification.

**Exam attempts as a state licensing candidate count toward the three-attempt limit for ARRT certification in the same discipline (e.g., radiography, nuclear medicine technology, radiation therapy, or a postprimary exam). Passing an exam as a state candidate counts as an attempt, but the passing score cannot be considered for future ARRT certification and registration. Contact ARRT for further information at (651) 687-0048 ext. 8560.**

- *Please direct all questions and personal information changes to your state licensing agency at the number listed above.*
- *Your score from this examination cannot be used for current or future certification and registration by the ARRT.*
- *Your score from this examination is valid only for state licensing purposes.*
- *Your exam results will be provided by your state licensing agency. Do not contact ARRT for your exam results.*

**See Reverse Side for Instructions on Scheduling Your Appointment**

(12/19)

To schedule, confirm, change, or cancel your examination date, time, or location - Call Pearson VUE at 1-800-632-9055

**Record Your Exam Scheduling Information Here**

- |                               |                      |
|-------------------------------|----------------------|
| • Call Center Representative: | Confirmation Number: |
| • Date:                       | Time:                |

**Scheduling or Changing Your Appointment**

It is your responsibility to contact Pearson VUE to schedule the date, time, and location of your exam. Your exam must be completed between the assigned exam window dates printed on this *Candidate Status Report*. If you fail to complete your exam during your assigned exam window, your file will close, and you will need to contact the FL Rad Tech Certification Office for new eligibility information.

Please call the Pearson VUE Call Center at **1-800-632-9055** to schedule your appointment. You may also schedule your appointment via the Internet at [www.pearsonvue.com/arrt](http://www.pearsonvue.com/arrt), where you will have to provide a return e-mail address. Shortly after scheduling your appointment, Pearson VUE will send an email confirmation letter to you listing your appointment time and date, testing center location, and directions to the testing center. See your *Examination Handbook* for appointment scheduling and confirmation information.

If you find it necessary to change your examination appointment, you must first call Pearson VUE to cancel your existing appointment in accordance with the guidelines printed in your *Examination Handbook* before requesting a new exam date or making changes in the test center location. Pearson VUE will charge a fee for each canceled or rescheduled appointment. (See your *Examination Handbook* for complete details.)

**Changing Your ARRT 90-Day Examination Window Dates**

If it is necessary to change your ARRT 90-day examination window, you must first call Pearson VUE to cancel your existing appointment BEFORE requesting an examination window change with the ARRT. Window dates cannot be changed if an appointment is scheduled. A completed Window Extension Request Form must be received at ARRT for approval on or before the last day of your current ARRT 90-day window. Window changes will NOT be extended beyond your current FL Rad Tech Certification Office license eligibility period listed on the front of this CSR. (See *Examination Handbook* for complete details.)

**Calculators**

Personal calculators are prohibited for examinations in all disciplines. You may use the basic 4-function calculator or scientific calculator provided on the computer or you may request a hand-held, basic 4-function calculator from the test center administrator.

**Results**

Examination results are not given at the test center or provided by the ARRT under any circumstances. Examination results will be posted on the FL Rad Tech Certification Office website. Please allow four weeks for posting of examination scores. If results do not appear on your state website within four weeks, please contact the FL Rad Tech Certification Office, not the ARRT.

**Appeals**

You must notify ARRT in writing of any negative situations that may have affected your exam performance by submitting a completed Exam Administration Appeal Form (located at [StateRHC.org](http://StateRHC.org)) within two days of your exam. ARRT will not investigate complaints it receives after results have been processed. You must fax your appeal to (651) 681-3295. (See *Examination Handbook* for complete details.)

**Notice of Possible Changes to Exam Content Specifications**

If you delay taking the exam after you receive this CSR, be aware that we periodically update the exam content specifications. You might need to prepare for new content on the exam. You can find the current exam content specifications at [www.StateRHC.org](http://www.StateRHC.org).

**PLEASE DIRECT ALL PERSONAL INFORMATION CHANGES TO THE FL RAD TECH CERTIFICATION OFFICE AT (850) 488-0595 AND BEFORE SCHEDULING AN APPOINTMENT**

**IF YOUR FL RAD TECH CERTIFICATION OFFICE ELIGIBILITY PERIOD HAS ENDED, YOU MUST CONTACT THEM DIRECTLY AT (850) 488-0595 FOR ELIGIBILITY INFORMATION**

(12/19)



To schedule, confirm, change, or cancel your examination date, time, or location - Call Pearson VUE at 1-800-632-9055

**Record Your Exam Scheduling Information Here**

- |                               |                      |
|-------------------------------|----------------------|
| • Call Center Representative: | Confirmation Number: |
| • Date:                       | Time:                |

**Scheduling or Changing Your Appointment**

It is your responsibility to contact Pearson VUE to schedule the date, time, and location of your exam. Your exam must be completed between the assigned exam window dates printed on this *Candidate Status Report*. If you fail to complete your exam during your assigned exam window, your file will close, and you will need to contact the FL Rad Tech Certification Office for new eligibility information.

Please call the Pearson VUE Call Center at **1-800-632-9055** to schedule your appointment. You may also schedule your appointment via the Internet at [www.pearsonvue.com/arrt](http://www.pearsonvue.com/arrt), where you will have to provide a return e-mail address. Shortly after scheduling your appointment, Pearson VUE will send an email confirmation letter to you listing your appointment time and date, testing center location, and directions to the testing center. See your *Examination Handbook* for appointment scheduling and confirmation information.

If you find it necessary to change your examination appointment, you must first call Pearson VUE to cancel your existing appointment in accordance with the guidelines printed in your *Examination Handbook* before requesting a new exam date or making changes in the test center location. Pearson VUE will charge a fee for each canceled or rescheduled appointment. (See your *Examination Handbook* for complete details.)

**Changing Your ARRT 90-Day Examination Window Dates**

If it is necessary to change your ARRT 90-day examination window, you must first call Pearson VUE to cancel your existing appointment BEFORE requesting an examination window change with the ARRT. Window dates cannot be changed if an appointment is scheduled. A completed Window Extension Request Form must be received at ARRT for approval on or before the last day of your current ARRT 90-day window. Window changes will NOT be extended beyond your current FL Rad Tech Certification Office license eligibility period listed on the front of this CSR. (See *Examination Handbook* for complete details.)

**Calculators**

Personal calculators are prohibited for examinations in all disciplines. You may use the basic 4-function calculator or scientific calculator provided on the computer or you may request a hand-held, basic 4-function calculator from the test center administrator.

**Results**

Examination results are not given at the test center or provided by the ARRT under any circumstances. Examination results will be posted on the FL Rad Tech Certification Office website. Please allow four weeks for posting of examination scores. If results do not appear on your state website within four weeks, please contact the FL Rad Tech Certification Office, not the ARRT.

**Appeals**

You must notify ARRT in writing of any negative situations that may have affected your exam performance by submitting a completed Exam Administration Appeal Form (located at [StateRHC.org](http://StateRHC.org)) within two days of your exam. ARRT will not investigate complaints it receives after results have been processed. You must fax your appeal to (651) 681-3295. (See *Examination Handbook* for complete details.)

**Notice of Possible Changes to Exam Content Specifications**

If you delay taking the exam after you receive this CSR, be aware that we periodically update the exam content specifications. You might need to prepare for new content on the exam. You can find the current exam content specifications at [www.StateRHC.org](http://www.StateRHC.org).

**PLEASE DIRECT ALL PERSONAL INFORMATION CHANGES TO THE FL RAD TECH CERTIFICATION OFFICE AT (850) 488-0595 AND BEFORE SCHEDULING AN APPOINTMENT**

**IF YOUR FL RAD TECH CERTIFICATION OFFICE ELIGIBILITY PERIOD HAS ENDED, YOU MUST CONTACT THEM DIRECTLY AT (850) 488-0595 FOR ELIGIBILITY INFORMATION**

(12/19)

## Computer-Based Testing Overview

After you have completed check-in procedures, test-center staff will show you to a work station and will make sure the computer is ready to deliver your exam. The testing session consists of four segments:

- 1. Introduction, Tutorial, and Non-Disclosure Agreement:** During this segment, the computer will verify your name and allow you to complete a tutorial if you choose. We strongly urge candidates to spend the few minutes to take the tutorial. You will also be asked to read and accept a non-disclosure agreement – it requires that all candidates agree to **not** copy any test questions or otherwise disclose the content of the exam. You must accept the terms of the non-disclosure agreement; if you do not respond within 2 minutes your exam session will end. The entire introductory segment will take anywhere from a few minutes up to 20 minutes, depending on how much time you spend reviewing the tutorial.
- 2. Examination Session:** You will be given the exam during this period. In addition to answering questions, you can mark questions for later review or even comment on questions. The clock will be running, so pace yourself. Most questions are in the standard multiple-choice format and require you to select one best answer. In addition, a small portion of the exam may consist of the question formats noted below:

- a. Select Multiple:** This format consists of a question or statement followed by a list of 4 to 10 response options. You are required to select all options that are correct.
- b. Sorted List:** This format presents a list of 4 to 8 options and requires you to place them in correct sequence. You accomplish this by using the mouse to “click-and-drag” the options into a box so that they end up in a specified order, such as numerical, alphabetical or chronological.
- c. Items with Hot Areas or Videos:** This format consists of a question accompanied by a medical image, drawing, graphic, or video.

To answer a ‘hot area’ question, place the cursor over the selected area and click the mouse; the highlighted areas are possible answers to the question. When selected, the area will become outlined and change color. To change your answer, move the mouse to another shaded area and click the mouse. The final selected shaded area will be recorded as your final answer.

For video items, you will need to read the question, open the exhibit, press the play arrow on the video, watch the video in its entirety, and then answer the question. You will not be able to move forward on the exam until you have opened and watched the entire video. The video controls are shown and described below. **Note: The videos are silent (no sound).**



Sample questions illustrating these formats can be obtained from the ARRT website (Examinations>Exam Format and Length), or by contacting the ARRT. In addition, the tutorial at the test center presents an example of each format.

- 3. Item Review and End Review:** After responding to all questions, you will have the opportunity to go back and review questions in the time remaining. You can change answers during the review. Once you select the “End Review” button you will no longer be able to go back to the exam. A sample review screen appears later in this Appendix.
- 4. Survey:** After the exam a short survey consisting of 13 questions will appear. Most people complete it in just a few minutes. The survey is important because it gives you the opportunity to let ARRT know about the quality of your testing experience. If something went wrong – or exceptionally right – this is the place to tell us.

The following pages illustrate the approximate appearance of a few of the more important computer screens. Taking a few minutes now to review these pages will help prepare you for exam day.

### Appearance of Test Questions

When the examination starts, the clock will be reset to the time allowed for the exam you are taking (see Exam Timing under the Exam Administration Day Section of the handbook to find the time allotted for your exam). Exam questions are presented in random order. The exam consists of a set number of scored questions plus several unscored pilot questions. The content specifications provide additional information about the number of questions and topics covered.

The screenshot shows a testing interface with a dark grey header bar. On the left, there are icons for 'Comment' and 'Calculator'. On the right, there is a clock showing 'Time Remaining 03:29:20', a counter showing '3 of 180', and an 'Eflag for Review' icon. Below the header, the main content area displays a question: 'The Alamo is located in the state labeled as number.' To the left of the question are four radio button options: A 1, B 2, C 3, and D 4. To the right is a map of Texas divided into four numbered regions (1, 2, 3, 4). At the bottom of the screen, there are 'Previous' and 'Next' navigation buttons. Five callout boxes provide explanations for these elements.

This button allows you to mark questions for later review. If uncertain of the best answer, then choose your best guess and flag the question for later review by clicking on the flag icon.

The clock indicates the time left to complete the exam.

You can comment on specific exam questions by clicking on the "Comment" button. The "Calculator" button gives access to an on-screen calculator (see next page).

The counter indicates which question you are on and the total number of questions on the exam.


Here is the exam question. Choose one best answer by clicking the appropriate oval or letter (A, B, C, D). If the question requires a graphic, it will also appear on the screen.

Click on these buttons to go back to the previous question or ahead to the next one.

### Online Calculator

To use the calculator, click on the “Calculator” button at the upper left side of the exam screen. You can operate the calculator by using the mouse to click on numbers or arithmetic operations. Alternatively, the keyboard can be used. **Note: Please make sure to check the display screen on the calculator to verify the correct entry of numbers.**

The “Modes” button on the calculator allows you to toggle between the Standard and Scientific calculators. Note that most calculations on the exam can be done with the Standard calculator. However, some candidates may wish to use the Scientific calculator for certain calculations.



## Exam Review

After you have completed all questions on the exam, a screen appears that allows you to go back to review questions. A filled-in flag icon appears next to any questions that you selected for review.

The screenshot shows the 'Exam Review' screen. At the top right, it says 'Time Remaining 03:25:44'. Below that is the title 'Exam Review'. Underneath is an 'Instructions' section with two numbered steps: 1. Review all of your questions and answers. 2. Review questions that are flagged for review. (Click the "flag" icon to change the review status.) A note below states: 'Although the "Review Incomplete" button appears, this button is not functional; all questions on the exam require an answer.' The main area is a grid titled 'Computed Tomography Section (0 Unseen/Incomplete)' containing 36 questions. Each question has a flag icon; Question 8 and Question 33 have filled-in flag icons. At the bottom, there is a control bar with four buttons: 'End Review', 'Review All', 'Review Incomplete', and 'Review Flagged'.

This button ends the exam. When you are done with your review, click this button to exit.

Once you click "End Review" you will no longer be able to review questions or change answers, **so be sure you are really ready to stop!**

You can return and review all questions on the exam by clicking on the "Review All" button.

You can return to the questions you selected for review by clicking on the "Review Flagged" button. To review all items on the exam, just click on "Review All."

If you click this button you will see that you have no incomplete questions, because skipping of questions is not an option on ARRT exams.

### After the Examination

After you click "End Review" and confirm that you will not be able to return to the exam, a screen will appear to remind you not to discuss questions and/or answers with anyone.

A short survey appears on the screen. It asks a few important questions about the quality of the test administration and provides a place for you to type any general comments. We appreciate your feedback.

## Computer-Based Testing Overview

After you have completed check-in procedures, test-center staff will show you to a work station and will make sure the computer is ready to deliver your exam. The testing session consists of four segments:

1. **Introduction, Tutorial, and Non-Disclosure Agreement:** During this segment, the computer will verify your name and allow you to complete a tutorial if you choose. We strongly urge candidates to spend the few minutes to take the tutorial. You will also be asked to read and accept a non-disclosure agreement – it requires that all candidates agree to **not** copy any test questions or otherwise disclose the content of the exam. You must accept the terms of the non-disclosure agreement; if you do not respond within 2 minutes your exam session will end. The entire introductory segment will take anywhere from a few minutes up to 20 minutes, depending on how much time you spend reviewing the tutorial.
2. **Examination Session – Modules:** The Limited Scope of Practice in Radiography Exam is delivered in modules. The modules are Core, Chest, Extremities, Skull/Sinuses, Spine, and Podiatric (refer to the Content Specifications for details). Candidates may take some or all modules, depending on the type of license offered by your state.

Which Modules. The computer will present **only** those modules that were assigned to you by your state licensing agency. Those same modules are printed on your *Candidate Status Report*.

Time Allowed. Each module is separately timed. The amount of time is determined by the number of questions in a module, at a rate of 1 minute per question. For example, the Core module has 115 questions, so you have up to 115 minutes to complete the Core module. The Core module includes 15 unscored (pilot) questions. The Chest module has 25 questions, and 25 minutes are allowed to complete that module. Each of the radiographic procedure modules include five additional unscored questions. It is important to pace yourself so that you complete each module within the allotted time.

Review Session. The computer requires that you answer every question. If you are unsure of an answer to a question, you can “mark” the question and come back to it later. After you have answered all questions in a module, a review screen allows you to go back to any questions you marked. You can change answers during the review. When done reviewing questions, you can end the module. Extra time is not given for the review session; it must be completed during the time allowed for each module. A sample review screen is presented later in this Appendix.

3. **Item Review and End Review:** After responding to all questions within a module, you will have the opportunity to go back and review questions in the time remaining. You can change answers during the review. Once you select the “End Review” button, the module ends and you will no longer be able to go back and review questions in that module. At this point, one of two things happen: (1) If you have additional modules to complete, the next module will appear; (2) If you do not have additional modules to complete, the exam ends. A sample review screen appears later in this Appendix.
4. **Survey:** After the exam a short survey consisting of 13 questions will appear. Most people complete it in just a few minutes. The survey is important because it gives you the opportunity to let ARRT know about the quality of your testing experience. If something went wrong – or exceptionally right – this is the place to tell us.

The following pages illustrate the approximate appearance of a few of the more important computer screens. Taking a few minutes now to review these pages will help prepare you for exam day.

### Appearance of Test Questions

When the examination starts, the clock will be reset to the time allowed for the module you are taking. Both the scored and unscored exam questions are presented in random order within each module. The content specifications provide additional information about the number of questions and topics covered.

The screenshot shows a testing interface with a dark grey header bar. On the left, there are icons for 'Comment' and 'Calculator'. On the right, there is a clock icon, the text 'Time Remaining 03:29:20', a counter '3 of 180', and an 'Flag for Review' icon. Below the header, the main content area contains a multiple-choice question: 'The Alamo is located in the state labeled as number.' followed by four radio button options: 'A 1', 'B 2', 'C 3', and 'D 4'. To the right of the question is a map of Texas divided into four numbered regions (1, 2, 3, 4). At the bottom of the screen, there are 'Previous' and 'Next' navigation buttons. Five callout boxes with lines pointing to specific elements provide the following explanations:

- Top-left callout:** This button allows you to mark questions for later review. If uncertain of the best answer, then choose your best guess and flag the question for later review by clicking on the flag icon.
- Top-right callout:** The clock indicates the time left to complete the module.
- Bottom-left callout:** You can comment on specific exam questions by clicking on the "Comment" button. The "Calculator" button gives access to an on-screen calculator (see next page).
- Bottom-right callout:** The counter indicates which question you are on and the total number of questions in the module you are in.
- Bottom-center callout:** Here is the exam question. Choose one best answer by clicking the appropriate oval or letter (A, B, C, D). If the question requires a graphic, it will also appear on the screen.
- Bottom-right callout:** Click on these buttons to go back to the previous question or ahead to the next one.

### Online Calculator

To use the calculator, click on the “Calculator” button at the upper left side of the exam screen. You can operate the calculator by using the mouse to click on numbers or arithmetic operations. Alternatively, the keyboard can be used. **Note: Please make sure to check the display screen on the calculator to verify the correct entry of numbers.**

The "Modes" button on the calculator allows you to toggle between the Standard and Scientific calculators. Note that most calculations on the exam can be done with the Standard calculator. However, some candidates may wish to use the Scientific calculator for certain calculations.



Some calculations may require the use of the natural logarithm function (“ln” key) or the  $e^x$  function (“2nd” key, then “ln” key). First press the key for the function that you would like, then enter the relevant number for the calculation.

## Exam Review

After you have completed all questions in a module, a screen appears that allows you to go back to review questions. A filled-in flag icon appears next to any questions that you selected for review.

The screenshot shows the 'Exam Review' screen. At the top right, a timer indicates 'Time Remaining 03:25:44'. Below the title 'Exam Review', there is an 'Instructions' section with two numbered steps: 1. Review all of your questions and answers. 2. Review questions that are flagged for review. (Click the "flag" icon to change the review status.) A note below states: 'Note: Although the "Review Incomplete" button appears, this button is not functional; all questions on the exam require an answer.' The main area is a grid titled 'Computed Tomography Section' with '(0 Unseen/Incomplete)' in the top right. The grid contains 36 questions, numbered 1 through 36, arranged in a 12x3 layout. Each question has a flag icon to its left. Questions 8, 19, and 33 have filled-in flag icons, while all others have empty flag icons. At the bottom of the grid, there is a navigation bar with four buttons: 'End Review' (with a right-pointing arrow), 'Review All' (with a flag icon), 'Review Incomplete' (with an 'X' icon), and 'Review Flagged' (with a flag icon).

This button ends the module. When you are done with your review, click this button to exit.

Once you click "End Review" you will no longer be able to review questions or change answers within that module, **so be sure you are really ready to stop!**

You can return and review all questions within the module by clicking on the "Review All" button.

You can return to the questions you selected for review by clicking on the "Review Flagged" button. To review all items within the module, just click on "Review All."

If you click this button you will see that you have no incomplete questions, because skipping of questions is not an option on ARRT exams.

### After the Examination

After you click "End Review" (at the end of your last module) and confirm that you will not be able to return to the exam, a screen will appear to remind you not to discuss questions and/or answers with anyone.

A short survey appears on the screen. It asks a few important questions about the quality of the test administration and provides a place for you to type any general comments. We appreciate your feedback.

# Potential Exam Disclosure Scenarios

| Scenario   | When it's OK  | When it's not OK  | Bottom line  |
|--|---|---|--|
| Educator asks candidates to “stop by” after the exam to “let me know how it went.”   | If the invitation and the feedback to the educator relates to their general experience (“I thought the test was not as difficult as I expected...”).  | This type of invitation from an educator may be misinterpreted by the candidate — and the student may think that the educator is asking the candidate to reveal copyrighted information.  | If the candidate is asked to reveal ARRT’s questions or their answer options, then he or she will need to report the educator to the ARRT Ethics Committee. The educator should stop the candidate immediately from revealing any exam content, since doing so may subject both the candidate and educator to ARRT’s ethics process.   |
| Candidate tells another candidate, “The test was very difficult — I felt like I didn’t have enough time.”                                    | The candidate is simply telling another candidate how they felt about the exam. This is all right because the candidate is not revealing any of ARRT’s questions or the answer options.               | One candidate (or potential candidate) asks another candidate about the specific questions.   | If ARRT’s questions or answer options are shared, these individuals may find themselves part of an ARRT ethics investigation and/or legal complaint.   |
| Candidate to educator: “You didn’t teach me about this question that asked [specific question]. I felt unprepared.”                          | Never.  | It is not all right and it will never be all right to reveal ARRT’s copyrighted questions (or answer options) to anyone.  | Candidates sign numerous documents stating that they will not share exam questions, and ARRT expects the candidates to abide by those contracts. Those who don’t may find themselves part of an ARRT ethics investigation and/or legal complaint.  |
| Candidate tells a potential candidate that there were multiple-choice and sorted-list questions on the test.                                 | This is public information, noted in the certification and registration handbook.   | It’s not all right to reveal anything beyond what’s in the handbook.  | Keep the conversation limited to what’s public information, such as the content specifications, and there’s no problem.  |
| Candidate asks another candidate, “I don’t think that I understood this question...[relates question]... Do you know what they were asking?” | Never.  | It is not all right and it will never be all right to reveal ARRT’s copyrighted questions (or answer options) to anyone.  | As noted two boxes up, candidates sign numerous documents stating that they will not share exam questions, and ARRT expects the candidates to abide by those contracts. Those who don’t may find themselves part of an ARRT ethics investigation and/or legal complaint.   |
| Candidate says to a potential candidate, “If I were you, I would bring a sweater — it was cold at the test site.”                            | This candidate is simply telling another candidate about their surroundings at the test site. This is all right because the candidate is not revealing any of ARRT’s questions or the answer options. | If it leads a candidate (or potential candidate) to ask another candidate about the specific questions.   | If ARRT’s questions or answer options are shared, these individuals may find themselves part of an ARRT ethics investigation and/or legal complaint.   |
| Potential candidate says to a candidate, “Were there a lot of questions on [specific topic]?”  | Never.  | This candidate should be aware of the topics that are contained in the exam from the content specifications published in the certification and registration handbooks and should not be asking for more specific information than is contained in that publication. | If the potential candidate is asking the candidate to reveal ARRT’s questions or the answer options, then this conversation violates both the <i>ARRT Standards of Ethics</i> and the legal contract that both the candidate and the potential candidate have signed. If asked this type of question, the potential candidate should be shown the content specifications and should be warned of the consequences of revealing ARRT’s copyrighted questions or their answer options. |



## ARRT Rules Agreement

Please review the following information and ask the Test Administrator if you have questions.

|     |  |
|-----|--|
| 1.  | ARRT has a zero-tolerance policy regarding possession of cell phones and other electronic devices at the test center. If you are found to be in possession of, or otherwise have access to, one of these devices after initial check-in (including during scheduled or unscheduled breaks), you will not be allowed to resume your exam or assessment, you will forfeit your exam or assessment fee, your score will be canceled, and it will count as an attempt in your three-attempt, three-year time period. For SSA participants, you will be assigned the full prescription for your discipline.<br>Should you bring an electronic device into the test center, you must turn off the device and store it in one of the test center's lockers before you enter the testing room.<br>Do not access your electronic device again until you have fully completed your exam or assessment. |
| 2.  | Jewelry that is wider than 1/4 in (1 cm) is not permitted inside the testing room, and you will be asked to remove it.   |
| 3.  | Do not use the noteboard provided by the Test Administrator until after you have responded to the Non-Disclosure Agreement. If you need a clean noteboard during the exam or assessment, you should raise your hand to get the Test Administrator's assistance. Return all items to the Test Administrator after completing your exam or assessment.   |
| 4.  | Eating, drinking, smoking, chewing gum, and making noise that creates a disturbance for other candidates is prohibited during the exam or assessment.  |
| 5.  | The Test Administrator will monitor you continuously while you complete your exam or assessment. The session may be videotaped or otherwise recorded for security or other purposes.   |
| 6.  | If you experience problems that affect your ability to complete your exam or assessment, notify the Test Administrator immediately by raising your hand. The Test Administrator cannot answer questions related to exam or assessment content and performance.   |
| 7.  | To request an unscheduled break, you must raise your hand to get the Test Administrator's attention. The exam or assessment timer will not stop while you are on an unscheduled break. The Test Administrator will sign you out after you leave the testing room. Before returning to your seat, the Test Administrator will sign you in; after being signed in, you may resume your exam or assessment.   |
| 8.  | You should not remove any items from your secure locker. If you must access a personal item, such as an item needed to take to the restroom, this is allowed after notifying the Test Administrator. However, if you access any other prohibited item from the secure locker (cell phone, books, notes, etc.), your score will be canceled, your testing fees will not be refunded, and it will count as an attempt in your three-attempt, three-year time period.<br><b>Note:</b> During scheduled breaks, RA and sonography candidates may access their locker in order to retrieve snacks. You may not leave the test center building during your scheduled break.  |
| 9.  | You may not leave the building for any reason (unless directed to leave by the Test Administrator); this includes all breaks. If you leave the building you will not be allowed to resume your exam or assessment, you will forfeit your exam or assessment fee, and your score will be canceled. The exam will count as an attempt in your three-attempt, three-year period. For SSA participants, you will be assigned the full prescription for your discipline.  |
| 10. | Do not remove copies of exam or assessment questions and answers from the testing room (including by writing on your person or clothing). Do not share exam or assessment questions and answers with anyone. Reproduction of exam or assessment questions and answers, in whole or part, constitutes a breach of your agreement, and you can/will be prosecuted in federal or state court. Depending upon your candidate or participant status, this will also result in score cancellation, future certification and registration ineligibility, and/or discontinuation of your certification and registration.   |
| 11. | After completing your exam or assessment, raise your hand. The Test Administrator will come to your workstation to ensure your exam or assessment has ended properly and will escort you from the testing room.  |
| 12. | If you do not follow the rules, are suspected of cheating or tampering with the computer, or demonstrate irregular behavior the issue will be reported to Pearson VUE, the ARRT, and your state licensing agency (if applicable). Your exam or assessment may be invalidated, the ARRT may take other action such as canceling your score, and you will not be refunded your exam or assessment fee.   |

**Candidate/Participant Statement:** *By providing a digital signature, I give Pearson VUE my explicit consent to retain and transmit my personal data and test responses to the Pearson VUE corporate office and the ARRT (either of which may be outside of the country in which I am testing). I understand the information provided above and agree to abide by the ARRT Rules Agreement. In addition, I understand that if I am found to be in violation of any rule listed above, this will constitute grounds for the ARRT to take appropriate punitive action up to and including terminating my participation in the exam or assessment, invalidating the results of this exam or assessment and any prior exam or assessment, and permanently barring me from all future exams or assessments. In addition, I understand I may be subject to an ARRT ethics investigation or even a federal court lawsuit for copyright infringement and/or breach of contract. Any information collected by an ARRT investigation may be forwarded to my state licensing agency for review of state ethics violations.*

ARRT Candidate Rules Agreement

Version 3.6 / October 2019

# Florida State Certification Examination Handbook Checklist

## **When you receive your Candidate Status Report (CSR) from ARRT...and before scheduling your exam you will want to check...**

- Does your name on your CSR match the name appearing on your two forms of required ID?
  - If your names do not match, do not schedule an appointment. Contact the FL Rad Tech Certification Office to make the necessary changes and have them notify ARRT so we can mail you a new CSR with your updated info.
  - Once you verify the changes to your CSR are correct, go ahead and schedule your exam.
- Name or address change after you receive your information from ARRT?
  - All changes must be made via the FL Rad Tech Certification Office.
- Be sure to note the different dates on your CSR. Your 90-day exam window is different than your 180-day FL Rad Tech Certification Office eligibility period.
  - You must schedule your exam for a time within the 90-day exam window printed on your CSR
  - Your 180-day FL Rad Tech Certification Office eligibility period allows you to complete 1 attempt within the 180-day period listed on your CSR.
- If you can't take your exam within your 90-day exam window, you can extend your window if you have time remaining in your 180-day FL Rad Tech Certification Office eligibility period.
  - Cancel any existing appointment.
  - Print the window extension request form located at [www.staterhc.org](http://www.staterhc.org).
  - Fax to ARRT at 651.681.3294 before the last day of your existing 90-day window.
  - A new CSR will be mailed to you once the request has been processed at ARRT.
- Required IDs at the test center.
  - Make sure your IDs meet ARRT's requirements listed in the handbook to prevent being turned away from the test center and losing your fee.
  - If you are unsure, cancel your appointment and reschedule when you are certain your IDs will be acceptable.
- Questions on exam results?
  - ARRT processes results each week and provides your score information to the FL Rad Tech Certification Office.
  - The FL Rad Tech Certification Office determines your pass/fail status, not ARRT.
  - Please allow up to 45 days for the FL Rad Tech Certification Office to post your results on their website.
  - Contact information for the FL Rad Tech Certification Office is on the inside cover of this handbook.

***Important Notice: State Licensing is Not ARRT Credentialing***

*A passing score on a state licensing examination does not make a candidate eligible for ARRT certification and registration. Candidates seeking ARRT certification and registration must have submitted an application directly to ARRT and must have met all other criteria for ARRT certification and registration. Those wishing only state licensing must meet criteria established by the state. Test scores earned as a state candidate may not be used for ARRT certification and registration.*

---

Direct questions regarding your state license application, the FL-DOH Rad Tech Certification Office 180-day eligibility period, or changes to your name, address, social security number, or date of birth to:

**Florida Department of Health**

**Radiologic Technology Certification Office**

**4052 Bald Cypress Way Bin C85**

**Tallahassee FL 32399-3285**

**Phone: (850) 488-0595**

**E-mail: [MQA\\_Rad-Tech@flhealth.gov](mailto:MQA_Rad-Tech@flhealth.gov)**

**Website: [www.floridahealth.gov/licensing-and-regulation/radiologic-technology/](http://www.floridahealth.gov/licensing-and-regulation/radiologic-technology/)**