10MHz / 5MHz / 1MHz PRECISION LCR METER



CE	RS-232	GPIB	PC Software
			-

The Smarter Way to Characterize Components

The LCR-8000G Series LCR meter, with test frequency up to 10MHz, provides accuracy, versatility and high resolution for a wide range of component measurements, even including DC resistance measurement and Voltage/Current monitoring. The Multi-Step function allows on-screen programming of customized measurement sequence with Pass/Fail indication. Each program includes 30 test steps and each test step can be set with selected parameters and test limits. Under Multi-Step operation, a tedious work routine can be done step by step automatically just at a press of a button. With Graph Mode, LCR-8110G, LCR-8105G and LCR-8101G display the component impedance response either over a wide range of test requency sweep or over a wide range of test voltage sweep in a graph chart. This gives an analysis result of either impedance vs. frequency or impedance vs. applied voltage all at a glance. GPIB and RS-232C interfaces are available as standard for instrument control and test result display on the PC. The rich features of LCR-8000G Series easily make your measurement tasks done at a very competitive price.

Wide Frequency Range With Friendly Intuitive User Interface

The LCR-8000G Series is designed to perform precision impedance measurements over a wide frequency range of 20Hz ~ 10MHz for LCR-8110G, 20Hz ~ 5MHz for LCR-8105G and 20Hz ~1MHz for LCR-8101G. The instrument is capable of measuring 11 different parameters with 0.1% basic accuracy, which meets the precision measurement requirements of components and modules used in the RF circuits. The large LCD display with single-layer operation menu of LCR-8000G Series provides users with ultimate convenience to plug and play without much learning time.

Multi Step Mode

The Multi-Step mode is capable of running a series of measurements of a component at a number of userdefined steps in sequence automatically. Total 64 programs can be saved into the non-volatile memory, and each program contains up to 30 test steps. The parameter and Hi/Lo limits can be set respectively for each test step. After a program being properly edited, the instrument can run through all the measurement steps either at a press of the button under the Manual Trigger selection, or automatically run the program by detecting the connection of a DUT under the Auto Trigger selection. When all the test steps are completed, the screen shows the measurement reading of the parameter being selected for each step with Pass, HI, or LO measurement result.

Graph Mode

The graph function shows the component characteristics in visual manner. Either voltage sweep or frequency sweep can be selected for horizontal scale. Just select the parameter, and set the start/stop voltage or the start/stop frequency of the sweep, LCR -8000G will run through the sequence of measurements and show the results on a graph. This graphical parameter measurement performs the characteristic verification of components and materials over the response to the changes in AC test frequency or AC test voltage without the need of complex programming of an external controller. When the graph gets out of the vertical range, the LCR-8000G Series can automatically adjust the scale to get full test information. On the graph the marker operation is available for detailed observation.

Pass/fail Function With Judgment Alarm

In the Pass/Fail test, primary parameter measurement result is compared with user-defined limits and the pass or fail result is then displayed. Three methods, including absolute limit, percentage and delta limit, are available for Hi(high) and Lo(low) limit setting based on the nominal test value. The Pass/Fail test checks whether the parameter measurement result sits within the limits, then display "PASS" for within the limits, or "LO" for lower than the low limit, or "HI" For higher than the high limit. A scale and bar for displaying measurement result is shown at the center of the screen to give a graphical identification, which greatly reduces operator's load in a long time inspection work. This scale and bar also facilitate the adjustment of the variable components. Either Pass or Fail result can be set with a buzzer alarm, which makes component or material sorting easy with sound identification.

LCR-8000G Series

FEATURES

- Wide Test Frequency 20Hz~10/5/1MHz
- 0.1% Basic Accuracy & 6 Digits Measurement Resolution
- Large LCD Display with Intuitive User Interface
- Full Measuring Functions with DUT V/I Monitor
- PASS/FAIL Function (abs , % , \triangle) with Judgment Alarm
- Average 1 ~ 256 Times
- DC Resistance Measurement
- Multi Step Mode
- Graph Mode
- Standard RS-232C / GPIB Interface

APPLICATIONS

- Education Lab and Training Institution
- Research & Design

• Quality Assurance Verification



LCR-8000G Series

`							
	20112 ~ 10101112/310112/	1MHz, 5 Digits, ±0.0059	%				
	100Ω						
BASIC ACCURACY *							
	AC (>2kHz) - MAX: 75ms, FAST: 150ms, MEDIUM: 450ms, SLOW: 600ms DC - MAX: 30ms, FAST: 60ms, MEDIUM: 120ms, SLOW: 900ms						
TEST SIGNAL LEVELS	Test Frequency ≤ 3MHz > 3MHz	Test Signal Level (rms) 10mV ~ 2V 10mV ~ 1V	Step 1mV/10mV 1mV/10mV	Accuracy 2%± 5mV 2%± 5mV			
SHORT CIRCUIT CURRENT	Max. 20mA						
MEASUREMENT RANGES	Mode R, Z, X, Rdc G, Y, B L C D Q θ		$\begin{tabular}{lllllllllllllllllllllllllllllllllll$				
	Impedance (Ζ), Phase Angle (θ), Inductance (L), Capacitance (C), AC Resistance (Rac), Quality Factor (Q), Dissipation Factor (D), Admittance (Y), Conductance (G), Reactance (X), Susceptance (B), DC Resistance (Rdc)						
SERIES OR PARALLEL EQUIVALENT CIRCUIT	C + R, C + D, C + Q, L + R, L + Q, L + D						
SERIES EQUIVALENT CIRCUIT ONLY	X + R, X + D, X + Q						
PARALLEL EQUIVALENT CIRCUIT ONLY	EQUIVALENT CIRCUIT ONLY C+G, B+G, B+D, B+Q, B+R, L+G						
POLAR FORM	Z + Phase Angle, Y + Phase Angle						
AVERAGE	1 ~256 times						
LCD DISPLAY	320 x 240 DOT-MATRIX						
INTERFACE	RS-232C, GPIB						
POWER SOURCE	AC 115V (+10%/-25%), AC 230V (+15% / -14%) (Selectable), 50/60Hz						
DIMENSIONS & WEIGHT	330(W) x 170(H) x 340(D)mm, Approx. 5kg						
Note : Basic accuracy varies with the temperture, frequency, A	C signal level and impedance	e of the device under test.	Specifications subje	ct to change without notice.	CR-8000GGD4DH		
ORDERING INFORMATION		SSORIES					
LCR-8110G 10 MHz Precision LCR Mete	er OPTI	User manual x 1 , Power cord x 1 , Test lead LCR-12 x 1 OPTION					
LCR-8105G 5 MHz Precision LCR Meter	·	B1 DC Bias Voltage Box					

PC Software LCR. Exe. ACCESSORIES SELECTION GUIDE

OPTIONAL ACCESSORIES SELECTION GUIDE						
ACCESSORY MODEL	BRIEF DESCRIPTION	LCR-8110G	LCR-8105G	LCR-8101G		
LCR-05	Test Fixture for axial & radial lead components	Δ	Δ	\checkmark		
LCR-06B	Test Lead with Kelvin clip (4 wire type)	Δ	Δ	√		
LCR-07	Test Lead with Alligator clip (2 wire type)	Δ	Δ	1		
LCR-08	Test Fixture (Tweezers) for SMD / Chip components	Δ	Δ	1		
LCR-12	Test Lead with Kelvin clip (4 wire type)	1	\checkmark	1		
LCR-15	Test Fixture for SMD / Chip components	1	1	1		
GTL-234	RS-232C cable	1	\checkmark	1		
GTL-248	GPIB Cable	1	\checkmark	1		
GRA-404	Rack Adapter Panel (19", 4U)	1	\checkmark	1		

Note : " Δ " means the accessories work with a frequency limitation (under 1MHz)



Global Headquarters

GOOD WILL INSTRUMENT CO., LTD.

No.7-1, Jhongsing Road, Tucheng Dist., New Taipei City 236, Taiwan T +886-2-2268-0389 F +886-2-2268-0639 E-mail: marketing@goodwill.com.tw

China Subsidiary

GOOD WILL INSTRUMENT (SUZHOU) CO., LTD. No. 521, Zhujiang Road, Snd, Suzhou Jiangsu 215011 China T +86-512-6661-7177 F +86-512-6661-7277

Malaysia Subsidiary

GOOD WILL INSTRUMENT (SEA) SDN. BHD. No. 1-3-18, Elit Avenue, Jalan Mayang Pasir 3, 11950 Bayan Baru, Penang, Malaysia T +604-6111122 F +604-6115225

Europe Subsidiary

GOOD WILL INSTRUMENT EURO B.V.

De Run 5427A, 5504DG Veldhoven, THE NETHERLANDS $\textbf{T} + 31(0)40\text{-}2557790 \hspace{0.1in} \textbf{F} + 31(0)40\text{-}2541194$

U.S.A. Subsidiary

INSTEK AMERICA CORP. 5198 Brooks Street Montclair, CA 91763, U.S.A. T +1-909-399-3535 F +1-909-399-0819

Japan Subsidiary

TEXIO TECHNOLOGY CORPORATION.

7F Towa Fudosan Shin Yokohama Bldg., 2-18-13 Shin Yokohama, Kohoku-ku, Yokohama, Kanagawa, 222-0033 Japan T +81-45-620-2305 F +81-45-534-7181

Korea Subsidiary

GOOD WILL INSTRUMENT KOREA CO., LTD.

Room No.503, Gyeonginro 775 (Mullae-Dong 3Ga, Ace Hightech-City B/D 1Dong), Yeongduengpo-Gu, Seoul 150093, Korea. T +82-2-3439-2205 F +82-2-3439-2207

India Subsidiarv

GW INSTEK INDIA LLP.

No.2707/B&C, 1st Floor UNNATHI Building, E-Block, Sahakara Nagar, Bengaluru-560 092. India T +91-80-6811-0600 F +91-80-6811-0626

<u>G</u>UINSTEK





