

## Getting to know the quick select buttons

Using the quick select buttons, the set functions are called up directly. Factory settings:

- Left quick select button ①: Adjust scaling.
- Right quick select button ②: Select image type.



**Warning!** Eye damage due to laser beam.

- Do not look into the laser beam and do not point the laser beam at people.

## Changing assignment of the quick select buttons

Change the left or right quick select button:

- 1 Call up: Move **Joystick** ① to the left or right.
- 2 Navigate: Move **Joystick** up / down.
- 3 Confirm: Press **[OK]** ①.  
or  
Cancel: Press **[Esc]** ②.



## Contact

### GREAT BRITAIN / IRELAND

Testo Ltd.  
Alton, Hampshire GU34 2QJ  
Tel. (14 20) 54 44 33, Fax (14 20) 54 44  
E-mail: info@testo.co.uk

### USA

Testo Inc.  
Flanders, NJ. 07836  
Tel. (973) 2 52 17 20, Fax (973) 2 52 17 29  
E-mail: info@testo.com

### AUSTRALIA

Testo Pty Ltd  
Croydon South VIC 3136  
Tel. (3) 98 00 46 77, Fax (3) 98 00 46 99  
E-mail: info@testo.com.au

The most up-to-date address details of our subsidiaries and agencies can be found in Internet at: [www.testo.com](http://www.testo.com)

# testo 875i · Thermal imager

## Brief instructions

en



## Commissioning

- 1 Push the **rechargeable battery** ① all the way into the rechargeable battery slot until this is flush with the bottom of the handle.
  - The thermal imager starts automatically.
- 2 Remove protective film from the **display** ②.
  - Instrument language (**Language**) and temperature unit (**Unit**) can be set:
- 3 Move **Joystick** ③ up / down to select the desired option.
- 4 Press **[OK]** ③.
- 5 Move **Joystick** up / down to change the setting.
- 6 Press **[OK]**.
- 7 Press **[Apply]** ④ function key to apply the settings.
- 8 Press **[⏻]** ⑤ to switch off the thermal imager.
- 9 Connect **mains unit** (☎) ⑥.
- 10 Completely charge battery (3 hours).
- 11 Insert **Memory card** in the card slot (**SD**) ⑦.
  - The camera is now ready for operation.

## Switching on / off

- Remove **protection cap** ① from the lens.
- Switch on: Press **[⏻]** ②.
  - The measurement view is opened after the warm-up period.
  - The thermal imager performs an automatic zeroing approx. every 60 sec. This can be recognized by a "click". The image is frozen briefly when this occurs.
- Switch off: Press **[⏻]** again.

## Focusing the picture

- Turn **lens** ① by hand.

## Recording the image

- 1 Freeze image: Press trigger button ② once.
- 2 Save image: Press trigger button again.
  - or
  - Discard image: Press **[Esc]** ①.

## Getting to know the menu

- 1 Open menu: Press **[OK]** ①.
- 2 Navigate: Move **Joystick** ① up / down.
- 3 Open submenu (only for menu points marked with ▶): Move joystick to the right.
- 4 Call up a function: Press **[OK]**.
  - or
  - Exit menu: Press **[Esc]** ②.

