N9021B MXA Signal Analyzer 10 Hz to 50 GHz





Table of Contents

Definition and Terms	3
Frequency and Time Specifications	. 4
Amplitude Accuracy and Range Specifications	. 6
Dynamic Range Specifications	. 9
Powersuite Specifications	13
General Specifications	14
Inputs and Outputs	16
Q analyzer	19
Q analyzer – Option B2X	22
Q analyzer – Option B5X	23
Real-time spectrum analyzer	24
Option RT1 and RT2	24

Definition and Terms

Specifications describe the performance of parameters covered by the product warranty and apply to the full temperature range of 0 to 55 °C, unless otherwise noted.

95th percentile values indicate the breadth of the population (approx. $2\ \sigma$) of performance tolerances expected to be met in 95 percent of the cases with a 95 percent confidence, for any ambient temperature in the range of 20 to 30 °C. In addition to the statistical observations of a sample of instruments, these values include the effects of the uncertainties of external calibration references. These values are not warranted. These values are updated occasionally if a significant change in the statistically observed behavior of production instruments is observed.

Typical values describe additional product performance information that is not covered by the product warranty. It is performance beyond specifications that 80 percent of the units exhibit with a 95 percent confidence level over the temperature range 20 to 30 °C. Typical performance does not include measurement uncertainty.

Nominal values indicate expected performance, or describe product performance that is useful in the application of the product, but are not covered by the product warranty.

The analyzer will meet its specifications when:

- It is within its calibration cycle
- Under auto couple control, except when Auto Sweep Time Rules = Accy
- Signal frequencies < 10 MHz, with DC coupling applied
- The analyzer has been stored at an ambient temperature within the allowed operating range for at least two hours before being turned on
- The analyzer has been turned on at least 30 minutes with Auto Align set to Normal, or, if Auto Align is set to off or partial, alignments must have been run recently enough to prevent an Alert message; if the Alert condition is changed from "Time and Temperature" to one of the disabled duration choices, the analyzer may fail to meet specifications without informing the user

This data sheet is a summary of the specifications and conditions for the N9021B MXA signal analyzer. For the complete specifications guide, visit:

www.keysight.com/find/N9021B



Quickly adapt to evolving test requirements

Industries from wireless to satellite communications require wider analysis bandwidth to meet demands for higher data throughput. As higher bandwidth technologies such as 5G NR move into mainstream use, engineers need tools for design validation and manufacturing that offer the accuracy, speed. and bandwidth to accelerate device development. Keysight's new N9021B MXA Signal Analyzer offers best-in-class bandwidth and phase noise for accurate and repeatable signal analysis across millimeter-wave

Frequency and Time Specifications

Option 532 10 Hz to 32 GHz Option 544 10 Hz to 44 GHz Option 550 10 Hz to 50 GHz Band LO Multiple (N) Swept or FFT, with FFT width ≤ 40 MHz FFT, with FFT width > 40 MHz 0 1 10 Hz to 3.6 GHz 10 Hz = 3.4 GHz 1 1 3.5 to 8.4 GHz 3.4 = 8.2 GHz 2 2 8.3 to 13.6 GHz 8.2 = 13.2 GHz 3 2 13.5 to 17.1 GHz 12.2 - 17.1 GHz 4 4 17.0 to 26.5 GHz 17.1 - 26.5 GHz 5 4 26.4 to 34.5 GHz 26.5 - 34.5 GHz 5 4 26.4 to 34.5 GHz 26.5 - 34.5 GHz Frequency reference Accuracy ± [(time since last adjustment x aging rate) + temperature stability + calibration accuracy] Prequency reference Accuracy ± [(time since last adjustment x aging rate) + temperature stability + calibration accuracy] Option PFR Standard ± 1 × 10 -7 / year ± 1 × 10 -6 / year ± 2 × 10 -6 Standard ± 10 × 3 × 10 × 8	Frequency range		DC coupled			
Option 544 10 Hz to 44 GHz Option 550 10 Hz to 50 GHz Band LO Multiple (N) Swept or FFT, with FFT width ≤ 40 MHz FFT, with FFT width > 40 MHz 0 1 10 Hz to 3.6 GHz 10 Hz - 3.4 GHz 1 1 3.5 to 8.4 GHz 3.4 – 8.2 GHz 2 2 8.3 to 13.6 GHz 8.2 – 13.2 GHz 3 2 13.5 to 17.1 GHz 13.2 – 17.1 GHz 4 4 17.0 to 26.5 GHz 17.1 – 26.5 GHz 5 4 26.4 to 34.5 GHz 26.5 – 34.5 GHz 6 8 34.4 to 50 GHz 34.5 – 50 GHz Frequency reference Accuracy ± [(time since last adjustment x aging rate) + temperature stability + calibration accuracy Option PFR Standard ± 1.5 x 10 -7 / year ± 1 x 10 -6 / year ± 1.5 x 10 -7 / year ± 1 x 10 -6 / year ± 1.5 x 10 -7 / year ± 2 x 10 -6 ± 1.5 x 10 -7 / year ± 2 x 10 -6 ± 1.5 x 10 -7 / year ± 2 x 10 -6 <td colspa<="" td=""><td colspan="2">Option 532</td><td colspan="3">10 Hz to 32 GHz</td></td>	<td colspan="2">Option 532</td> <td colspan="3">10 Hz to 32 GHz</td>	Option 532		10 Hz to 32 GHz		
Band	·		10 Hz to 44 GHz			
Band	Option 550					
1 1 1 3.5 to 8.4 GHz 3.4 − 8.2 GHz 2 2 8.3 to 13.6 GHz 8.2 − 13.2 GHz 3 2 13.5 to 17.1 GHz 13.2 − 17.1 GHz 4 4 4 17.0 to 26.5 GHz 17.1 - 26.5 GHz 5 4 26.4 to 34.5 GHz 26.5 − 34.5 GHz 6 8 34.4 to 50 GHz 34.5 − 50 GHz Frequency reference Accuracy		LO Multiple (N)	Swept or FFT, with FFT width	า ≤ 40 MHz	FFT, with FFT width > 40 MHz	
2 2 8.3 to 13.6 GHz 8.2 - 13.2 GHz 3 2 13.5 to 17.1 GHz 13.2 - 17.1 GHz 4 4 17.0 to 26.5 GHz 17.1 - 26.5 GHz 5 4 26.4 to 34.5 GHz 26.5 - 34.5 GHz 6 8 34.4 to 50 GHz 34.5 - 50 GHz Frequency reference Accuracy ± [(time since last adjustment x aging rate) + temperature stability + calibration accuracy] Aging rate Option PFR Standard ± 1.5 x 10 -7 / years Temperature stability Option PFR Standard 20 to 30 °C ± 1.5 x 10 -8 ± 2 x 10 -8 Standard Achievable initial calibration accuracy Option PFR Standard Achievable initial calibration accuracy 0ption PFR Standard Besidual FM (with option PFR) ≤ (0.25 Hz × N) _{pe} in 20 ms (nominal) Frequency reference accuracy (start, stop, center, marker) ± (marker frequency x frequency reference accuracy + 0.100 Hz) ± (marker frequency x frequency reference accuracy + 0.141 Hz)	0	1	10 Hz to 3.6 GHz		10 Hz – 3.4 GHz	
3	1	1	3.5 to 8.4 GHz		3.4 – 8.2 GHz	
4 4 4 17.0 to 26.5 GHz 17.1 − 26.5 GHz 2 6.4 to 34.5 GHz 26.5 − 34.5 GHz 34.5 − 50 GHz Frequency reference Accuracy	2	2	8.3 to 13.6 GHz		8.2 – 13.2 GHz	
5	3	2	13.5 to 17.1 GHz		13.2 – 17.1 GHz	
Frequency reference Accuracy	4	4	17.0 to 26.5 GHz		17.1 – 26.5 GHz	
Accuracy ± [(time since last adjustment x aging rate) + temperature stability + calibration accuracy] Aging rate Option PFR Standard ± 1 x 10 -7 / year ± 1 x 10 -6 / year ± 1.5 x 10 -7 / 2 years Temperature stability Option PFR Standard 20 to 30 °C ± 1.5 x 10 -8 ± 2 x 10 -6 Full temperature range ± 5 x 10 -8 ± 2 x 10 -6 Achievable initial calibration accuracy Option PFR Standard ± 4 x 10 -8 ± 1.4 x 10 -6 Residual FM (with option PFR) ≤ (0.25 Hz × N) _{P-P} in 20 ms (nominal) Residual FM (Standard) ≤ (10 Hz x N) _{P-P} in 20 ms (nominal) Frequency readout accuracy (start, stop, center, marker) ± (marker frequency x frequency reference accuracy + 0.25 % x span + 5 % x RBW + 2 Hz + 0.5 x horizontal resolution ¹) Marker frequency counter Accuracy ± (marker frequency x frequency reference accuracy + 0.100 Hz) Delta counter accuracy ± (delta frequency x frequency reference accuracy + 0.141 Hz) Counter resolution 0.001 Hz Frequency span (FFT and swept mode) Range 0 Hz (zero span), 10 Hz to maximum frequency of instrument Resolution 2 Hz Accuracy Stepped/Swept ± (0.25 % x span + horizontal resolution)	5	4	26.4 to 34.5 GHz		26.5 – 34.5 GHz	
Accuracy ± [(time since last adjustment x aging rate) + temperature stability + calibration accuracy] Aging rate Option PFR Standard ± 1 × 10 -7 / year ± 1.5 × 10 -7 / 2 years Temperature stability Option PFR Standard 20 to 30 °C ± 1.5 × 10 -8 × 2 × 10 -6 Full temperature range ± 5 × 10 -8 £ 2 × 10 -6 Achievable initial calibration accuracy Option PFR Standard ± 4 × 10 -8 £ 1.4 × 10 -6 Residual FM (with option PFR) ≤ (0.25 Hz × N) _{PP} in 20 ms (nominal) Frequency readout accuracy (start, stop, center, marker) ± (marker frequency x frequency reference accuracy + 0.25 % x span + 5 % x RBW + 2 Hz + 0.5 x horizontal resolution ¹) Marker frequency counter Accuracy ± (marker frequency x frequency x frequency reference accuracy + 0.100 Hz) Delta counter accuracy ± (delta frequency x frequency reference accuracy + 0.141 Hz) Counter resolution 0.001 Hz Frequency span (FFT and swept mode) Range 0 Hz (zero span), 10 Hz to maximum frequency of instrument Resolution 2 Hz Accuracy Stepped/Swept ± (0.25 % x span + horizontal resolution)	6	8	34.4 to 50 GHz		34.5 – 50 GHz	
Aging rate Option PFR Standard ±1 x 10 -7 / year ±1 x 10 -6 / year ±1.5 x 10 -7 / 2 years ±1.5 x 10 -6 / year Temperature stability Option PFR Standard 20 to 30 °C ±1.5 x 10 -8 / ±2 x 10 -6 ±2 x 10 -6 Full temperature range ±5 x 10 -8 / ±2 x 10 -6 ±4 x 10 -6 Achievable initial calibration accuracy Option PFR Standard Residual FM (with option PFR) ≤ (0.25 Hz x N) _{P-P} in 20 ms (nominal) Residual FM (Standard) ≤ (10 Hz x N) _{P-P} in 20 ms (nominal) Frequency readout accuracy (start, stop, center, marker) ± (marker frequency x frequency accuracy + 0.25 % x span + 5 % x RBW + 2 Hz + 0.5 x horizontal resolution ¹) Marker frequency counter Accuracy ± (marker frequency x frequency reference accuracy + 0.100 Hz) Delta counter accuracy ± (delta frequency x frequency reference accuracy + 0.141 Hz) Counter resolution 0.001 Hz Frequency span (FFT and swept mode) Range 0 Hz (zero span), 10 Hz to maximum frequency of instrument Resolution 2 Hz Accuracy	Frequency referen	се				
# 1 x 10 -7 / year # 1 x 10 -8 / year # 1.5 x 10 -7 / 2 years ± 1.5 x 10 -7 / 2 years	Accuracy		± [(time since last adjustment x	aging rate) + to	emperature stability + calibration accuracy]	
# 1.5 x 10 -7 / 2 years Temperature stability	Aging rate		Option PFR	Standard		
# 1.5 x 10 -7 / 2 years Temperature stability Option PFR Standard 20 to 30 °C # 1.5 x 10 -8 # 2 x 10 -6 Full temperature range # 5 x 10 -8 # 2 x 10 -6 Achievable initial calibration accuracy Option PFR Standard # 4 x 10 -8 # 1.4 x 10 -6 Residual FM (with option PFR) \$ (0.25 Hz x N)_{p-p} in 20 ms (nominal) Residual FM (Standard) \$ (10 Hz x N)_{p-p} in 20 ms (nominal) Frequency readout accuracy (start, stop, center, marker) # (marker frequency x frequency reference accuracy + 0.25 % x span + 5 % x RBW + 2 Hz + 0.5 x horizontal resolution 1) Marker frequency counter Accuracy # (marker frequency x frequency reference accuracy + 0.100 Hz) Delta counter accuracy # (delta frequency x frequency reference accuracy + 0.141 Hz) Counter resolution 0.001 Hz Frequency span (FFT and swept mode) Range 0 Hz (zero span), 10 Hz to maximum frequency of instrument Resolution 2 Hz Accuracy Stepped/Swept # (0.25 % x span + horizontal resolution)			± 1 × 10 ⁻⁷ / year	± 1 × 10 ⁻⁶	/ year	
Temperature stability 20 to 30 °C ± 1.5 × 10 -8 ± 2 × 10 -6 Full temperature range ± 5 × 10 -8 Achievable initial calibration accuracy Option PFR Standard ± 4 × 10 -8 ± 1.4 × 10 -6 Residual FM (with option PFR) S(0.25 Hz × N) _{P-P} in 20 ms (nominal) Residual FM (Standard) Frequency readout accuracy (start, stop, center, marker) ± (marker frequency x frequency reference accuracy + 0.25 % x span + 5 % x RBW + 2 Hz + 0.5 x horizontal resolution ¹) Marker frequency counter Accuracy ± (marker frequency x frequency x frequency x frequency reference accuracy + 0.100 Hz) Delta counter accuracy ± (delta frequency x frequency reference accuracy + 0.141 Hz) Counter resolution 0.001 Hz Frequency span (FFT and swept mode) Range 0 Hz (zero span), 10 Hz to maximum frequency of instrument Resolution 2 Hz Accuracy Stepped/Swept ± (0.25 % x span + horizontal resolution)			± 1.5 × 10 ⁻⁷ / 2 years			
Full temperature range Achievable initial calibration accuracy Option PFR Standard ± 4 × 10 -8 Escidual FM (with option PFR) Scindard (0.25 Hz × N) _{p-p} in 20 ms (nominal) Scidual FM (Standard) Frequency readout accuracy (start, stop, center, marker) ± (marker frequency x frequency reference accuracy + 0.25 % x span + 5 % x RBW + 2 Hz + 0.5 x horizontal resolution 1) Marker frequency counter Accuracy ± (marker frequency x frequency reference accuracy + 0.25 % x span + 5 % x RBW + 2 Hz + 0.5 x horizontal resolution 1) Delta counter accuracy ± (delta frequency x frequency reference accuracy + 0.100 Hz) Delta counter accuracy ± (delta frequency x frequency reference accuracy + 0.141 Hz) Counter resolution 0.001 Hz Frequency span (FFT and swept mode) Range 0 Hz (zero span), 10 Hz to maximum frequency of instrument Resolution 2 Hz Accuracy Stepped/Swept ± (0.25 % x span + horizontal resolution)	Temperature stabi	lity		Standard		
Achievable initial calibration accuracy Detion PFR ± 4 × 10 -8 ± 1.4 × 10 -6 Residual FM (with option PFR) Standard ± 4 × 10 -8 ± 1.4 × 10 -6 Residual FM (Standard) Standard) Standard ± 1.4 × 10 -6 Exempla Standard (constitution of the properties of th	20 to 30 °C		$\pm 1.5 \times 10^{-8}$ $\pm 2 \times 10^{-6}$			
$ \pm 4 \times 10^{-8} $ Residual FM (with option PFR) ≤ (0.25 Hz × N) _{P-P} in 20 ms (nominal) Residual FM (Standard) ≤ (10 Hz × N) _{P-P} in 20 ms (nominal) Frequency readout accuracy (start, stop, center, marker) ± (marker frequency x frequency reference accuracy + 0.25 % x span + 5 % x RBW + 2 Hz + 0.5 x horizontal resolution ¹) Marker frequency counter Accuracy ± (marker frequency x frequency reference accuracy + 0.100 Hz) Delta counter accuracy ± (delta frequency x frequency reference accuracy + 0.141 Hz) Counter resolution 0.001 Hz	Full temperature r	ange	± 5 × 10 ⁻⁸ ± 2 × 10 ⁻⁶			
Residual FM (with option PFR) ≤ (0.25 Hz × N) _{P-P} in 20 ms (nominal) Residual FM (Standard) ≤ (10 Hz × N) _{P-P} in 20 ms (nominal) Frequency readout accuracy (start, stop, center, marker) ± (marker frequency x frequency reference accuracy + 0.25 % x span + 5 % x RBW + 2 Hz + 0.5 x horizontal resolution ¹) Marker frequency counter Accuracy ± (marker frequency x frequency reference accuracy + 0.100 Hz) Delta counter accuracy ± (delta frequency x frequency reference accuracy + 0.141 Hz) Counter resolution 0.001 Hz Frequency span (FFT and swept mode) Range 0 Hz (zero span), 10 Hz to maximum frequency of instrument Resolution 2 Hz Accuracy Stepped/Swept ± (0.25 % x span + horizontal resolution)	Achievable initial	calibration accuracy	Option PFR	Standard		
Residual FM (Standard) ≤ (10 Hz x N) _{P-P} in 20 ms (nominal) Frequency readout accuracy (start, stop, center, marker) ± (marker frequency x frequency reference accuracy + 0.25 % x span + 5 % x RBW + 2 Hz + 0.5 x horizontal resolution ¹) Marker frequency counter Accuracy ± (marker frequency x frequency reference accuracy + 0.100 Hz) Delta counter accuracy ± (delta frequency x frequency reference accuracy + 0.141 Hz) Counter resolution 0.001 Hz Frequency span (FFT and swept mode) Range 0 Hz (zero span), 10 Hz to maximum frequency of instrument Resolution 2 Hz Accuracy Stepped/Swept ± (0.25 % x span + horizontal resolution)			± 4 × 10 ⁻⁸	± 1.4 × 10 ⁻⁶	6	
Frequency readout accuracy (start, stop, center, marker) ± (marker frequency x frequency reference accuracy + 0.25 % x span + 5 % x RBW + 2 Hz + 0.5 x horizontal resolution ¹) Marker frequency counter Accuracy ± (marker frequency x frequency reference accuracy + 0.100 Hz) Delta counter accuracy ± (delta frequency x frequency reference accuracy + 0.141 Hz) Counter resolution 0.001 Hz Frequency span (FFT and swept mode) Range 0 Hz (zero span), 10 Hz to maximum frequency of instrument Resolution 2 Hz Accuracy Stepped/Swept ± (0.25 % x span + horizontal resolution)	Residual FM (with	option PFR)	≤ (0.25 Hz × N) _{p-p} in 20 ms (n	ominal)		
± (marker frequency x frequency reference accuracy + 0.25 % x span + 5 % x RBW + 2 Hz + 0.5 x horizontal resolution ¹) Marker frequency counter Accuracy ± (marker frequency x frequency reference accuracy + 0.100 Hz) Delta counter accuracy ± (delta frequency x frequency reference accuracy + 0.141 Hz) Counter resolution 0.001 Hz Frequency span (FFT and swept mode) Range 0 Hz (zero span), 10 Hz to maximum frequency of instrument Resolution 2 Hz Accuracy Stepped/Swept ± (0.25 % x span + horizontal resolution)	Residual FM (Star	ndard)	≤ (10 Hz × N) _{p-p} in 20 ms (nor	minal)		
Accuracy ± (marker frequency x frequency reference accuracy + 0.100 Hz) Delta counter accuracy ± (delta frequency x frequency reference accuracy + 0.141 Hz) Counter resolution 0.001 Hz Frequency span (FFT and swept mode) Range 0 Hz (zero span), 10 Hz to maximum frequency of instrument Resolution 2 Hz Accuracy Stepped/Swept ± (0.25 % x span + horizontal resolution)	Frequency readou	t accuracy (start, stop	, center, marker)			
Accuracy ± (marker frequency x frequency reference accuracy + 0.100 Hz) Delta counter accuracy ± (delta frequency x frequency reference accuracy + 0.141 Hz) Counter resolution 0.001 Hz Frequency span (FFT and swept mode) Range 0 Hz (zero span), 10 Hz to maximum frequency of instrument Resolution 2 Hz Accuracy Stepped/Swept ± (0.25 % x span + horizontal resolution)	± (marker frequen	cy x frequency refere	nce accuracy + 0.25 % x span	+ 5 % x RBW	/ + 2 Hz + 0.5 x horizontal resolution 1)	
Delta counter accuracy ± (delta frequency x frequency reference accuracy + 0.141 Hz) Counter resolution 0.001 Hz Frequency span (FFT and swept mode) Range 0 Hz (zero span), 10 Hz to maximum frequency of instrument Resolution 2 Hz Accuracy Stepped/Swept ± (0.25 % x span + horizontal resolution)	Marker frequency	counter				
Counter resolution 0.001 Hz Frequency span (FFT and swept mode) Range 0 Hz (zero span), 10 Hz to maximum frequency of instrument Resolution 2 Hz Accuracy Stepped/Swept ± (0.25 % x span + horizontal resolution)	Accuracy		± (marker frequency x freque	ncy reference	e accuracy + 0.100 Hz)	
Frequency span (FFT and swept mode) Range 0 Hz (zero span), 10 Hz to maximum frequency of instrument Resolution 2 Hz Accuracy Stepped/Swept ± (0.25 % x span + horizontal resolution)	Delta counter acci	uracy	± (delta frequency x frequence	y reference a	accuracy + 0.141 Hz)	
Range 0 Hz (zero span), 10 Hz to maximum frequency of instrument Resolution 2 Hz Accuracy Stepped/Swept ± (0.25 % x span + horizontal resolution)	Counter resolution	n	0.001 Hz			
Resolution 2 Hz Accuracy Stepped/Swept ± (0.25 % x span + horizontal resolution)	Frequency span (F	FFT and swept mode)				
Accuracy Stepped/Swept ± (0.25 % x span + horizontal resolution)	Range		0 Hz (zero span), 10 Hz to maximum frequency of instrument		uency of instrument	
Stepped/Swept ± (0.25 % x span + horizontal resolution)	Resolution					
	Accuracy					
FFT ± (0.1% x span + horizontal resolution)	-		± (0.25 % x span + horizontal	resolution)		
	FFT					

^{1.} Horizontal resolution is span/(sweep points - 1).

Sweep time and triggering				
Range	Span = 0 Hz		1 µs to 6000 s	
	Span ≥ 10 Hz		1 ms to 4000 s	
Accuracy	Span = 0 Hz, swept		± 0.01 % (nominal)	
	Span ≥ 10 Hz, FFT		± 40 % (nominal)	
	Span = 0 Hz		± 0.01 % (nominal)	
Trigger	Free run, line, video, e	externa	al 1, external 2, RF burst, p	eriodic timer
Trigger delay	Span = 0 or FFT		-150 to +500 ms	
	Span ≥ 10 Hz, swept		0 μs to 500 ms	
	Resolution		0.1 µs	
Time gating				
Gate methods	Gated LO; gated vide	o; gate	ed FFT	
Gate length range	100.0 ns to 5.0 s (Exc	ept me	ethod = FFT)	
Gate delay range	0 to 100.0 s			
Gate delay jitter	33.3 ns p-p (nominal)			
Sweep (trace) point range				
All spans	1 to 100,001			
Resolution bandwidth (RBW)				
EMI bandwidths (CISPR comp	oliant)		200 Hz, 9 kHz, 120 kHz,	1 MHz
EMI bandwidths (Mil STD 461	compliant)		10 Hz, 100 Hz, 1 kHz, 10	kHz, 100 kHz, 1 MHz
Range (with -3 dB bandwidth,	standard)		1 Hz to 3 MHz (10% steps	s), 4, 5, 6, 8 MHz
With option B2X/B5X and Opti	ion RBE			, 60, 70, 80, 100, 133, 150, 200, and alyzer mode and zero span
Bandwidth accuracy (power)				
1 Hz to 750 kHz			± 1.0 % (± 0.044 dB)	
820 kHz to 1.2 MHz (< 3.6	GHz CF)		± 2.0 % (± 0.088 dB)	
1.3 to 2 MHz (< 3.6 GHz C	F)		± 0.07 dB (nominal)	
2.2 to 3 MHz (< 3.6 GHz C	F)		± 0.15 dB (nominal)	
4 to 8 MHz (< 3.6 GHz CF)			± 0.25 dB (nominal)	
Bandwidth accuracy (-3 dB)	1 Hz to 1.3 MHz		± 2% (nominal)	
Selectivity (-60 dB/-3 dB)		4.1: 1 (nominal)		
Video Bandwidth (VBW)				
Range 1 Hz to 3 MHz (10% steps),		4, 5,6, 8 MHz, and wide op	pen (labeled 50 MHz)	
Accuracy	±6%, nominal	±6%, nominal		
Analysis bandwidth ¹				
Maximum bandwidth	Option B2X	255 MHz		
	Option B5X		MHz	

^{1.} Analysis bandwidth is the instantaneous bandwidth available around a center frequency over which the input signal can be digitized for further analysis or processing in the time, frequency, or modulation domain

Amplitude Accuracy and Range Specifications

Amplitude range				
Measurement range	Preamp Off	Displayed average noise level (DANL) to +30 dBm		
	Preamp On	Displayed average noise level (DANL) to +20 dBm		
Input attenuator range	0 to 70 dB in 2 dB steps			
Maximum safe input level				
Average total power	+30 dBm (1 W)	10 up pulse width 110/ duty evels and input attenuation		
Peak pulse power	+50 dBm (100 W)	< 10 µs pulse width, < 1% duty cycle, and input attenuation ≥ 30 dB		
DC volts	±0.2 Vdc			
Display range				
Log scale	0.1 to 1 dB/division in 0.1 dE	3 steps		
	1 to 20 dB/division in 1 dB steps (10 display divisions)			
Linear scale	10 divisions			
Scale units	dBm, dBmV, dBμV, dBmA, d	dBμA, V, W, A, dBuV/m, dBuA/m, dBpT, dBG, dBpW		
Electronic attenuator (option EA	Electronic attenuator (option EA3)			
Frequency range	10 Hz to 3.6 GHz ¹			
Attenuation range				
Electronic attenuator range	0 to 24 dB, 1 dB steps			
Full attenuation range	0 to 94 dB, 1 dB steps (Mechanical + Electronic)			

^{1.} Frequency range of option EA3 varies according to sweep types. Please refer to the frequency band definition on page 4

Preamplifier		
Frequency range	Option P32	100 kHz to 32 GHz
	Option P44	100 kHz to 44 GHz
	Option P50	100 kHz to 50 GHz
Gain	100 kHz to 3.6 GHz	+20 dB nominal
	3.6 to 26.5 GHz	+35 dB, nominal
	26.5 to 50 GHz	+40 dB, nominal
Noise figure	100 kHz to 3.6 GHz	11 dB, nominal
	3.6 to 8.4 GHz	9 dB, nominal
	8.4 to 13.6 GHz	10 dB, nominal
	13.6 to 50 GHz	DANL + 176.24 dB, nominal

Frequency response		Specification	95th percentile
20 to 30°, preselector cer	tering applied above 3.6 GHz		
Preamp Off	20 Hz to 10 MHz	±0.43 dB	±0.23 dB
10 dB attenuation	10 to 50 MHz	±0.43 dB	±0.21 dB
	50 to 3.6 GHz	±0.36 dB	±0.22 dB
	3.5 to 5.2 GHz	±1.5 dB	±0.76 dB
	5.2 to 8.4 GHz	±1.3 dB	±0.56 dB
	8.3 to 13.6 GHz	±1.8 dB	±0.67 dB
	13.5 to 17.1 GHz	±1.8 dB	±0.62 dB
	17.0 to 22.0 GHz	±1.8 dB	±0.73 dB
	22.0 to 26.5 GHz	±2.3 dB	±0.76 dB
	26.4 to 34.5 GHz	±2.3 dB	±0.82 dB
	34.4 to 50 GHz	±3.0 dB	±1.21 dB
Preamp On	100 kHz to 50 MHz	±0.7 dB	±0.31 dB
0 dB attenuation	50 MHz to 3.6 GHz	±0.55 dB	±0.25 dB
	3.5 to 5.2 GHz	±1.8 dB	±0.78 dB
	5.2 to 8.4 GHz	±1.8 dB	±0.63 dB
	8.3 to 13.6 GHz	±2.1 dB	±0.51 dB
	13.5 to 17.1 GHz	±2.3 dB	±0.8 dB
	17.0 to 22.0 GHz	±2.6 dB	±0.94 dB
	22 to 26.5 GHz	±3.3 dB	±0.96 dB
	26.4 to 34.5 GHz	±2.8 dB	±1.04 dB
	34.4 to 50 GHz	±3.9 dB	±1.37 dB
Input attenuation switchin	g uncertainty		
Attenuation > 2 dB,	50 MHz (ref frequency)	± 0.20 dB	± 0.08 dB, typical
Preamp off, Relative to	20 Hz to 3.6 GHz		± 0.3 dB, nominal
10 dB	3.5 to 8.4 GHz		± 0.5 dB, nominal
	8.3 to 13.6 GHz		± 0.7 dB, nominal
	13.5 to 26.5 GHz		± 0.7 dB, nominal
	26.4 to 50 GHz		± 1.0 dB, nominal
Total absolute amplitude a	occuracy		
		t signal –10 to –50 dBm, RF preselectory reference level, any scale, σ = nomin	
Preamp Off	At 50 MHz	± 0.45 dB, ± 0.19 dB (95% pc	ercentile)
	At all frequencies	Specification: ± (0.45 dB + freq re 95% percentile: ± (0.19 dB + freq r percentile)	sponse)
Preamp On	At all frequencies	± (0.49 dB + frequency response)	

Input voltage standing wave ratio (VSWR)		95% percentile
Preamp Off,	10 MHz to 3.6 GHz	1.125
Input atten 10 dB	3.5 to 8.4 GHz	1.162
	8.3 to 13.6 GHz	1.217
	13.5 to 17.1 GHz	1.262
	17.0 to 26.5 GHz	1.319
	26.4 to 34.5 GHz	1.546
	34.4 to 50 GHz	1.676
Preamp On,	10 MHz to 3.6 GHz	1.386
Input atten 0 dB	3.5 to 8.4 GHz	1.539
	8.3 to 13.6 GHz	1.385
	13.5 to 17.1 GHz	1.345
	17.0 to 26.5 GHz	1.372
	26.4 to 34.5 GHz	1.571
	34.4 to 50 GHz	1.725
RBW switching uncertainty (reference to	o 30 kHz RBW)	
1 Hz to 1.5 MHz RBW	± 0.05 dB	
1.6 to 3 MHz RBW	± 0.10 dB	
4, 5, 6, 8 MHz RBW	± 1.0 dB	
Reference level		
Range	Log scale	-170 to +30 dBm in 0.01 dB steps
	Linear scale	Same as log (707 pV to 7.07 V)
Accuracy	0 dB	
Display scale switching uncertainty		
Switching between linear and log	0 dB	
Log scale/div switching	0 dB	
Display scale fidelity		
-10 dBm < mixer level < -80 dBm	± 0.10 dB total	
Detector type		
Normal, peak, sample, negative peak, I	og power average. RMS averag	ge, and voltage average

Dynamic Range Specifications

1 dB gain compression (two-tone) (At 1 kHz RBW with 100 kHz tone spacing, 20 to 30°C) **Preamp Off** 20 MHz to 3.6 GHz +5 dBm, nominal 3.6 to 16 GHz +8 dBm, nominal 16 to 26.5 GHz +7 dBm, nominal 26.5 to 50 GHz 0 dBm, nominal Preamp On 10 MHz to 3.6 GHz -14 dBm, nominal 3.6 to 26.5 GHz Tone spacing 100 kHz to 20 MHz -28 dBm, nominal Tone spacing > 70 MHz -20 dBm, nominal 26.5 to 50 GHz -30 dBm, nominal

Displayed average noise level (DANL)

Input terminated, 1 Hz RBW, sample or average detector, averaging type = Log, 0 dB input attenuation, IF Gain = High, 20 to 30 °C, with wideband options (MPB, B2X, or B5X)

Preamp Off

10 Hz		-123 dBm, nominal
20 Hz		-129 dBm, nominal
100 Hz		-126 dBm, nominal
1 kHz		-146 dBm, nominal
9 kHz to 5 MHz		-147 dBm, typical
5 to 10 MHz	-155 dBm	-158 dBm, typical
10 MHz to 1.2 GHz	-154 dBm	-157 dBm, typical
1.2 to 2.1 GHz	-152 dBm	-155 dBm, typical
2.1 to 3 GHz	-151 dBm	-154 dBm, typical
3 to 3.6 GHz	-150 dBm	-153 dBm, typical
3.5 to 4.2 GHz	-143 dBm	-147 dBm, typical
4.2 to 6.6 GHz	-144 dBm	-148 dBm, typical
6.6 to 8.4 GHz	-147 dBm	-149 dBm, typical
8.3 to 13.6 GHz	-147 dBm	-149 dBm, typical
13.5 to 14 GHz	-143 dBm	-147 dBm, typical
14 to 17.1 GHz	-145 dBm	-148 dBm, typical
17 to 22.5 GHz	-145 dBm	-146 dBm, typical
22.5 to 26.5 GHz	-139 dBm	-143 dBm, typical
26.4 to 30 GHz	-140 dBm	-143 dBm, typical
30 to 34.5 GHz	-138 dBm	-143 dBm, typical
34.4 to 37 GHz	-134 dBm	-139 dBm, typical
37 to 40 GHz	-132 dBm	-138 dBm, typical
40 to 49 GHz	-130 dBm	-136 dBm, typical
49 to 50 GHz	-128 dBm	-135 dBm, typical

Displayed average noise level (continued)			
Preamp On	100 kHz to 5 MHz		-159 dBm, typical
	5 to 10 MHz	-163 dBm	-167 dBm, typical
	10 MHz to 1.2 GHz	-164 dBm	-166 dBm, typical
	1.2 to 2.1 GHz	-163 dBm	-165 dBm, typical
	2.1 to 3.6 GHz	-162 dBm	-164 dBm, typical
	3.5 to 8.4 GHz	-158 dBm	-161 dBm, typical
	8.3 to 13.6 GHz	-160 dBm	-162 dBm, typical
	13.5 to 17.1 GHz	-161 dBm	-163 dBm, typical
	17 to 20 GHz	-160 dBm	-162 dBm, typical
	20 to 26.5 GHz	-158 dBm	-160 dBm, typical
	26.4 to 30 GHz	-157 dBm	-159 dBm, typical
	30 to 34.5 GHz	-155 dBm	-158 dBm, typical
	34.5 to 37 GHz	-153 dBm	-157 dBm, typical
	37 to 40 GHz	-152 dBm	-155 dBm, typical
	40 to 44 GHz	-149 dBm	-154 dBm, typical
	44 to 46 GHz	-149 dBm	-154 dBm, typical
	46 to 50 GHz	-146 dBm	-151 dBm, typical

DANL with noise floor extension (option NF2)

DANL improvement exceeds 9 dB with 95% confidence in the average of all bands, paths, frequency options and signal path option (MPB).

Band	Frequency	Preamp Off	Preamp On
0, f > 20 MHz	10 Hz to 3.5 GHz	-163 dBm	-174 dBm
1	3.5 to 8.4 GHz	-159 dBm	-172 dBm
2	8.3 to 13.6 GHz	-159 dBm	-172 dBm
3	13.5 to 17.1 GHz	-159 dBm	-173 dBm
4	17.0 to 26.5 GHz	-154 dBm	-169 dBm
5	26.4 to 34.5 GHz	-153 dBm	-167 dBm
6	34.4 to 50 GHz	-144dBm	-158 dBm

0	0111100000112	TTTGBIII	100 02111
Spurious response			
Residual responses	200 kHz to 8.4 GHz	(swept)	-100 dBm
	Zero span or FFT o	r other frequencies	-100 dBm nominal
Images response	10 MHz to 3.6 GHz		-80 dBc, -108 dBc typical
f ±645 MHz,	3.5 to 13.6 GHz		-78 dBc, -87 dBc typical
Mixer level -10 dBm	13.5 to 17.1 GHz		-74 dBc, -85 dBc typical
	17.0 to 22 GHz		-70 dBc, -81 dBc typical
	22 to 26.5 GHz		-68 dBc, -77 dBc typical
	26.5 to 34.5 GHz		-70 dBc, -94 dBc typical
	34.4 to 42 GHz		-60 dBc, -79 dBc typical
	42 to 50 GHz		-75 dBc, nominal
LO related spurious (f >600 l	LO related spurious (f >600 MHz from carrier)		
	10 MHz to 3.6 GHz		-90 dBc typical

Other spurious	Mixer level	Response
Carrier frequency ≤ 3 GHz		-80 dBc nominal
Carrier frequency 3 to 26.5 GHz		
First RF order (f ≥ 10 MHz from carrier)	-10 dBm	-80 dBc + 20log(N¹), including IF feedthrough, LO harmonic mixing responses
Higher RF order (f ≥ 10 MHz from carrier)	-40 dBm	-80 dBc + 20log(N1), including higher order mixer response
Carrier frequency > 26.5 GHz		
First RF order (f ≥ 10 MHz from carrier)	-30 dBm	-90 dBc nominal
Higher RF order (f ≥ 10 MHz from carrier)	-30 dBm	-90 dBc nominal

^{1.} N is the LO multiplication factor. Refer to page 4 for the N value verses frequency ranges.


Second harmonic distortion (SHI)				
Preamp Off	Source frequency		Distortion	SHI
Mixer level at -15	10 MHz to 1.0 GHz		-63 dBc	+48 dBm, +55 dBm typical
dBm	1.0 to 1.8 GHz		-60 dBc	+45 dBm, +57 dBm typical
	1.75 to 3 GHz		-69 dBc	+54 dBm, +60 dBm typical
	3 to 6.5 GHz		-74 dBc	+59 dBm, +67 dBm typical
	6.5 to 10 GHz		-72 dBc	+57 dBm, +70 dBm typical
	10 to 13.25 GHz		-65 dBc	+50 dBm, +61 dBm typical
	13.2 to 25 GHz		-70 dBc nominal	+55 dBm nominal
Preamp On	Source frequency	Preamp level ²	Distortion	SHI
	10 MHz to 1.8 GHz	-45 dBm	-78 dBc	+33 dBm nominal
	1.8 to 13.25 GHz	-50 dBm	-60 dBc	+10 dBm nominal
	13.25 to 25 GHz	-50 dBm	-50 dBc	0 dBm nominal

^{2.} Preamp level = Input level - Input Attenuation

Third-order intermodulation	distortion (TOI)	
Two –18 dBm tones at input	at input mixer with tone separation at 100 kHz	z, 20 to 30 °C
Preamp Off	10 to 150 MHz	+14.5 dBm, +19.5 dBm typical
	150 to 300 MHz	+16 dBm, +20 dBm typical
	300 MHz to 1.1 GHz	+17 dBm, +21 dBm typical
	1.1 to 3.6 GHz	+21 dBm, +22.5 dBm typical
	3.5 to 8.4 GHz	+18 dBm, +20 dBm typical
	8.3 to 13.6 GHz	+18 dBm, +23 dBm typical
	13.5 to 17.1 GHz	+13 dBm, +16.5 dBm typical
	17.0 to 26.5 GHz	+13 dBm, +16 dBm typical
	26.4 to 34.5 GHz	+12 dBm, +19 dBm typical
	34.4 to 50 GHz	+8 dBm, +12 dBm typical
Preamp On Two tone at prea	mp input	
Two -45 dBm	10 MHz to 500 MHz	+4 dBm nominal
	500 MHz to 3.6 GHz	+4.5 dBm nominal
Two -50 dBm	3.6 to 26.5 GHz	-15 dBm nominal ³

^{3.} Exception for frequencies between 13.6 to 17.6 GHz, TOI is -18 dBm nominal

Phase noise	Offset	Specification	Typical
20 to 30 °C, CF = 1 GHz	10 Hz		-80 dBc/Hz nominal
	100 Hz	-94 dBc/Hz	-100 dBc/Hz typical
	1 kHz	-121 dBc/Hz	-124 dBc/Hz typical
	10 kHz	-129 dBc/Hz	-130 dBc/Hz typical
	100 kHz	-129 dBc/Hz	-130 dBc/Hz typical
	1 MHz	-145 dBc/Hz	-146 dBc/Hz typical
	10 MHz	-155 dBc/Hz	-158 dBc/Hz typical



Powersuite Specifications

Channel Power			
Amplitude accuracy, W-CDMA or IS95 (20 to 30 °C, attenuation = 10 dB)	± 0.82 dB	± 0.23 dB (95th percentile)	
Occupied bandwidth			
Frequency accuracy		± [span/1000] nominal	
Adjacent channel power	Adjacent	Alternate	
Accuracy, W-CDMA (ACLR) (at specific mixer levels and	ACLR ranges)		
MS	± 0.14 dB	± 0.18 dB	
BTS	± 0.49 dB	± 0.42 dB	
Dynamic range			
Without noise correction	-73 dB typical	-79 dB typical	
With noise correction	-78 dB typical	-82 dB typical	
Offset channel pairs measured	1 to 6		
ACP measurement and transfer time (fast method)	10 ms nominal (σ = 0.2 dB)		
Multiple number of carriers measured	Up to 12		
Power statistics CCDF			
Histogram resolution 0.01 dB			
Harmonic distortion			
Maximum harmonic number	10th		
Result Fundamental power (dBm), relative harmonics power (dBc),total harmonic distortion in %			
Intermod (TOI)			
Measure the third-order products and intercepts from tw	o tones		
Burst power			
Methods	Power above threshold, power within burst width		
Result		Single burst output power, average output power, max. power, minimum power within burst, burst width	
Spurious emission			
W-CDMA (1 to 3.6 GHz) table-driven spurious signals; s	search across regions		
Dynamic range	81.3 dB	82.2 dB typical	
Absolute sensitivity	-84.5 dBm	-89.5 dBm typical	

Spectrum emission mask (SEM)		
cdma2000® (750 kHz offset)		
Relative dynamic range (30 kHz RBW)	78.6 dB	84.8 dB typical
Absolute sensitivity	-99.7 dBm	-104.7 dBm typical
Relative accuracy	± 0.12 dB	
3GPP W-CDMA (2.515 MHz offset)		
Relative dynamic range (30 kHz RBW)	81.9 dB	88.1 dB typical
Absolute sensitivity	-99.7 dBm	-104.7 dBm typical
Relative accuracy	± 0.16 dB	

General Specifications

Temperature range			
Operating	0 to 55 °C	Altitude ≤2,300 m	
	0 to 47 °C	Altitude =4,600 m	
Storage	−40 to 70°C		
Altitude	4,600 m (approx. 15,000 f	4,600 m (approx. 15,000 feet)	
Relative humidity	95% relative humidity, no humidity at 55°C	95% relative humidity, non-condensing up to 40°C and decreasing linearly to 50% relative humidity at 55°C	
	From 40°C to 50°C, the dew point	From 40°C to 50°C, the maximum % relative humidity follows the line of constant dew point	
Environment	Indoor use		

EMC

Complies with the essential requirements of the European EMC Directive as well as current editions of the following standards (dates and editions are cited in the Declaration of Conformity):

- IEC/EN 61326-1
- CISPR 11 Group 1, Class A
- AS/NZS CISPR 11
- ICES/NMB-001

This ISM device complies with Canadian ICES-001

Cet appareil ISM est conforme à la norme NMB-001 du Canada

Safety

Complies with European Low Voltage Directive 2014/35/EU

- IEC/EN 61010-1: 2010 AMD1: 2016 / EN61010-1: 2010+A1: 2019; IEC61010-2-030: 2017 / EN 61010-2-030: 2010
- Canada: CAN/CSA-C22.2 No.61010-1-12, UPD1: 2015, UPD2: 2016, AMD1:2018; CAN/CSA-C22.2 No. 61010-2-030-18
- USA: ANSI/UL Std. No. 61010-1:2012 AMD1:2018; ANSI/UL Std No.61010-2-030:2018

Display		
Resolution	1280 x 800	
Size	269 mm (10.6 in.) diagonal (nomina	al) capacitive multi-touch screen
Data storage		
Internal	Removable solid state drive (≥ 256	GB) and secure digital SD memory device
External	Supports USB 3.0/2.0 compatible memory devices	
Weight (without options)		
Net	25.5 kg (56.2 lbs) (nominal)	
Shipping	37.5 kg (82.7 lbs) (nominal)	
Dimensions		
Height	177 mm (7.0 in)	
Width	426 mm (16.8 in)	
Length	556 mm (21.9 in)	

Calibration cycle

The recommended calibration cycle is one year; calibration services are available through Keysight service centers

Acoustic noise emission	Geraeuschemission
LpA < 70 dB	LpA < 70 dB
Operator position	Am Arbeitsplatz
Normal position	Normaler Betrieb
Per ISO 7779	Nach DIN 45635 t.19

Environmental stress

Samples of this product have been type tested in accordance with the Keysight Environmental Test Manual and verified to be robust against the environmental stresses of storage, transportation, and end-use; those stresses include, but are not limited to, temperature, humidity, shock, vibration, altitude, and power line conditions; test methods are aligned with IEC 60068-2 and levels are similar to MIL-PRF-28800F Class 3

Power requirements			
Voltage and frequency (nominal)	100/120 V, 50/60/400 Hz	The instruments can operate with mains supply	
	220/240 V, 50/60 Hz	voltage fluctuations up to ± 10% of the nominal voltage	
Power consumption			
On	630 W maximum		
Standby	45 W		

Inputs and Outputs

Front panel

RF input		
Option 532, 544, 550	2.4mm male, 50 Ω (nominal) (sta	andard)
External Mixing (Option EXM)		
Connection port		
Connector	SMA, female	
Impedance	50 Ω, nominal	
Functions	Triplexed for LO output, IF input,	and mixer bias
Mixer bias range	± 10 mA in 10 μA step	
IF input center frequency		
IF BW path < 25 MHz	322.5 MHz	
IF BW path = 40 MHz	250.0 MHz	
IF BW path = 255 MHz	750 MHz	
IF BW path = 510 MHz	877.148375 MHz	
LO output frequency range	3.75 to 14.0 GHz	
Probe power		
Voltage/current	+15 Vdc, ± 7% at 150 mA max (r	nominal)
· ·	-12.6 Vdc, ± 10% at 150 mA ma	x (nominal)
Probes supported	1130A, 1131A, 1132A, 1134A	
Active probe	1161A	
Passive probe	-5 dB (0-10 MHz, nominal)	
Input return loss	-0 dB (10-40 MHz, nominal)	
USB ports		
Host (3 ports)		
Standard	Compatible with USB 2.0	
Connector	USB Type-A female	
Output current		
Port marked with lightning bolt	1.2 A (nominal)	
Port not marked with lightning bolt	0.5 A (nominal)	
Headphone jack	Miniature stereo audio jack 3.5 n	nm

Rear panel

10 MHz out	
Connector	BNC female, 50 Ω (nominal)
Output amplitude	≥ 0 dBm (nominal)
Frequency	10 MHz x (1+ frequency reference accuracy)
Ext Ref In	
Connector	BNC female, 50 Ω (nominal)
Input amplitude range	-5 to 10 dBm (nominal)
Input frequency	1 to 50 MHz (nominal)
Frequency lock range	± 2 x 10-6 of specified external reference input frequency
Trigger 1 and 2 inputs	
Connector	BNC female
Impedance	10 kΩ (nominal)
Trigger level range	–5 to 5 V
Trigger 1 and 2 outputs	
Connector	BNC female
Impedance	50 kΩ (nominal)
Trigger level range	0 to 5 V (CMOS)
Monitor output	
Connector	VGA compatible, 15-pin mini D-SUB
Format	XGA (60 Hz vertical sync rates, non-interlaced) Analog RGB
Resolution	1024 x 768
Noise source drive +28 V (pulsed)	
Connector	BNC female
SNS Series noise source	For use with Keysight Technologies' SNS series noise sources
Analog out	
Connector	BNC female (used by Option YAS and N9063EM0E analog demodulation measurement application)
USB ports	
Master (2 ports)	
Standard	Compatible with USB 3.0
Connector	USB Type-A female
Output current	0.5 A (nominal)
Slave (1 port)	
Standard	Compatible with USB 3.0
Connector	USB Type-A female

GPIB interface	
Connector	IEEE-488 bus connector
GPIB codes	SH1, AH1, T6, SR1, RL1, PP0, DC1, C1, C2, C3, C28, DT1, L4, C0
GPIB mode	Controller or device
LAN TCP/IP interface	
Standard	1000Base-T
Connector	RJ45 Ethertwist
IF output	
Connector	SMA female, shared by CR3, CRP
Impedance	50 Ω nominal

2 nd IF output, Option CR3	Center frequency
SA mode	322.5 MHz
IQ analyzer with IF BW ≤ 25 MHz	322.5 MHz
IQ analyzer with IF path 40 MHz	250 MHz
IQ analyzer with IF path 255 MHz	750 MHz
IQ analyzer with IF path 510 MHz	877.1484375 MHz
Conversion gain	-1 to +4 dB (nominal) plus RF frequency response
Bandwidth	
Low band	Up to 1 GHz nominal
High band, with preselector bypass	Depends on RF center frequency
Programmable IF output, Option CRP	
Center frequency	
Range	10 to 75 MHz (user selectable)
Resolution	0.5 MHz
Conversion Gain	-1 to +4 dB (nominal) plus RF frequency response
Bandwidth	
Output at 70 MHz	100 MHz nominal
Lower output frequencies	Subject to folding
Residual output signals	≤ -88 dBm nominal

IQ Analyzer

Frequency				
Band	LO Multiple (N)	IF BW ≤ 40 MHz		IF BW > 40 MHz
0	1	10 Hz to 3.6 GHz		10 Hz – 3.4 GHz
1	1	3.5 to 8.4 GHz	3.5 to 8.4 GHz	
2	2	8.3 to 13.6 GHz		8.2 – 13.2 GHz
3	2	13.5 to 17.1 GHz		13.2 – 17.1 GHz
4	4	17.0 to 26.5 GHz		17.1 – 26.5 GHz
5	4	26.4 to 34.5 GHz		26.5 – 34.5 GHz
6	8	34.4 to 50 GHz		34.5 – 50 GHz
Frequency span				
Option B2X	20 Hz – 255 MHz			
Option B5X	20 Hz – 510 MHz			
Resolution bandwidth	Overall	100 mHz to 3 MHz		
	Span = 1 MHz	50 Hz to 1 MHz		
(spectrum measurement)	Span = 10 kHz	1 Hz to 10 kHz		
	Span = 100 Hz	100 mHz to 100 Hz		
Window shapes	Flat top, Uniform, Hai (K-B 70/90/110 dB)	nning, Gaussian, Blackma	an, Blackman-Harri	is, Kaiser Bessel
Analysis bandwidth				
Option B2X	255 MHz			
Option B5X	510 MHz			
IF frequency response (standard	d 10 MHz IF path)			
IF frequency response (demod	ulation and FFT response rela	ative to the center frequer	ncy)	
Center frequency	Span	Preselector	Max. error	RMS
f < 3.6 GHz	≤ 10 MHz	NA	± 0.3 dB	0.04 dB, nominal
3.6 GHz ≤ f ≤ 26.5 GHz	≤ 10 MHz	Off	± 0.3 dB	0.02 dB, nominal
26.5 < f ≤ 50 GHz	≤ 10 MHz	Off	± 0.35 dB	0.026 dB, nominal
IF phase linearity (BW ≤ 10 MHz)			
Center frequency	Span	Preselector	Peak-to-Peak	RMS (nominal)
≤ 3.6 GHz	≤ 10 MHz	N/A	0.4° nominal	0.1°
> 3.6 GHz	≤ 10 MHz	Off	0.4° nominal	0.1°
Dynamic range				
Clipping level at mixer	Center frequency	≥ 20 MHz		
IF gain = Low	-10 dBm	-8 dBm nominal		
IF gain = High	-20 dBm	-17.5 dBm nominal		
Data acquisition (Standard 10 M	Hz IF path)			
Time record length				
IQ analyzer	4,999,999 IQ sample	pairs	Waveform mea	surement
Advanced tool	Data packing		89600 VSA software or fast capture	
	32-bit	64-bit		

Length (IQ pairs)	536 MSa (229 Sa)	268 MSa (228 Sa)	2 GB total memo	ry
Length (time units)	Samples/Sample rate (I	Q pairs)		
Sample rate				
IQ pairs	1.25 × IFBW			
ADC resolution	16 bits			

25 MHz analysis bandwidth	(Standard 25 MHz IF patl	n, licensed as B25)		
IF frequency response (der	modulation and FFT resp	oonse relative to t	he center frequency, 20 to	o 30°C	
Center frequency	Span	Preselector	Max. error	RMS (nominal)	
< 3.6 GHz	10 to ≤ 25 MHz	N/A	±0.45 dB	±0.04 dB	
3.6 GHz ≤ f ≤ 26.5 GHz	10 to ≤ 25 MHz	On		±0.40 dB	
3.6 GHz ≤ f ≤ 26.5 GHz	10 to ≤ 25 MHz	Off	±0.42 dB	±0.05 dB	
26.5 GHz < f ≤ 50 GHz	10 to ≤ 25 MHz	On		±0.50 dB	
26.5 GHz < f ≤ 50 GHz	10 to ≤ 25 MHz	Off	±0.44 dB	±0.03 dB	
IF phase linearity					
Center frequency	Span	Preselector	Peak-to-Peak (nominal)	RMS (nominal)	
20 MHz ≤ f < 3.6 GHz	≤ 25 MHz	N/A	0.6°	0.14°	
f ≥ 3.6 GHz	≤ 25 MHz	Off	1.9°	0.42°	
Dynamic range					
Full scale (ADC clipping)	Default settings, signal at CF				
IF gain = Low	Band				
	0	-8 dBm nominal			
	1 to 4	-7 dBm nominal			
IF gain = High	Band	Mixer level			
	0	-18 dBm nominal, subject to gain limitations			
	1 to 6	-17 dBm nominal, subject to gain limitations			
Effect of signal frequency ≠	CF	Up to ±3 dB no	minal		
Data Acquisition Time record length					
IQ analyzer	4,999,999 IQ sample	pairs	Waveform measu	rement	
Advanced tool	Data packing		89600 VSA softwa	are or fast capture	
	32-bit	64-bit			
Length (IQ pairs)	536 MSa (2 ²⁹ Sa)	268 MSa (2 ²⁸ S	a) 2 GB total memo	ory	
Length (time units)	Samples/Sample rat	e (IQ pairs)			
Sample rate					
IQ pairs	1.25 × IFBW				
ADC resolution	16 bits				

40 MHz analysis bandwidth (Stan	dard 40 MHz IF path, lid	censed as B40)			
IF frequency response (demodul	ation and FFT respons	se relative to the	center fred	quency, 20 to 30°C	
Center frequency	Span	Preselector	Max. erro	or	RMS (nominal)
$30 \text{ MHz} \le f < 3.6 \text{ GHz}$ $3.6 \text{ GHz} \le f \le 8.4 \text{ GHz}$ $8.4 \text{ GHz} \le f \le 26.5 \text{ GHz}$ $26.5 \text{ GHz} < f \le 34.4 \text{ GHz}$ $34.4 \text{ GHz} < f \le 50 \text{ GHz}$	≤ 40 MHz ≤ 40 MHz ≤ 40 MHz ≤ 40 MHz ≤ 40 MHz	N/A Off Off Off	±0.35 dE ±0.46 dE ±0.67 dE	3, ±0.30 dB typical 3, ±0.25 dB typical 3, ±0.33 dB typical 3, ±0.25 dB typical 3, ±0.35 dB typical	±0.08 dB ±0.08 dB ±0.08 dB ±0.1 dB ±0.1 dB
IF phase linearity					
Center frequency	Span	Preselector	Peak-to-	Peak (nominal)	RMS (nominal)
20 MHz ≤ f < 3.6 GHz	≤ 40 MHz	N/A	0.5°	, ,	0.10°
f ≥ 3.6 GHz	≤ 40 MHz	Off	3.6°		0.98°
Dynamic range					
SFDR (spurious-free dynamic ran	nge)				
Signal frequency within ±12 MHz	-	Band 0 1 to 6	-80 dBc	nominal nominal	
Signal frequency within ±18 MHz of center		Band 0 1 to 6	-78 dBc	c nominal c nominal	
Signal frequency anywhere within analysis BW		Band 0 1 to 6		c nominal c nominal	
Full scale (ADC clipping)					
Default settings, signal at CF					
IF gain = Low		Band	Mixer lev		
		0 1 to 4 5 to 6	-8 dBm i -7 dBm i -11 dBm		
IF gain = High		Band	Mixer lev	/el	
		0 1 to 2 3 to 4 5 to 6	-13 dBm -17 dBm -16 dBm -15 dBm) 	
Effect of signal frequency ≠ CF			Up to ±4	dB nominal	
Data Acquisition					
Time record length (IQ pairs)					
IQ analyzer	4,999,999 IQ sam	nple pairs		Waveform measure	ement
Advanced tools	32-bit packing	64-bit packin	g	89600 VSA softwar	re or fast capture
Length (IQ sample pairs)	536 MSa	268 MSa		2 GB total memory	
Length (Time units)	Samples/Sample	rate (IQ pairs)			
Sample rate					
IQ pairs	IFBW x 1.25				
ADC resolution	12 bits				

IQ Analyzer – Option B2X

IF frequency response (demodula	ation and FFT response r	elative to the center frequ	ency, 20 to 30°C	
Center frequency	Span	Preselector	Max. error	RMS (nominal)
400 MHz ≤ f < 1 GHz	≤ 255 MHz	N/A	±0.8 dB, ±0.4 dB typical	,
1 GHz ≤ f < 3.4 GHz	≤ 255 MHz	N/A	±0.5 dB, ±0.3 dB typical	
3.4 GHz ≤ f ≤ 8.2 GHz	≤ 255 MHz	Off	±0.5 dB, ±0.35 dB typica	
8.2 GHz ≤ f ≤ 26.5 GHz	≤ 255 MHz	Off	±0.6 dB nominal	±0.2 dB
26.5 GHz ≤ f ≤ 50 GHz	≤ 255 MHz	Off	±0.8 dB nominal	±0.2 dB
IF phase linearity				
Center frequency	Span	Preselector	Peak-to-Peak (nominal)	RMS (nominal)
20 MHz ≤ f < 3.4 GHz	≤ 255 MHz	N/A	3°	0.6°
3.4 GHz ≤ f < 26.5 GHz	≤ 255 MHz	Off	2°	0.5°
26.5 GHz ≤ f ≤ 50 GHz	≤ 255 MHz	Off	4°	0.8°
Dynamic range	= 200 Mil i2	, on		
SFDR (spurious-free dynamic rar	nge)			
Signal frequency anywhere within	-	-78 dBc nominal		
Full scale clipping				
Default settings, signal at CF				
IF gain = Low	Band	Mixer level		
	0	+2 dBm nominal		
	1 to 2	+3 dBm nominal		
	3 to 4	0 dBm nominal		
	5 to 6	-11 dBm nominal		
IF gain = High	Band	Mixer level		
IF gain offset = 0 dB	0	-3 dBm nominal		
	1 to 2	-6 dBm nominal		
	3 to 4	-9 dBm nominal		
	5 to 6	-11 dBm nominal		
Effect of signal frequency ≠ CF		Up to ±4 dB nominal		
Data Acquisition Time record length (IQ pairs)				
IQ analyzer	4,999,999 IQ sample	pairs	Waveform measureme	ent
Advanced tools	32-bit packing	64-bit packing	89600 VSA or fast captu	ire
Length (IQ sample pairs)	1073 MSa (230 Sa)	536 MSa (229 Sa)	4 GB total memory (opti	on DP4)
Length (Time units)	Length of IQ sample p	pairs/sample rate (IQ pairs	s)	
Sample rate				
IQ pairs	Minimum of (Span x 1	.25, 300 MSa/s)		
ADC resolution	14 bits			

IQ Analyzer – Option B5X

IF frequency response (demodula	tion and EET response rel	ative to the center free	allency 20 to 30°C	
Center frequency	Span	Preselector	Max. error	RMS (nominal)
600 MHz ≤ f < 3.4 GHz	≤ 500 MHz	N/A	±0.75 dB, ±0.41 dB typical	±0.1 dB
3.4 GHz ≤ f < 8.2 GHz	≤ 500 MHz	Off	±0.75 dB, ±0.41 dB typical	±0.1 dB
8.2 GHz ≤ f ≤ 26.5 GHz	≤ 510 MHz	Off	±0.8 dB nominal	±0.5 db
26.5 GHz ≤ f ≤ 50 GHz	≤ 510 MHz	Off	±1.0 dB nominal	
IF phase linearity	2 010 WH12	, on	±1.0 dB florillidi	
Center frequency	Span	Preselector	Peak-to-Peak (nominal)	RMS (nominal)
20 MHz ≤ f < 3.4 GHz	≤ 510 MHz	N/A	5°	1.0°
3.4 GHz ≤ f < 26.5 GHz	≤ 510 MHz	Off	6°	1.4°
26.5 GHz ≤ f ≤ 50 GHz	≤ 510 MHz	Off	7°	1.6°
	2 0 10 WH 12	Oli	•	1.0
Dynamic range		_	1	
SFDR (spurious-free dynamic ran	,			
Signal frequency anywhere within	analysis BW	-75 dBc nominal		
Full scale clipping				
Default settings, signal at CF				
IF gain = Low	Band	Mixer level		
	0	+1 dBm nominal		
	1 to 2	+3 dBm nominal		
	3 to 4	0 dBm nominal		
	5 to 6	-11 dBm nominal		
IF gain = High	Band	Mixer level		
IF gain offset = 0 dB	0	-4 dBm nominal		
	1 to 2	-9 dBm nominal		
	3 to 4	-13 dBm nominal		
	5 to 6	-11 dBm nominal		
Effect of signal frequency ≠ CF		Up to ±4 dB nomir	nal 	
Data Acquisition Time record length (IQ pairs)				
IQ analyzer	4,999,999 IQ sample p	pairs	Waveform measurer	nent
Advanced tools	32-bit packing	64-bit packing	89600 VSA software	or fast capture
Length (IQ pairs)	-			
IFBW ≤ 255.176 MHz	1073 MSa (230 Sa)	536 MSa (229 Sa)	4 GB total memory (opt. DP4)
IFBW > 255.176 MHz	2147 MSa (231 Sa)	1073 MSa (230 Sa		•
Length (Time units)	Length of IQ sample pa		, , ,	•
Sample rate				
IFBW ≤ 255.176 MHz	Minimum of (Span x 1.	25, 300 MSa/s)		
IFBW > 255.176 MHz		mum of (Span x 1.25, 600 MSa/s)		
ADC resolution	14 bits	·		

Real-time Spectrum Analyzer

Option RT1 and RT2

Real-time analysis				
Real-time analysis bandwidth				
Option RT1	Up to 509.47 MHz	Analysis bandwidth determines the maximum		
Option RT2	Up to 509.47 MHz	real-time bandwidth		
Option DUA	Up to 2 x 255 MHz at same	Up to 2 x 255 MHz at same center frequency, requires Option B5X		
Minimum detectable signal dura	Minimum detectable signal duration with > 60 dB 3.33 ns, with option B2X or B5X			
Minimum signal duration with 10	0% POI at full amplitude range	For frequency mask triggering (FMT)		
Option RT1	17.3 µs	Signal is at mask level		
Option RT2	3.57 µs	Signal is at mask level		
Minimum acquisition time	104 µs			
FFT rate	292,969/s			

Learn more at: www.keysight.com

For more information on Keysight Technologies' products, applications or services, please contact your local Keysight office. The complete list is available at: www.keysight.com/find/contactus

